

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

H200

RQ:

2021-08-23-53

Project:

SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 07/22/2021

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Lillard	08/26/21	0930	23.3	757	8.5	8.51	99.8	—	1.08	P/F	L/F
ICV	//	//	0934	23.3	757	8.5	8.51	99.7	—	1.05	P/F	L/F
CCV	Andrew Luering	8/26/21	1457	25.1	755	8.2	8.22	99.7	—	1.08	P/F	L/F
CCV			1457								P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Lillard	08/26/21	0936	1105E62	05/23	1000	1000	P/F	L/F
ICV	//	//	0938	1102792	01/23	100	101.3	P/F	L/F
CCV	Andrew Luering	8/26/21	1459	1105C52	05/2023	5000	4997	P/F	L/F
CCV			1459					P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Lillard	08/26/21	0940	210039	10/23	7.0	23.2	7.01	-23.1	P/F	L/F
Calibr.	I		0942	2104A60	03/23	4.0	23.1	4.01	157.3	P/F	L/F
Calibr.	I		0944	1102792	01/23	10.0	23.1	10.00	-192.4	P/F	L/F
ICV	I		0946	1102792	01/23	10.0	23.2	10.04	-193.1	P/F	L/F
CCV	Andrew Luering	8/26/21	1501	2104A60	03/2023	4.0	22.4	3.98	155.6	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 -169.3 , slope from 4 to 7 180.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 07/22/2021

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Lillard	08/26/21	0948	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: