

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodor

RQ: 2021-10-25-07

Project: SMR/CN

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/19/2021

DO DEP SOP FD 1500	Name	Date	Time GT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	11/1/21	1000	22.9	759.0	8.6	8.58	99.8	—	1.07	(P) F	(L) F
ICV			1004	22.9	759.1	8.6	8.58	99.9	—	1.07	(P) F	(L) F
CCV			1400	23.9	757.1	8.4	8.40	99.7	—	1.07	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Conduct. EX 1200	Name	Date	Time GT-ET	Lot#	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	11/1/21	1006	1108F51	08/23	1000	1000	(P) F	(L) F
ICV			1008	1102792	21/23	100	99.6	(P) F	(L) F
CCV			1404	1105551	10/6/23	100,000	102146	(P) F	(L) F
CCV				1107614				P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP EX 1100	Name	Date	Time GT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	11/1/21	1010	210039	10/23	7.0	23.0	7.01	-24.2	(P) F	(L) F
Calibr.			1012	2107E40	07/23	4.0	23.1	4.00	152.9	(P) F	(L) F
Calibr.			1014	1106829	11/22	10.0	23.0	10.00	-198.2	(P) F	(L) F
ICV			1016	1106829	11/22	10.0	23.0	10.00	-198.3	(P) F	(L) F
CCV			1406	2107E40	07/23	4.0	23.2	4.01	152.3	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range $\pm 180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/19/2021

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time GT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	11/1/21	1018	0.50	0.50	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: