

Chemical Analysis Report

SE-DIST-2021-08-11-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **SMP - 3129B1 August**
Request ID: **RQ-2021-08-09-80**
Customer: **SE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
2199 South Rock Road
Fort Pierce, FL 34945
Attn: Drew Liddick

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 02-SEP-2021 13:03

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: South Prong Sebastian @ Park Dock

Collection Date/Time: 08/10/2021 14:15

Field ID: G5SE0079

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2268810	EPA 180.1 Rev. 2.0	Turbidity	3.4		NTU	P402346	
	SM 2320 B-2011	Total Alkalinity	100		mg CaCO3/L	P402594	
	SM 2540 D-2011	TSS	3	I	mg/L	P402753	
	SM 4500 F-C-2011	Fluoride	0.29		mg F/L	P402597	
2268811	EPA 350.1 Rev. 2.0	Ammonia-N	0.13		mg N/L	P402580	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P402887	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.045		mg N/L	P402362	
	EPA 365.1 Rev. 2.0	Total-P	0.45		mg P/L	P402462	
	SM 5310 B-2011	Organic Carbon	21		mg C/L	P402919	
2268812	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.28		mg P/L	P402341	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P402346

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402580

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402887

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402362

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402462

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402341

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P402594

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-2011
Batch ID: P402753

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P402597

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P402919

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402580

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402887

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402362

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402462

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	109		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402341

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.4		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P402594

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P402597

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.6		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P402919

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P402580

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268659	Ammonia-N	98.2	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P402887

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268693	Kjeldahl Nitrogen	107	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P402362

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268660	NO2NO3-N	101	99.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P402462

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2266313	Total-P	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P402341

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268877	O-Phosphate-P	95.8	95.9	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P402597

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268420	Fluoride	104	104	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P402919

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2268467	Organic Carbon	98.6	96.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P402346

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268797	Turbidity	1.20	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P402580

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268659	Ammonia-N	1.15	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P402887

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268693	Kjeldahl Nitrogen	3.63	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P402362

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268660	NO2NO3-N	0.975	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P402462

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2266313	Total-P	0.320	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P402341

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268877	O-Phosphate-P	0.104	Spike	P	0 - 20

Reference Method: SM 2320 B-2011

Batch ID: P402594

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268420	Total Alkalinity	1.27	Sample	P	0 - 5.2

Reference Method: SM 4500 F-C-2011

Batch ID: P402597

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268420	Fluoride	0.495	Spike	P	0 - 2.1

Quality Assurance Report

Precision

Reference Method: SM 5310 B-2011
Batch ID: P402919

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2268467	Organic Carbon	1.46	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A107186

Included Lab Sample IDs: 2268812

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	96.7	97.0	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A107187

Included Lab Sample IDs: 2268810

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A107200

Included Lab Sample IDs: 2268811

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A107271

Included Lab Sample IDs: 2268811

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	108	107	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A107292

Included Lab Sample IDs: 2268811

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.3	98.7	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A107300

Included Lab Sample IDs: 2268810

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	99.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A107300

Included Lab Sample IDs: 2268810

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A107443

Included Lab Sample IDs: 2268811

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.0	95.6	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A107458

Included Lab Sample IDs: 2268811

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	106	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					1.20	
EPA 350.1 Rev. 2.0	Ammonia-N	99.2	98.2	100			1.15
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	103	107			3.63
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	101	99.6			0.975
EPA 365.1 Rev. 2.0	Total-P	109	104	104			0.320
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.4	95.8	95.9			0.104
SM 2320 B-2011	Total Alkalinity	102				1.27	
SM 4500 F-C-2011	Fluoride	98.6	104	104			0.495
SM 5310 B-2011	Organic Carbon	95.5	98.6	96.6			1.46

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2268810
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2268811
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2268811
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2268811
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2268811
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2268812
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2268810
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2268810
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2268810
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2268811

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/11/2021			08/11/2021 15:39	Sarah A Nixon	2268810
EPA 350.1 Rev. 2.0	08/11/2021			08/17/2021 13:25	Roberta Tatti	2268811
EPA 351.2 Rev. 2.0	08/11/2021	08/24/2021 13:34	Benjamin T. Nordmann	08/25/2021 11:24	Virginia P. Brown	2268811
EPA 353.2 Rev. 2.0	08/11/2021			08/12/2021 10:10	Beverly Y. Sanders	2268811
EPA 365.1 Rev. 2.0	08/11/2021	08/13/2021 15:00	Simonne.V. Calhoun	08/16/2021 13:59	Shengyu Ding	2268811
EPA 365.1 Rev. 2.0 dissolved	08/11/2021			08/11/2021 16:37	Patrick M. Roche	2268812
SM 2320 B-2011	08/11/2021			08/12/2021 22:57	Dale L. Simmons	2268810
SM 2540 D-2011	08/11/2021			08/16/2021 15:36	Yijie Li	2268810
SM 4500 F-C-2011	08/11/2021			08/12/2021 22:57	Dale L. Simmons	2268810
SM 5310 B-2011	08/11/2021			08/23/2021 20:26	Madeline Gruver	2268811