

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: STITCH

RQ: 2022-02-21-58

Project: Barium Treat

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/3/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Vincent McClure	2/22/22	844	21.1	768.7	9.0	9.04	101.3	-	1.094	Ⓟ/F	Ⓟ/F
ICV	↓	↓	846	20.6	768.8	9.1	9.10	101.8	-	1.094	Ⓟ/F	Ⓟ/F
CCV	↓	↓	5:20	21.6	767.0	8.9	8.98	101.9	-	1.094	Ⓟ/F	Ⓟ/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Vincent McClure	2/22/22	851	210917B	09/22	1000	999	Ⓟ/F	Ⓟ/F
ICV	↓	↓	854	210926A	05/22	100	101	Ⓟ/F	Ⓟ/F
CCV	↓	↓	525	211122B	12/22	1407 1407	1407	Ⓟ/F	Ⓟ/F
CCV	↓	↓	525	210917C	09/22	50000	49865	Ⓟ/F	Ⓟ/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Vincent McClure	2/22/22	857	2112B19	12/23	7.02	21.7	6.98	-225	Ⓟ/F	Ⓟ/F
Calibr.	↓	↓	901	2108F51	8/23	4.00	21.0	3.82	155.4 155.8	Ⓟ/F	Ⓟ/F
Calibr.	↓	↓	904	2108H37	2/23	10.04	21.5	10.02	195.4 195.4	Ⓟ/F	Ⓟ/F
ICV										P/F	L/F
CCV	↓	↓	529	2112B19	12/23	7.02	21.6	7.10	-25.0	Ⓟ/F	Ⓟ/F
CCV	↓	↓	532	2108H37	2/23	10.04	21.6	10.04	-197.0	Ⓟ/F	Ⓟ/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 172.9, slope from 4 to 7 177.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? ☒ YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Vincent McClure	2/22/22	841	0.00	0.0	Ⓟ/F	Ⓟ/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: