

RQ- 2022-02-21-23

Login Station ID: #4

Shipping Method: ☒ FedEx | ☐ UPS | ☐ HD | ☐ Greyhound

Date/Time of Receipt: 03/17 /22 @ 9:25

[illegible]

***All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.**

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes _____ No X

Evidence Tape on Bottles?

Yes _____ No _____ NA X

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes ☒ No ☐

Caps on tight?

Yes V No

Sufficient Sample Volume:

Yes No

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? *(Blue dot on container)*

Yes _____ No X Init: 32

Chlorine detected? (One container checked per Field ID)

Yes _____ No _____ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

****Acid Preserved Samples: pH< 2.0?**

Yes ✓ No NA Init:

****W-1-4-DIOX (Green dot on container) preserved to pH<4.0?**

Yes _____ No _____ NA X Init: ~

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by:

Date: 03/17/22

Event contains NCRs (Y/N)? Y

Event ID

NE-DIST-2022-03-17-02 (SMP)
Date Received: 17-MAR-2022

Date Received: 17-MAR-2022 09:25

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	#4 Temp Gun EXP:10/19/2022
pH Test Strips 0 – 6	223819AV EXP:08/30/2023
pH Paper 0 – 3	N/A
pH Paper 0 – 13	207621 EXP:03/15/2024
Chlorine Test Strips	0070 EXP:02/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 3/16/2022
Customer: NE-DIST
RQ: RQ-2022-02-21-23
Collected By: Kimberly Appleby, Vincent McClure
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB


Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
UNNAMED STREAM @ CR 121	 19010127
FB	 FB03162022
Tributary to Deep Creek @ Kinard Rd.	 19010083

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 3/16/2022
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:



Rec'd/Inspected Date: 3 - 17 - 2022

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: STITCH RQ: 2022-02-21-23 Project: Nassau West

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/3/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Vincent Muller	3/16/22	8:19	21.1	761.5	8.9	/	100.1	/	1.07	0/F	0/F
ICV	↓	↓	8:21	21.1	761.6	8.9	8.95	100.3	/	1.07	0/F	0/F
CCV	↓	↓	8:32	21.3	760.3	8.8	8.84	99.7	/	1.07	0/F	0/F
CCV			1432								P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75. Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mos/cm	Meter Reading μ mos/cm	Pass / Fail	Lab / Field
Calibr.	Vincent Muller	3/16/22	8:25	210917B	09/22	1000	1000	0/F	0/F
ICV	↓	↓	8:27	211120B	12/22	1413	1412	0/F	0/F
CCV	↓	↓	8:34	210917A	09/22	100	99.6	0/F	0/F
CCV			1434					P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Vincent Muller	3/16/22	8:31	2112019	12/23	7.02	21.5	7.02	-29.2	0/F	0/F
Calibr.	↓	↓	8:35	2201088	1/24	4.00	21.5	4.0	145.5	0/F	0/F
Calibr.	↓	↓	8:37	2202A63	7/23	10.04	21.5	10.04	-199.9	0/F	0/F
ICV	↓	↓	8:49	2112019	12/23	7.02	21.5	7.06	-29.8	0/F	0/F
CCV	↓	↓	8:55	2201088	1/24	4.00	21.7	3.99	145.9	0/F	0/F
CCV			1437							P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 10.7 slope from 4 to 7 14.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES NO / NA (not surf. water project)
 If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/3/22
 If NO, what will be used? (each one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Vincent Muller	3/16/22	8:50	0.00	0.00	0/F	0/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



RQ-2022-02-21-23


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Kimberly Appleby, Vincent McClure

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  19010127										
Location: UNNAMED STREAM @ CR 121										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTH)
3/16/2022	1100 EST	4.21	42.7	16.1	4.55	0.3	0.4 ☐ VOB (S)	0.7	62.8	0.03
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	A
		Lot # SA1105060		Exp: 4/30/22						
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	A
		Lot # NA1105100		Exp: 6/7/22						
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #		Syringe Lot#						
Biochemical Oxygen Demand (BP-1L)		BODTEST OV-BODUNIN				☐ Ice				
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice			1	A
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		Formalin Lot #		Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice			1	A
		Contract Lab								
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp. Date		
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI		WE8321-MS		W-PSNP-TQ		W-CT-CL-R	MCAA Lot#	
		W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA** Buffer Solution)				W-TR-SQ-R		
		W-PCB-TQ-R								
Phytoplankton/Periphyton		DTY-QN-C		PKD-QN-BV		☐ Ice				
Name: Kimberly Appleby Vincent McClure						Signature: 			Date: 03/16/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: UNNAMED STREAM @ CR 121**Date:** 3/16/2022**WIN/Field ID:** 19010127**WBID:** 2178**ENR:** -**Latitude:** 30.51667**County:** NASSAU**Org ID:** 21FLA**Longitude:** -81.99292**Receiving Body of Water:** Nassau - St. Marys**RQ:** 2022-02-21-23

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒	☐	☒	☐	☐
☒	Vincent McClure	☐	☒	☐	☒	☒
☐		☐	☐	☐	☐	☐
☐		☐	☐	☐	☐	☐

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Secchi	
Secchi Equipment ID: Disco	

Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other _____	☐ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☒ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# STITCH	
Photos: NO	Flow: FLOW	Tide:	Tide Times: High: Low
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: CLOUDY, LIGHT SPORADIC RAIN. HEAVY RAIN OVER PAST WEEKEND. 0.75 INCHES OF RAIN PAST 48 HRS. HIGHER FLOW RATE THAN USUAL.			

Duplicate Sample	
Time Zone: EST Time: _____ Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID: _____	

Blank	
Time Zone: EST Time: 1105 Field ID: FB03162022	(Blank Type_DDMMYYYY)
DI Source: NE ROC Lab Location: UNNAMED STREAM @ CR 121	
DI Type: Field Blank Equipment ID: _____	

DI Container (Name/ID): Cargirl	DI Container Type: Carboy
--	----------------------------------

LIMS EVENT ID: _____ WIN Import ID: _____

Enter Field Parameters, Verify Station ID In LIMS DMT: _____ (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: _____ (Initials/Date)

Scan/Save Field Sheets _____ (Initials/Date) Migrate Data to WIN: _____ (Initials/Date) Verify Data In WIN: _____ (Initials/Date)



RQ- 2022-02-21-23

Customer: WQAP

Project ID: SMP

Collected By: Kimberly Appleby, Vincent McClure

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

BLANK



Field ID: FB03162022

Comments:

Matrix: Field Blank

(Check one) ☐ COMPOSITE ☒ GRAB

Date

Time

DI Source: NE ROC Lab

3/16/2022

1105

EST

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA1105060 Exp: 4/30/22	☒ Ice ☒ H2SO4	☒ <2	1	C
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP Lot # NA1105100 Exp: 6/7/22	☒ Ice ☒ HNO3*	☒ <2	1	C
Filtered Nutrients (P125ML)	☐ W-PO4-F Filter ☐ Field Filtered 0.45 µm Filter Lot # Syringe Lot #	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-IC / TDS	☒ Ice		1	C
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice	MCAA Exp. Date		
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R	☐ Ice	MCAA Lot#		
Name: Kimberly Appleby Vincent McClure		Signature:		Date: 03/16/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PLAN FIELD SHEET

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Tributary to Deep Creek @ Kinard Rd.

Date: 3/16/2022

WIN/Field ID: 19010083

WBID: 2196

ENR:

Latitude: 30.44125

County: NASSAU

Org ID: 21FLA

Longitude: -82.01734

Receiving Body of Water: Deep Creek

RQ: 2022-02-21-23

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒	☒	☒	☒	☒
☒	Vincent McClure	☒	☒	☒	☒	☒
☐		☐	☐	☐	☐	☐
☐		☐	☐	☐	☐	☐

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☐ Low ☒ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# STITCH	
Photos: YES	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: PARTLY CLOUDY, LIGHT RAIN DURING MORNING. HEAVY RAIN OVER PAST WEEKEND. 0.75 INCHES OF RAIN LAST 48 HRS. HIGHER THAN USUAL FLOW RATE. UPSTREAM IS A BRAIDED SYSTEM, TOOK SAMPLE IN MAIN FLOW UPSTREAM OF CULVERT DUE TO ACCESS.			

Duplicate Sample	
Time Zone: EST	Time: Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone:	Time: Field ID: (Blank Type_DDMMYYYY)
DI Source:	Location:
DI Type:	Equipment ID:
DI Container (Name/ID):	DI Container Type:

LIMS EVENT ID:	WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date)	QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)
Scan/Save Field Sheets (Initials/Date)	Migrate Data to WIN: (Initials/Date)
Verify Data In WIN: (Initials/Date)	



RQ-2022-02-21-23


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Kimberly Appleby, Vincent McClure

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  19010083										
Location: Tributary to Deep Creek @ Kinard Rd.										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/16/2022	1135 EST	6.58	67.6	16.7	4.4	0.15	0.33 ☒ VOB (S)	0.33	56.6	0.03
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	A
		Lot #	SA1105060		Exp:	4/30/22				
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	A
		Lot #	NA1105100		Exp:	6/7/22				
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #			Syringe Lot #					
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice				
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice			1	A
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	A
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ CR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ CR-GFD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-TR-SQ-R W-PCB-TQ-R				☐ Ice		MCAA Lot#		
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice				
Name: Kimberly Appleby Vincent McClure					Signature: 				Date: 03/16/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

Date of Request: 21-JAN-2022
 Created By: PARRISH_J On: 21-JAN-2022
 Modified By: AYRES_J On: 16-MAR-2022
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: Nassau West 2022

Send Coolers To:

Phone: 904-256-1700
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: P
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 16-MAR-2022
 Biology Request Reviewed By:
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped: YES On: 08-FEB-2022
 Sampling Kit Packed By: MCCOY_I
 Preservatives Needed: NO
 Date to Receive Samples: 21-FEB-2022
 Logged In By: ROBERTS_Z On: 24-FEB-2022

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf.Fresh)Nutrients, Metals, Bacteria
 Group B:(Surf.Fresh) Metals, Bacteria
 Group C:(Blank)Nutrient,Metal

Packing Notes:
 None

Bottle Group A (Water) With 5 Samples:

Bottle Type: P-1L	Number of Bottles: 5	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 5	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WO-THI	Number of Bottles: 5	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML	Number of Bottles: 5	Preserved With: HNO ₃
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃
Hardness		

Bottle Group A (Water) With 5 Samples:

Bottle Type: P-500ML	Number of Bottles: 5	Preserved With: HNO3
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML	Number of Bottles: 5	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 5	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group B (Water) With 2 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 2	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Bottle Group B (Water) With 2 Samples:

Bottle Type: P-500ML Number of Bottles: 2 Preserved With: HNO3
 W-ICPMS Template: DEFAULT (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

*Boron
 Aluminum
 Antimony
 Arsenic
 Barium
 Beryllium
 Cadmium
 Chromium
 Cobalt
 Copper
 Lead
 Manganese
 Molybdenum
 Nickel
 Selenium
 Silver
 Thallium

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3
 W-HARD Template: DEFAULT (SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Hardness

W-ICP Template: SMP-FRESH (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Calcium
 Iron
 Magnesium
 Potassium
 Sodium
 Strontium
 Tin
 Titanium
 Vanadium
 Zinc

W-ICPMS Template: DEFAULT (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

*Boron
 Aluminum
 Antimony
 Arsenic
 Barium
 Beryllium
 Cadmium
 Chromium
 Cobalt
 Copper
 Lead
 Manganese
 Molybdenum
 Nickel
 Selenium
 Silver
 Thallium

Bottle Type: P-125ML Number of Bottles: 1 Preserved With: FILTER-HNO3
 W-HARD-F Template: DEFAULT (SM 2340B) Hardness by calculation, in dissolved aqueous matrices, reported as mg/L CaCO3
 Hardness

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: ICE And H2SO4
 W-NH3 Template: DEFAULT (EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
 W-NO2NO3 Template: DEFAULT (EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-500ML		Number of Bottles: 1	Preserved With: ICE And H2SO4
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L	
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices	
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices	
Bottle Type: P-1L		Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L	
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices	
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices	
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices	
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices	
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.