

Chemical Analysis Report

NE-DIST-2022-02-23-03

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Little Trout CM Retrieve**

Request ID: **RQ-2022-02-21-58**

Customer: **NE-DIST**


Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 18-MAR-2022 18:43



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Little Trout River at Old Kings Rd.

Collection Date/Time: 02/22/2022 11:35

Field ID: 20030739

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2307744	EPA 180.1 Rev. 2.0	Turbidity	2.6		NTU	P410270	
	SM 2320 B-2011	Total Alkalinity	45		mg CaCO3/L	P410659	
	SM 2540 D-2011	TSS	2	I	mg/L	P410805	
	SM 4500 F-C-2011	Fluoride	0.061	I	mg F/L	P410662	
2307745	EPA 350.1 Rev. 2.0	Ammonia-N	0.027		mg N/L	P410627	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.93		mg N/L	P410549	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.022		mg N/L	P410624	
	EPA 365.1 Rev. 2.0	Total-P	0.050		mg P/L	P410485	
	SM 5310 B-2011	Organic Carbon	24		mg C/L	P410541	
2307746	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.023		mg P/L	P410263	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P410270

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410627

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P410549

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410624

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P410485

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P410263

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P410659

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P410805

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P410662

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P410541

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P410270

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	97.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410627

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P410549

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410624

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P410485

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P410263

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P410659

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	96.8		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P410662

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.6		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P410541

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P410627

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307774	Ammonia-N	95.8	97.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P410549

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307334	Kjeldahl Nitrogen	98.0	99.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P410624

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307745	NO2NO3-N	92.1	90.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P410485

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307726	Total-P	106	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P410263

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307668	O-Phosphate-P	94.2	95.8	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P410662

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2308113	Fluoride	96.7	96.7	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P410541

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2307667	Organic Carbon	96.2	96.0	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P410270

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307585	Turbidity	7.31	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P410627

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307774	Ammonia-N	1.86	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P410549

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307334	Kjeldahl Nitrogen	1.84	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P410624

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307745	NO2NO3-N	1.96	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P410485

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307726	Total-P	1.93	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P410263

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307668	O-Phosphate-P	1.68	Spike	P	0 - 20

Reference Method: SM 2320 B-2011

Batch ID: P410659

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2308113	Total Alkalinity	2.96	Sample	P	0 - 5.2

Reference Method: SM 4500 F-C-2011

Batch ID: P410662

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2308113	Fluoride	0.0	Spike	P	0 - 2.1

Quality Assurance Report Precision

Reference Method: SM 5310 B-2011
Batch ID: P410541

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2307667	Organic Carbon	0.0602	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A110574

Included Lab Sample IDs: 2307746

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.4	99.3	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A110576

Included Lab Sample IDs: 2307744

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 5310 B-2011

Run ID: A110701

Included Lab Sample IDs: 2307745

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.4	96.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A110729

Included Lab Sample IDs: 2307745

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A110732

Included Lab Sample IDs: 2307745

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	96.0	95.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110733

Included Lab Sample IDs: 2307745

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.7	98.3	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A110751

Included Lab Sample IDs: 2307744

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.7	95.2	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A110751

Included Lab Sample IDs: 2307744

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	98.1	P/P	90 - 110

Quality Assurance Report

Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110781

Included Lab Sample IDs: 2307745

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	97.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity	97.8				7.31	
EPA 350.1 Rev. 2.0	Ammonia-N	97.0	95.8	97.6			1.86
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	98.0	99.9			1.84
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	98.0	92.1	90.1			1.96
EPA 365.1 Rev. 2.0	Total-P	102	106	104			1.93
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	94.2	95.8			1.68
SM 2320 B-2011	Total Alkalinity	96.8				2.96	
SM 4500 F-C-2011	Fluoride	97.6	96.7	96.7			0.0
SM 5310 B-2011	Organic Carbon	96.0	96.2	96.0			0.0602

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2307744
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2307745
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2307745
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2307745
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2307745
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2307746
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2307744
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2307744
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2307744
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2307745

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	02/23/2022			02/23/2022 15:51	Khloe J Berg	2307744
EPA 350.1 Rev. 2.0	02/23/2022			03/03/2022 12:05	Ping Hua	2307745
EPA 351.2 Rev. 2.0	02/23/2022	03/03/2022 12:56	Samantha L Bates	03/04/2022 15:50	Alexandra J Mattheus	2307745
EPA 353.2 Rev. 2.0	02/23/2022			03/03/2022 09:55	Beverly Y. Sanders	2307745
EPA 365.1 Rev. 2.0	02/23/2022	03/02/2022 15:00	Simonne.V. Calhoun	03/03/2022 12:12	Shengyu Ding	2307745
EPA 365.1 Rev. 2.0 dissolved	02/23/2022			02/23/2022 17:26	James L Waggerby	2307746
SM 2320 B-2011	02/23/2022			03/02/2022 21:16	Dale L. Simmons	2307744
SM 2540 D-2011	02/23/2022			02/28/2022 15:12	Yijie Li	2307744
SM 4500 F-C-2011	02/23/2022			03/02/2022 21:16	Dale L. Simmons	2307744
SM 5310 B-2011	02/23/2022			03/02/2022 01:37	Khloe J Berg	2307745