

RQ-2022-03-07-24

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 3/10/2022 9:35

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
2.8 0.8				9 9		✓			5225 5544 6446 5225 5544 6457
						✓			

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☐ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☐ No ☐Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☒ No ☐ Init: IM

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☒ Init: IM

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☒ No ☐ NA ☐ Init: IM

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☐ NA ☒ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM

Date: 3/10/2022

Event contains NCRs (Y/N)? ☒

Event ID:

LSJR South West 2022

NE-DIST-2022-03-10-02 (SMP)

Page 1 of 2, 06/15/2021

Date Received: 10-MAR-2022_09:35

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/19/22
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	207621 exp: 3/15/24
Chlorine Test Strips	0070 exp: 02/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 3/9/2022
Customer: NE-DIST
RQ: RQ-2022-03-07-24

Collected By: Brian Davis, Katrin Villinger
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Swimming Pen CR @ SR 220	 20030191
Swimming Pen Creek at Eagles Creek Dr.	 G2NE1188
LITTLE BLACK CR AT SR 21	 20030629
LITTLE BLACK CREEK AT SR 220	 20030667
BULL CR SW CROSSING CR 215	 20030480
PETERS CREEK @ ROSEMARY HILL RD	 20030571

RELINQUISHED BY(SIGNATURE):

Katrin Villinger

DATE/TIME: 3/9/2022

1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

IM

Rec'd/Inspected Date: *3/10/22*

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "x" this box if there are qualified data on this page

Meter ID: Bruce

RO: 2022-03-07-24

Project: SMP LSPR South West

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/3/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Villinger	3/9/22	7:35	21.4	759.6	8.8	—	99.9	—	10254	P/F	L/F
ICV	↓	↓	7:37	21.4	759.6	8.8	8.85	100.0	—	10254	P/F	L/F
CCV	B. Pruis	↓	16:30	20.4	758.5	9.0	9.00	99.9	—	103	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	K. Villinger	3/9/22	7:39	210917E	9/22	1000	1000	P/F	L/F
ICV	↓	↓	7:42	210426A	5/22	100	101.7	P/F	L/F
CCV	B. Pruis	↓	16:32	211122B	12/22	1413	1376	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K. Villinger	3/9/22	7:45	2112019	12/23	7.02	21.8	4.02	-15.0	P/F	L/F
Calibr.	↓	↓	7:47	2201088	1/24	4.00	21.5	4.00	164.3	P/F	L/F
Calibr.	↓	↓	7:50	2202463	7/23	10.03	21.8	10.03	-184.5	P/F	L/F
ICV	↓	↓	7:54	2201088	1/21	4.00	21.7	4.01	162.4	P/F	L/F
CCV	B. Pruis	↓	16:54	2202463	7/23	10.04	21.3	10.10	188.9	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.5, slope from 4 to 7 179.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K. Villinger	3/9/22	7:55	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



RQ-2022-03-07-24


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Katrin Villinger, Brian Davis

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:										
20030191										
Location:		Swimming Pen CR @ SR 220								
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)
3/9/2022	1010 EST	6.02	70.7	23.2	7.28	0.3	0.6 ☒ VOB (S)	0.6	1354	0.68
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	A
		Lot # SA1105060		Exp: 04/30/22						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ NO3*		☐ <2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #				Syringe Lot #				
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice				
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHL / W-COLOR / W-SD4-IC / W-				☒ Ice			1	A
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	A
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ CR-DG3				☐ Ice				
		☐ PCR-GULL2 ☐ CR-GFD								
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI WE8321-MS W-PSNP-TQ W-CT-CL-R				☐ Ice		MCAA Lot#		
		W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-TR-SQ-R								
		W-PCB-TQ-R								
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice				
Name: Katrin Villinger Brian Davis					Signature: 				Date: 03/09/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION**

STRATEGIC MONITORING PLAN FIELD SHEET

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Swimming Pen CR @ SR 220

Date: 3/9/2022

WIN/Field ID: 20030191

WBID: 2410

ENR: -

Latitude: 30.09997

County: CLAY

Org ID: 21FLA

Longitude: -81.74634

Receiving Body of Water: Lower St. Johns

RQ- 2022-03-07-24

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Katrin Villinger	☒		☒					
☒	Brian Davis		☒		☒	☒	☒		
☐									
☐									

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Secchi	
Secchi Equipment ID:	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID# BRUCE		
Photos: YES	Flow: FLOW	Tide: N/A	Tide Times: High: Low
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SPIO ID Numbers:			
Notes: CLOUDY, SLIGHT WIND			

Duplicate Sample	
Time Zone: EST	Time: Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone: Time: Field ID: (Blank Type_DDMMYYYY)	
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
-------------------------	--------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets BD 03/09/2022 (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2022-03-07-24

Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Katrin Villinger, Brian Davis

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  G2NE1188										
Location: Swimming Pen Creek at Eagles Creek Dr.										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/9/2022	1035 EST	5.96	69.7	23.3	7.17	0.1	0.2 ☒ VOB (S)	0.2	241.4	0.11
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice H2SO4*		☒ <2		
		Lot #			Exp:					
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	B
		Lot #	NA1105100		Exp:	06/07/22				
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #			Syringe Lot #					
BOD (BP-1L)		☐ OV-BODUNIN				☒ Ice				
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHL-F / W-COLOR / W-SO4-IC / W-				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E-Coli ☐ Coli ☐ Enterococci				☒ Ice			1	B
Molecular (QPCR-P-500ML)		☐ CR-HF183 ☐ PCR-BacR ☐ CR-DG3 ☐ CR-GULL2 ☐ CR-GFD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI WE8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-TQ-R W-CAR8-AA (5 ml of MCAA** Buffer Solution) W-TR-SQ-R W-PCB-TQ-R				☐ Ice		MCAA Lot#		
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice				
Name: Katrin Villinger Brian Davis						Signature: 			Date: 03/09/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

Meter Passed End
of Day CCV
Yes / No**WIN Station Name:** Swimming Pen Creek at Eagles Creek Dr.**Date:** 3/9/2022**WIN/Field ID:** G2NE1188**WBID:** 2410**ENR:** -**Latitude:** 30.08903**County:** CLAY**Org ID:** 21FLA**Longitude:** -81.72675**Receiving Body of Water:** Lower St. Johns**RQ:** 2022-03-07-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Katrin Villinger		☒		☒	☒
☒	Brian Davis	☒		☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	

Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID# BRUCE
Photos: YES Flow: FLOW Tide: FALLING	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:	
S&IO ID Numbers:	
Notes:	

Duplicate Sample	
Time Zone: EST Time: Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone: Time: Field ID:	(Blank Type_DDMMYYYY)
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
--------------------------------	---------------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets BD 03/09/2022 Migrate Data to WIN: Verify Data In WIN:



RQ-2022-03-07-24


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Katrin Villinger, Brian Davis

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here							Comments:				
WIN ID/Field ID:  20030667											
Location: LITTLE BLACK CREEK AT SR 220											
Matrix: Fresh							(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTH)	
3/9/2022	1115 EST	3.22	36.5	21.4	6.72	0.3	0.6 ☐ VOB (S)	1.6	179.1	0.08	
	1120 EST	3.65	35.7	21.3	6.69	1.1	Central Lab please use top row for login purposes		179.3	0.08	
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice H2SO4*		<2			
		Lot #		Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☒ Ice NO3*		<2			
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #				Syringe Lot #					
BOD (BP-1L)		☐ OV-BODUNIN				☒ Ice					
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice					
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CLIC / W-FCROR / W-SO4-IC / W-				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	C	
Molecular (QPCR-P-500ML)		☐ CR-HF183 ☐ CR-BacR ☐ CR-DG3 ☐ CR-GULL2 ☐ CR-GFD				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.			
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☒ W-CT-CL-R ☒ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R				☒ Ice		MCAA Lot#	2	C	
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice					
Name: Katrin Villinger Brian Davis					Signature: 					Date: 03/16/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

Meter Passed End

of Day CCV

Yes / No

WIN Station Name: LITTLE BLACK CREEK AT SR 220

Date: 3/9/2022

WIN/Field ID: 20030667

WBID: 2368

ENR: -

Latitude: 30.10508

County: CLAY

Org ID: 21FLA

Longitude: -81.80228

Receiving Body of Water: Lower St. Johns

RQ: 2022-03-07-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Katrin Villinger		☒		☒	☒
☒	Brian Davis	☒		☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	

Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# BRUCE	
Photos: YES	Flow: FLOW	Tide: RISING	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes:			

Duplicate Sample	
Time Zone: EST	Time: Field ID: N/A (WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone: Time: Field ID: (Blank Type_DDMMYYYY)	
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
-------------------------	--------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets BD 03/09/2022 (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2022-03-07-24

Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Katrin Villinger, Brian Davis

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  20030629										
Location: LITTLE BLACK CR AT SR 21										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/9/2022	1155 EST	2.17	24.7	23.2	6.81	0.3	0.3 ☐ VOB (S)	0.7	190.5	0.09
	EST						Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice H2SO4*		<2		
		Lot #		Exp:						
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☒ Ice HNO3*		<2		
		Lot #		Exp:						
Filtered Nutrients (P125ML)		☐ W-PC4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #				Syringe Lot #				
BOD (BP-1L)		☐ OV-BODUNIN				☒ Ice				
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-FCO / W-SO4-IC / W-				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	C
Molecular (QPCR-P-500ML)		☐ CR-HF183 ☐ PCR-BacR ☐ CR-DG3 ☐ CR-GULL2 ☐ PCR-GFD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		☒ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☒ W-CT-CL-R ☒ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☒ W-PCB-TQ-R				☒ Ice		MCAA Lot#	4	C
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice				
Name: Katrin Villinger Brian Davis					Signature: 				Date: 03/09/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

Meter Passed End

of Day CCV

Yes / No

WIN Station Name: LITTLE BLACK CR AT SR 21

Date: 3/9/2022

WIN/Field ID: 20030629

WBID: 2368

ENR: -

Latitude: 30.11805

County: CLAY

Org ID: 21FLA

Longitude: -81.80842

Receiving Body of Water: Lower St. Johns

RQ- 2022-03-07-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Katrin Villinger		x		x	x
☒	Brian Davis	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	

Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID#
	BRUCE

Photos: YES	Flow: FLOW	Tide:	Tide Times:
			High: Low:

Biological Samples Collected:	☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:
-------------------------------	---

SBIO ID Numbers:

Notes:
CLOUDY, RAINY.

Duplicate Sample	
Time Zone: EST Time: Field ID: N/A	(WIN ID_DUP)

WIN DMT Activity ID:

Blank	
Time Zone: Time: Field ID:	(Blank Type_DDMMYYYY)
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
-------------------------	--------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets BD 03/09/2022 (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2022-03-07-24


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Brian Davis, Katrin Villinger

Field Report Prepared By: Brian Davis

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID:  20030571											
Location: PETERS CREEK @ ROSEMARY HILL RD											
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)	
3/9/2022	1410 EST	8.24	96.4	23.1	6.97	0.15	0.3 ☒ VOB (S)	0.3	78.9	0.04	
	EST						Central Lab please use top row for login purposes				
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice H2SO4*		<2			
		Lot #		Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☒ Ice NO3*		<2			
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #				Syringe Lot #					
BOD (BP-1L)		☐ OV-BODUNIN				☒ Ice					
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice					
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☒ W-FW-QLDC				Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	D	
Molecular (QPCR-P-500ML)		☒ CR-HF183 ☒ CR-BacR ☒ CR-DG3				☒ Ice			1	D	
		☐ CR-GULL2 ☐ CR-GFD									
Tracers (BG-500ML)		☒ W-E8321-DI ☒ W-E8321-MS				☒ Ice		MCAA Exp.	1	D	
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI WE8321-MS W-PSNP-TQ W-CT-CL-R				☐ Ice		MCAA Lot#			
		W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-TR-SQ-R									
		W-PCB-TQ-R									
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice					
Name: Brian Davis Katrin Villinger					Signature: 					Date: 03/09/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

**Meter Passed End
of Day CCV
Yes / No****WIN Station Name:** PETERS CREEK @ ROSEMARY HILL RD**Date:** 3/9/2022**WIN/Field ID:** 20030571**WBID:** 2444**ENR:** -**Latitude:** 30.00749**County:** CLAY**Org ID:** 21FLA**Longitude:** -81.74817**Receiving Body of Water:** Lower St. Johns**RQ:** 2022-03-07-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Brian Davis		☒		☒	☒
☒	Katrin Villinger		☒		☒	☒
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	

Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID# BRUCE
Photos: YES Flow: FLOW Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:	
SBIO ID Numbers:	
Notes: SLIGHT RAIN, CLOUDY, VERY WINDY	

Duplicate Sample	
Time Zone: EST Time: Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone: Time: Field ID:	(Blank Type_DDMMYY)
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
--------------------------------	---------------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

(Initials/Date)

(Initials/Date)

Scan/Save Field Sheets BD 03/09/2022

(Initials/Date)

Migrate Data to WIN:

(Initials/Date)

Verify Data In WIN:

(Initials/Date)



RQ-2022-03-07-24

Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Katrin Villinger, Brian Davis

Field Report Prepared By: Katrin Villinger

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  20030480										
Location: BULL CR SW CROSSING CR 215										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
3/9/2022	1320 EST	5.31	59.5	20.9	6.55	0.3	0.7 ☐ VOB (S)	1.2	77	0.04
	EST						Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☐ 2504*		<2		
		Lot #				Exp:				
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☒ Ice ☐ HNO3*		<2		
		Lot #				Exp:				
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
		Filter Lot #				Syringe Lot #				
BOD (BP-1L)		☐ OV-BODUNIN				☒ Ice				
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh) (P-1L)		☒ Turbidity / W-F / W-TSS / TURBIDITY / W-FCIF / W-FCIF-OR / W-SO44C / W-				☐ Ice				
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E-Coli ☐ Coli ☐ Enterococci				☒ Ice			1	E
Molecular (QPCR-P-500ML)		☐ CR-HF183 ☐ PCR-BacR ☐ CR-DG3 ☐ CR-GULL2 ☐ CR-GFD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R				☐ Ice		MCAA Lot#		
Phytoplankton/Periphyton		☐ DTY-QN-C ☐ PKD-QN-C				☐ Ice				
Name: Katrin Villinger Brian Davis					Signature: 				Date: 03/09/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: BULL CR SW CROSSING CR 215**Date:** 3/9/2022**WIN/Field ID:** 20030480**WBID:** 2446**ENR:** =**Latitude:** 30.02051**County:** CLAY**Org ID:** 21FLA**Longitude:** -81.9197**Receiving Body of Water:** Lower St. Johns**RQ:** 2022-03-07-24

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Katrin Villinger		☒		☒	☒
☒	Brian Davis	☒		☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☐ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☒ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded	Meter ID# BRUCE
Photos: YES Flow: FLOW Tide:	Tide Times: High: Low
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:	
SPIO ID Numbers:	
Notes: SUNNY, CLEAR, PARTLY CLOUDY	

Duplicate Sample	
Time Zone: EST Time: Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:	

Blank	
Time Zone: Time: Field ID:	(Blank Type_DDMMYYYY)
DI Source: Location:	
DI Type: Equipment ID:	

DI Container (Name/ID):	DI Container Type:
--------------------------------	---------------------------

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets BD 03/09/2022 Migrate Data to WIN: Verify Data In WIN:

Date of Request: 04-FEB-2022
 Created By: PARRISH_J On: 04-FEB-2022
 Modified By: JASROTIA_P On: 16-FEB-2022
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: LSJR South West 2022

Send Coolers To:

Phone: 904-256-1700
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: S
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 16-FEB-2022
 Biology Request Reviewed By: JASROTIA_P On: 16-FEB-2022
 Sampling Kit Required: YES Ship on: 25-FEB-2022
 Sampling Kit Shipped: YES On: 23-FEB-2022
 Sampling Kit Packed By: ROBERTS_Z
 Preservatives Needed: NO
 Date to Receive Samples: 07-MAR-2022
 Logged In By:

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf.Fresh) Nutrient, Bacteria
 Group B: (Surf.Fresh) Metals, Bacteria
 Group C: (Surf.fresh) CTCLR,PCLTQR,PCBTQR Bacteria
 Group D: (Surf.Fresh) Bacteria, Tracer, PCR
 Group E: (Surf.fresh) Bacteria

Packing Notes:**Bottle Group A (Water) With 2 Samples:**

Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 2	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WO-THI	Number of Bottles: 2	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L

Bottle Group A (Water) With 2 Samples:

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H2SO4
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340B) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Calcium
Iron
Magnesium
Potassium
Sodium
Strontium
Tin
Titanium
Vanadium
Zinc

W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
---------	-------------------	---

*Boron
Aluminum
Antimony
Arsenic
Barium
Beryllium
Cadmium
Chromium
Cobalt
Copper
Lead
Manganese
Molybdenum
Nickel
Selenium
Silver
Thallium

Bottle Group C (Water) With 2 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 2	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Type: BG-500ML FILL TO	Number of Bottles: 6	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

Bottle Group D (Water) With 1 Samples:

Bottle Type: QPCR-P-500ML	Number of Bottles: 1	Preserved With: ICE
PCR-BACR	Template: DEFAULT	(SOP-PCR-1.4) Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction
PCR-DG3	Template: DEFAULT	(SOP-PCR-1.5) Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction
PCR-HF183	Template: DEFAULT	(SOP-PCR-1.0) Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Group D (Water) With 1 Samples:

Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

Bottle Group E (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.