

Login Checklist

RQ- 2022 - 04 - 18 - 15

Login Station ID: # 3

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 04/20/2022 / 09:25

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
2.5C				9		X			5225 5545 4960
1.8C				9		X			5225 5545 4970

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes X No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes X No NA
ABS

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA X

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes X No _____

Caps on tight?

Yes X No _____

Sufficient Sample Volume:

Yes X No _____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes X No _____ Init: ABS /

Chlorine detected? (One container checked per Field ID)

Yes _____ No X Init: ABS /

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes X No _____ NA _____ Init: ABS / KK

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes _____ No _____ NA X Init: ABS /

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: ABS /

Date: 04-20-2022

Event contains NCRs (Y/N)? N

Event ID:

NE - DIST - 2022 - 04 - 22 - 01

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IRTG # 3 EXP:10/19/2022
pH Test Strips 0 – 6	LOT#223819AV EXP:8/30/2023
pH Paper 0 – 3	N/A
pH Paper 0 – 13	LOT 207621 EXP:3/15/2024
Chlorine Test Strips	0064 / 000005 abs 0070 / 2023 / 02

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 4/19/2022 **Collected By:** Jeremy Parrish, Katrin Villinger
Customer: NE-DIST **Sampling Agency:** FDEP
RQ: RQ-2022-04-18-15 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Sixteen Mile Creek at Ashley St.	 20030688
FB	 FB_04192022

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 4/19/2022

1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK

Rec'd/Inspected Date: 04/20/2022

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSE DTC

RQ: 2022-04-18-15

Project: SE-SIXHILL

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

2/3/2022

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J.P. MSH	4/19/22	820	20.7	764	8.9	8.9	100.5	-	1.015	P/F	L/F
ICV	J	4/19/22	821	21.2	764	8.9	8.93	100.5	-	1.015	P/F	L/F
CCV	K. Villinger	↓	1619	21.6	7634	8.8	8.77	99.5	-	1.015	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J.P. MSH	4/19/22	822	220225A	3/23	1000	1000	P/F	L/F
ICV	J	↓	824	220917A	9/22	100	100.8	P/F	L/F
CCV	K. Villinger	↓	1620	211122B	12/22	1413	1421	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J.P. MSH	4/19/22	825	2112B19	12/23	7.	21.9	7.02	-40	P/F	L/F
Calibr.	J	↓	830	220108B	1/24	4.	21.6	4.00	138	P/F	L/F
Calibr.	J	↓	835	2202A63	7/23	10.	21.9	10.04	-211	P/F	L/F
ICV	J	↓	839	220488	1/24	4	21.8	4.08	134	P/F	L/F
CCV	K. Villinger	↓	1621	2202A63	7/23	10.03	21.8	10.06	-244.1	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 171, slope from 4 to 7 178 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J.P. MSH	4/19/22	840	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



RQ-2022-04-18-15


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Jeremy Parrish, Katrin Villinger

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID:  20030688										
Location: Sixteen Mile Creek at Ashley St.										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTH)
4/19/2022	1145 EST	6.3	73.9	23.2	7.27	0.3	0.65 ☐ VOB (S)	0.743	715	0.35
							Central Lab please use top row for login purposes			
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA1105060 Exp: 4/30/22				☒ Ice ☒ H2SO4*		☒ <2	1	A
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD Lot # NA1105100 Exp: 6/7/22				☒ Ice ☒ HNO3*		☐ <2	1	A
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter Filter Lot # R1eb01329				☒ Ice			1	A
Biochemical Oxygen Demand (BP-1L)		BODTEST OV-BODUNIN				☐ Ice				
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHL / W-COLOR / W-SO4-C / W-DS				☒ Ice			1	A
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC Formalin Lot # 8280122TR				☒ Formalin			2	B
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab				☒ Ice			1	A
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GPD				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp. Date		
Pesticides (BG-500ML) (BG-1L)		☒ W-E8321-DI ☒ WE8321-MS ☒ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CAR8-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R ☐ PKD-QN-BV				☒ Ice		MCAA Lot#	4	A
Phytoplankton/Periphyton		DTY-QN-C				☐ Ice				
Name: Jeremy Parrish Katrin Villinger						Signature: 			Date: 04/19/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PLAN FIELD SHEET

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Sixteen Mile Creek at Ashley St. Date: 4/19/2022
WIN/Field ID: 20030688 WBID: 2589 ENR: - Latitude: 29.67
County: ST. JOHNS Org ID: 21FLA Longitude: -81.46
Receiving Body of Water: Deep Creek RQ: 2022-04-18-15

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish		☒		☒	☒
☒	Katrin Villinger		☒	☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☒ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# DESI	
Photos: YES	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☐ None ☐ HA ☒ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: SUNNY / SCI COLLECTION TIME: 1145			

Duplicate Sample	
Time Zone: EST Time: Field ID: N/A (WIN ID_DUP)	
WIN DMT Activity ID:	

Blank	
Time Zone: EST Time: 1150 Field ID: FB 04192022 (Blank Type_DDMMYYYY)	
DI Source: NE ROC Location: Sixteen Mile Creek @ Ashley St.	
DI Type: Field Blank Equipment ID:	
DI Container (Name/ID): Car Man DI Container Type: Carboy	

LIMS EVENT ID: WIN Import ID:
Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:
(Initials/Date) (Initials/Date) (Initials/Date)



RQ- 2022-04-18-15

Customer: WQAP

Project ID: SMP

Collected By: Jeremy Parrish, Katrin Villinger

Field Report Prepared By: Jeremy Parrish, Katrin Villinger

Sampling Agency: FDEP

BLANK



Field ID: FB_04192022

Comments:

Matrix: Field Blank

(Check one) ☐ COMPOSITE ☒ GRAB

Date

Time

DI Source: NE ROC

4/19/2022

1150

EST

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # SA1308030 Exp: 12/1/22	☒ Ice ☒ H2SO4	☒ <2	1	C
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP Lot # NA1286110 Exp: 11/16/22	☒ Ice ☐ H2SO4	☐ <2	1	C
Filtered Nutrients (P125ML)	☒ W-PO4-F Filter ☒ Field Filtered 0.45 µm Filter Lot # r1eb01329 Syringe Lot # Na	☒ Ice		1	C
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-IC / TDS	☒ Ice		1	C
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice	MCAA Exp. Date		
Pesticides (BG-500ML) (BG-1L)	☒ W-E8321-DI ☒ WE8321-MS ☒ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☒ W-TR-SQ-R	☒ Ice	MCAA Lot#	3	C
Name: Jeremy Parrish Katrin Villinger		Signature: 		Date: 04/19/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



DIVERSIFIED
ENVIRONMENTAL LABORATORIES, INC.

CHAIN OF CUSTODY RECORD

Project Number:
(Lab Use Only)

3653 Regent Boulevard, Suite # 509
Jacksonville, FL 32224
P. (904) 807-9625 E. info@deliab.com
F. (904) 807-9627

Page 1 of 1

Company Name Florida Department of Environmental Protection
Address 8800 Baymeadows Way West Ste 100
City Jacksonville State FL Zip 32256
Attn. Jeremy Parrish
Tel. E-mail or Fax DEP-overflowmicro@Florida-DEP.gov
Project Name SCT Sixteen Mile Creek @ Ashley St.
Project Number RG-2022-04-18-15
Project Location

Transfer #	Sample # (Lab Use Only)	Sample Identification	Date Sampled	Time Sampled	Sample Matrix Code	Specific Cond. (Field)	Comp	Grab
1		26030088	4/19/22	1145	NPW	745		✓
2								
3								
4								
5								
6								
7								
8								
9								
0								

Sampler Name/Signature *Katrina Villinger*

E. Coli - Quantitray	Enterococci - Quantitray	Coliform, Fecal - MF 3 Dilutions
----------------------	--------------------------	----------------------------------

Container Information (Qty, Size, & Type)

NPW	NPW	NPW
-----	-----	-----

Container Preservation Codes

6	6	6
---	---	---

Sample Kit or Container Lot # NA

Sample Matrix Codes:

A. Air	P. Petroleum
B. Bulk	S. Soil/Sludge
W. Wipe	O. Other
PW. Potable Water	
NPW. Non-Potable Water	

Preservation Codes:

0. None	5. Na2S2O3
1. H2SO4	6. Crushed Ice (4C)
2. HNO3	7. C6H8O6
3. HCl	8. NH4Cl
4. NaOH	9. Other (List)

Thermal Preservation:

In Crushed Ice? ☒ Yes / ☐ No
Cooler Temp: 0.2 °C
Sample Cont. Temp: 0.8 °C

Turnaround Requested:

☒ Standard (10 Business Days)
☐ Semi-Rush (3-5 Business Days)
☐ Rush (1-2 Business Days)

Turnaround times are estimated & subject to change. Surcharges apply for all Rush samples.

Billing Information:

Invoice with Report
Invoice to Accounting
Req. #:
P.O. #:

Special Instructions:

Transfer Numbers: 1

Relinquished By: Katrina Villinger

Date / Time: 4/19/22 1530

Received By: *DP* Date / Time: 4/19/22 1530

VC



Certification #E821059

Lab General Form 059, Rev. #1, Rev. Date: 02/07/22, Effective Date: 02/07/22.

Visit us online & order kits at: www.deliab.com

Date of Request: 11-MAR-2022
 Created By: PARRISH_J On: 11-MAR-2022
 Modified By: AYRES_J On: 15-APR-2022
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: Calf Creek SCI and Moncrief SCI

Send Coolers To:

Phone: 904-256-1700
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@floridadep.gov

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 15-APR-2022
 Biology Request Reviewed By: AYRES_J On: 15-APR-2022
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped: YES On: 04-APR-2022
 Sampling Kit Packed By: SHAIK_A
 Preservatives Needed: NO
 Date to Receive Samples: 18-APR-2022
 Logged In By: SHAIK_A On: 20-APR-2022

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@floridadep.gov

Comments: Group A: Nutrients, Metals, Tracers, Q-PSNP-TQ, W-TR-SQ-R, Bacteria
 Group B: SCI, Algae_ID
 Group C:(Blank)Nutrient, Metals,Tracer,PSNP TQ, TRSQR

 W-TR-SQ-R (sunscreen)

Wash your hands thoroughly and wear gloves. Refrain from using sunscreens on the day of sample collection.

#####

Bottle Group A (Water) With 2 Samples:

Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 2	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WI-THI	Number of Bottles: 2	Preserved With: ICE
NE-DIV-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by Diversified Environmental Laboratories in Jacksonville
Bottle Type: BG-500ML	Number of Bottles: 2	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP/96%	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO ₃

Bottle Group A (Water) With 2 Samples:

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 2	Preserved With: FILTER-ICE
W-PO4-F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L
Bottle Type: BG-1L	Number of Bottles: 6	Preserved With: ICE
W-PSNP-TQ	Template:	(EPA 8270E) Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS
W-TR-SQ-R	Template: DEFAULT	(EPA 8270E) Analysis of Sunscreen Compounds in Water by GC/MS-SIM

Bottle Group B (Biological) With 2 Samples:

Bottle Type: PJ-2L	Number of Bottles: 4	Preserved With: Formalin
MI-FW-QLDC	Template: DEFAULT	(SOP-IZ06) No. Taxa of FW macroinverts, qual samp, 20 dipnets.
Bottle Type: PT-50ML	Number of Bottles: 2	Preserved With: ICE
ALGAL_ID	Template: DEFAULT	(SOP-AB05) Qualitative assessment of algal taxa present without counts

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: BG-1L	Number of Bottles: 2	Preserved With: ICE
W-PSNP-TQ	Template: DEFAULT	(EPA 8270E) Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS
W-TR-SQ-R	Template: DEFAULT	(EPA 8270E) Analysis of Sunscreen Compounds in Water by GC/MS-SIM

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.

