

RQ-20 22-04-25-09

Login Checklist

Login Station ID: #2

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 4/26/2022 9:23

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #	
		YES	HNO ₃		Formalin	Present?		*Intact?		
						Yes	No	Yes		No
2.1 1.2				12 9		✓			922555461869 3225 5546 1858	

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☐

CONTAINER CHECK

Evidence Tape on Bottles?

Yes ☐ No ☐ NA ☒Evidence Tape intact? Yes ☐ No ☐

If Criminal, photographs taken?

Yes ☒ No ☐Containers intact? Yes ☒ No ☐

Caps on tight?

Yes ☒ No ☐

Sufficient Sample Volume:

Yes ☒ No ☐If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☒ No ☐ Init: IM

Chlorine detected? (One container checked per Field ID)

Yes ☐ No ☒ Init: IM

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ☒ No ☐ NA ☐ Init: IM

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes ☐ No ☒ NA ☐ Init: IM

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: IM

Date: 4/26/2022

Event contains NCRs (Y/N)? ☒

Event ID:

LSJR Main and HA/RPS
NE-DIST-2022-04-26-02 (SMP)Page 1 of 2, 06/15/2020, Form ID-020-122
Date Received: 26-APR-2022 09:23

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #2 exp:10/19/22
pH Test Strips 0 – 6	
pH Paper 0 – 3	
pH Paper 0 – 13	207621 exp: 3/15/24
Chlorine Test Strips	0070 exp: 02/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check (W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 4/25/2022
Customer: NE-DIST
RQ: RQ-2022-04-25-09

Collected By: Jeremy Parrish, Kimberly Appleby
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
BIG POTTSBURG CR BOWDEN RD	 20030068
BIG POTTSBURG CR AT BELFORT RD	 20030064
Miller Creek at Atlantic Blvd.	 G2NE0005
LITTLE POTTSBURG CREEK AT EMERSON	 20030070
FB	 FB04252022
LITTLE POTTSBURG CREEK AT BEDFORD ROAD	 20030756

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 4/25/2022
1600EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:



Rec'd/Inspected Date: 4/26/22

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSI Basic RQ: 2022-04-25-09 Project: SMP LSI ma

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/3/22 2/3/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Pun	4/25/22	830	21.0	765	9.0	9.0	100	-	1.03	P/F	L/F
ICV	J	4/25/22	831	20.9	765	9.0	9.0	100	-	1.03	P/F	L/F
CCV	Kim Appleby	↓	1346	23.9	764.5	8.5	8.49	100.6	-	1.03	P/F	L/F
CCV	↓	↓									P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J Pun	4/25/22	832	270225A	3/23	1000	1000	P/F	L/F
ICV	J	↓	833	210917A	9/22	100	98.3	P/F	L/F
CCV	Kim Appleby	↓	1350	211122B	12/22	1413	1409	P/F	L/F
CCV	↓	↓		211122A	12/22	15000	14996	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Pun	4/25/22	835	21120518	12/23	7.	21.1	7.00	-14.8	P/F	L/F
Calibr.	J	↓	837	2201088	1/24	4.	21.3	4.00	164.8	P/F	L/F
Calibr.	J	↓	840	2202A65	7/23	10.	21.4	10.00	-184	P/F	L/F
ICV	J	↓	842	2201088	1/24	4.	21.3	4.00	160.3	P/F	L/F
CCV	Kim Appleby	↓	1352	2202A63	7/23	10.02	22.7	10.03	-188.20	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
If mV are recorded: slope from 7 to 10 170, slope from 4 to 7 178.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/3/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J Pun	4/25/22	846	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



RQ-2022-04-25-09


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Jeremy Parrish, Kimberly Appleby

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID:  20030070											
Location: LITTLE POTTSBURG CREEK AT EMERSON											
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPTH)	
4/25/2022	1110 EST	4.44	50.4	21.2	7.36	0.15	0.288 ☒ VOB (S)	0.288	537	0.26	
	EST						Central Lab please use top row for login purposes				
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	C	
		Lot # SA1308030		Exp: 12/01/22							
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	C	
		Lot # NA1286110		Exp: 11/16/22							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #		Syringe Lot#							
Biochemical Oxygen Demand (BP-1L)		BODTEST OV-BODUNIN				☐ ICE					
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice			1	C	
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC		Formalin Lot #		Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☐ Enterococci				☐ Ice					
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp. Date			
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R ☐ PKD-QN-BV				☐ Ice		MCAA Lot#			
Phytoplankton/Periphyton		☐ JTY-QN-C				☐ Ice					
Name: Jeremy Parrish Kimberly Appleby					Signature: 					Date: 04/25/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION**

STRATEGIC MONITORING PLAN FIELD SHEET

January 2022

Meter Passed End

of Day CCV

Yes / No

WIN Station Name: LITTLE POTTSBURG CREEK AT EMERSON

Date: 4/25/2022

WIN/Field ID: 20030070

WBID: 2284B

ENR:

Latitude: 30.291527

County: DUVAL

Org ID: 21FLA

Longitude: -81.615111

Receiving Body of Water: Arlington River

RQ- 2022-04-25-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish		☒		☒	☒
☒	Kimberly Appleby	☒		☒		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Secchi	
Secchi Equipment ID: D4	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☒ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# BRUCE	
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: 700 Low: 1310
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: PARTLY CLOUDY.			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone: EST	Time: 1115	Field ID: FB04252022	(Blank Type_DDMMYYYY)
DI Source: NE ROC Lab	Location: LITTLE POTTSBURG CREEK AT EMERSON		
DI Type: Field Blank	i	Equipment ID:	
DI Container (Name/ID): CarNB		DI Container Type: Carboy	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2022-04-25-09

Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Kimberly Appleby, Jeremy Parrish

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID: 											
G2NE0005											
Location: Miller Creek at Atlantic Blvd.											
Matrix: Marine						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)	
4/25/2022	1035 EST	5.33	64.8	23.2	7.11	0.12	0.24 ☒ VOB (S)	0.24	6310	3.46	
							Central Lab please use top row for login purposes				
	EST										
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☐ <2	1	A	
		Lot # SA1308030		Exp:		12/01/22					
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3* ☐ <2					
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #				Syringe Lot #					
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice					
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A	
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☒ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☒ Ice			1	A	
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ Coli ☐ Coli ☐ Enterococci ☐ Contract Lab				☐ Ice					
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.			
Pesticides (BG-500ML) (BG-1L)		☐ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R				☐ Ice		MCAA Lot#			
Phytoplankton/Periphyton		☐ DTY-QN-C ☐ PKD-QN-C				☐ Ice					
Name: Kimberly Appleby Jeremy Parrish					Signature: 					Date: 04/25/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

**Meter Passed End
of Day CCV**
Yes / No**WIN Station Name:** Miller Creek at Atlantic Blvd.**Date:** 4/25/2022**WIN/Field ID:** G2NE0005**WBID:** 2287B**ENR:****Latitude:** 30.305989**County:** DUVAL**Org ID:** 21FLA**Longitude:** -81.628833**Receiving Body of Water:** St Johns River Above Trout River**RQ-** RQ-2022-04-25-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	x		x		
☒	Jeremy Parrish		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☒ Other Bridge	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# BRUCE	
Photos: No	Flow: FLOW	Tide: FALLING	Tide Times: High: 700 Low: 1310
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: PARTLY CLOUDY			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets Migrate Data to WIN: Verify Data In WIN:



RQ-2022-04-25-09

Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Kimberly Appleby, Jeremy Parrish

Field Report Prepared By: Kimberly Appleby

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID:  20030064											
Location: BIG POTTSBURG CR AT BELFORT RD											
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)	
4/25/2022	950 EST	6.42	71.7	20.7	7.4	0.07	0.15 ☒ VOB (S)	0.15	343.5	0.16	
	EST						Central Lab please use top row for login purposes				
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4* ☐ <2					
		Lot #		Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3* ☐ <2					
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #		Syringe Lot #							
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice					
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice					
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ Coli ☐ Coli ☐ Enterococci ☐ Contract Lab				☐ Ice					
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.			
Pesticides (BG-500ML) (BG-1L)		☒ W-E8321-DI ☐ WE8321-MS ☐ W-PSNP-TQ ☒ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☒ W-PCB-TQ-R				☒ Ice		MCAA Lot#	3	E	
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice					
Name: Kimberly Appleby Jeremy Parrish					Signature: 					Date: 04/18/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

**Meter Passed End
of Day CCV
Yes / No****WIN Station Name:** BIG POTTSBURG CR AT BELFORT RD**Date:** 4/25/2022**WIN/Field ID:** 20030064**WBID:** 2265C**ENR:****Latitude:** 30.257619**County:** DUVAL**Org ID:** 21FLA**Longitude:** -81.580408**Receiving Body of Water:** Arlington River**RQ:** 2022-04-25-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Kimberly Appleby	☒		☒		
☒	Jeremy Parrish		☒		☒	
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Secchi	
Secchi Equipment ID: D4	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☒ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# BRUCE	
Photos: No	Flow: FLOW	Tide:	Tide Times: High: Low
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: PARTLY CLOUDY.			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WIN Import ID:

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) QC Station ID, Field Parameters In LIMS DMT: (Initials/Date)

Scan/Save Field Sheets (Initials/Date) Migrate Data to WIN: (Initials/Date) Verify Data In WIN: (Initials/Date)



RQ-2022-04-25-09


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Jeremy Parrish, Kimberly Appleby

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments:					
WIN ID/Field ID:  20030068											
Location: BIG POTTSBURG CR BOWDEN RD											
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE					
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)	
4/25/2022	1010 EST	4.75	53.1	20.9	7.12	0.25	0.533 ☒ VOB (S)	0.533	310	0.15	
	EST						Central Lab please use top row for login purposes				
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group	
Preserved Nutrients (P-500ML)		☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H2SO4* ☐ <2					
		Lot #		Exp:							
Metals (P-500ML)		☐ W-ICPMS ☐ W-ICP ☐ W-HARD				☐ Ice ☐ HNO3* ☐ <2					
		Lot #		Exp:							
Filtered Nutrients (P125ML)		☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice					
		Filter Lot #		Syringe Lot #							
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice					
Chlorophyll (BP-1L)		☐ CHLSUITE-W				☐ Ice					
Physical/Aggregate (Fresh) (P-1L)		☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice					
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice					
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin					
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice					
Microbiology (Contract Lab Bottle)		☐ Coli ☐ Coli ☐ Enterococci ☐ Contract Lab				☐ Ice					
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice					
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.			
Pesticides (BG-500ML) (BG-1L)		☒ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☒ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CARB-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☒ W-PCB-TQ-R				☒ Ice		MCAA Lot#	3	E	
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice					
Name: Jeremy Parrish Kimberly Appleby					Signature: 					Date: 04/25/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PLAN FIELD SHEET

January 2022

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: BIG POTTSBURG CR BOWDEN RD

Date: 4/25/2022

WIN/Field ID: 20030068

WBID: 2265C

ENR:

Latitude: 30.26424

County: DUVAL

Org ID: 21FLA

Longitude: -81.59001

Receiving Body of Water: Arlington River

RQ- 2022-04-25-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish		x		x	
☒	Kimberly Appleby	x		x		
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☐ Normal ☒ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded		Meter ID# BRUCE	
Photos: NO	Flow: FLOW	Tide:	Tide Times: High: Low:
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:			
SBIO ID Numbers:			
Notes: WEATHER: PARTLY CLOUDY			

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: **WIN Import ID:**

Enter Field Parameters, Verify Station ID In LIMS DMT: (Initials/Date) **QC Station ID, Field Parameters In LIMS DMT:** (Initials/Date)

Scan/Save Field Sheets: (Initials/Date) **Migrate Data to WIN:** (Initials/Date) **Verify Data In WIN:** (Initials/Date)



RQ- 2022-04-25-09

Customer: WQAPProject ID: SMPCollected By: Jeremy Parrish, Kimberly ApplebyField Report Prepared By: Jeremy Parrish, Kimberly ApplebySampling Agency: FDEP

BLANK

Field ID: FB04252022

Comments:

Matrix: Field Blank

(Check one) ☐ COMPOSITE ☒ GRAB

Date

Time

DI Source: NE ROC Lab

4/25/2022

1115

EST

Parameter Suite	Parameter Methods	Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP Lot # <u>SA1308030</u> Exp: <u>12/01/22</u>	☒ Ice ☒ H2SO4	☒ <2	1	D
Metals (P-500ML)	☒ W-ICPMS ☒ W-ICP Lot # <u>NA1286110</u> Exp: <u>11/16/22</u>	☒ Ice ☒ H2SO4	☒ <2	1	D
Filtered Nutrients (P125ML)	☐ W-PO4-F Filter ☐ Field Filtered 0.45 µm Filter Lot # Syringe Lot #	☐ Ice			
Physical/Aggregate (Fresh) (P-1L)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-IC / TDS	☒ Ice		1	D
Physical/Aggregate (Marine) (P-1L)	☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY	☐ Ice			
Tracers (BG-500ML)	☐ W-E8321-DI ☐ W-E8321-MS	☐ Ice	MCAA Exp. Date:		
Pesticides (BG-500ML) (BG-1L)	☐ W-E8321-DI ☐ W-E8321-MS ☐ W-PSNP-TQ ☐ W-CT-CL-R ☐ W-PCL-TQ-R ☐ W-CAR-AA (5 ml of MCAA** Buffer Solution) ☐ W-TR-SQ-R ☐ W-PCB-TQ-R	☐ Ice	MCAA Lot#		
Name: <u>Jeremy Parrish</u> <u>Kimberly Appleby</u>		Signature:		Date: <u>04/25/2022</u>	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution



RQ-2022-04-25-09


Customer: NE-DIST

Project ID: STRATEGIC MONITORING PLAN

Collected By: Jeremy Parrish, Kimberly Appleby

Field Report Prepared By: Jeremy Parrish

Sampling Agency: FDEP

Place Label Here						Comments:				
WIN ID/Field ID: 						Microbiology sample submitted to contract laboratory.				
20030756										
Location: LITTLE POTTSBURG CREEK AT BEDFORD ROAD										
Matrix: Fresh						(Check one) ☒ GRAB ☐ COMPOSITE				
Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (umhos/cm)	Salinity (PPTH)
4/25/2022	1140 EST	6.16	69.7	21.4	7.42	0.13	0.266 ☒ VOB (S)	0.266	470.9	0.23
							Central Lab please use top row for login purposes			
	EST									
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☐ H2SO4*		☒ <2	1	B
		Lot # SA1308030		Exp:		12/01/22				
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	B
		Lot # NA1286110		Exp:		11/16/22				
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 um Filter				☒ Ice			1	B
		Filter Lot # R1EB01329				Syringe Lot #				
BOD (BP-1L)		☐ OV-BODUNIN				☐ Ice				
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	B
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice			1	B
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC				☐ Formalin				
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice				
Microbiology (Contract Lab Bottle)		☒ Coli ☐ Coli ☐ Enterococci				☒ Ice			1	B
		☒ Contract Lab								
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3 ☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice				
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice		MCAA Exp.		
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI W-E8321-MS W-PSNP-TQ W-CT-CL-R W-PCL-TQ-R W-CARB-AA (5 ml of MCAA** Buffer Solution) W-TR-SQ-R W-PCB-TQ-R				☐ Ice		MCAA Lot#		
Phytoplankton/Periphyton		DTY-QN-C PKD-QN-C				☐ Ice				
Name: Jeremy Parrish Kimberly Appleby						Signature: 			Date: 04/25/2022	

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

**FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION****STRATEGIC MONITORING PLAN FIELD SHEET**

January 2022

**Meter Passed End
of Day CCV
Yes / No****WIN Station Name:** LITTLE POTTSBURG CREEK AT BEDFORD ROAD**Date:** 4/25/2022**WIN/Field ID:** 20030756**WBID:** 2284B**ENR:****Latitude:** 30.286333**County:** DUVAL**Org ID:** 21FLA**Longitude:** -81.614527**Receiving Body of Water:** Arlington River**RQ:** 2022-04-25-09

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Jeremy Parrish		☒		☒	☒
☒	Kimberly Appleby	☒		☒	☒	
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☐ Syringe	☒ Pole
Depth measured with: Meter	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☒ From Shore	☐ Electric Motor
☐ Other	☐ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# BRUCE							
Photos: No		Flow: FLOW		Tide:		Tide Times: High: Low:							
Biological Samples Collected: ☐ None ☒ HA ☐ SCI ☒ RPS ☐ Algae ID ☒ LVS ☐ LVI Other:													
SBIO ID Numbers:													
Notes: WEATHER: PARTLY CLOUDY													

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDMMYYYY)
DI Source:	Location:		
DI Type:	Equipment ID:		
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: **WIN Import ID:****Enter Field Parameters, Verify Station ID In LIMS DMT:** **QC Station ID, Field Parameters In LIMS DMT:****Scan/Save Field Sheets** **Migrate Data to WIN:** **Verify Data In WIN:**



DIVERSIFIED
ENVIRONMENTAL LABORATORIES, INC.

CHAIN OF CUSTODY RECORD

Project Number:
(Lab Use Only)

3653 Regent Boulevard, Suite # 509
Jacksonville, FL 32224
P. (904) 807-9625 E. info@delilab.com
F. (904) 807-9627

Page 1 of 1

Company Name <u>Florida Department of Environmental Protection</u>										Container Information (Qty, Size, & Type)				Sample Matrix Codes:		
Address <u>8800 Baymeadows Way W</u>										NPW	NPW	NPW				A. Air P. Petroleum
City <u>Jacksonville</u> State <u>FL</u> Zip <u>32256</u>																B. Bulk S. Soil/Sludge
Attn.										Container Preservation Codes				Sample Kit or Container Lot #		
Tel. <u>(904) 256-1709</u> E-mail or Fax <u>Dep-overflowmicro@dep.state.fl.us</u>										6	6	6				
Project Name <u>LSJR Main</u>										Parameters				Preservation Codes:		
Project Number <u>RA-2022-04-25-09</u>										E. Coli - Quantitray	Enterococci - Quantitray	Coliform, Fecal - MF 3 Dilutions				0: None 5. Na2S2O3
Project Location																1. H2SO4 6. Crushed Ice (4C)
Sampler Name/Signature <u>Kim Appleby</u>														2. HNO3 7. C6H8O6		
														3. HCl 8. NH4Cl		
														4. NaOH 9. Other: (List)		
														Thermal Preservation:		
														In Crushed Ice? ☒ Yes / No		
														Cooler Temp: <u>0.5</u> °C		
														Sample Cont. Temp: <u>1.9</u> °C		
														Turnaround Requested:		
														☒ Standard (10 Business Days)		
														☐ Semi-Rush (3-5 Business Days)		
														☐ Rush (1-2 Business Days)		
														Turnaround times are estimated & subject to change. Surcharges apply for all Rush samples.		
														Billing Information:		
														☐ Invoice with Report		
														☐ Invoice to Accounting		
														Req. #:		
														P.O. #:		

Special Instructions:				Transfer Numbers	Relinquished By	Date / Time	Received By	Date / Time
				1	Kim Appleby	4/25/22 1333		04/25/22 1333



Certification #E821059

Date of Request: 14-MAR-2022
 Created By: PARRISH_J On: 14-MAR-2022
 Modified By: JASROTIA_P On: 06-APR-2022
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: LSJR Main and HA/RPS

Send Coolers To:

Phone: 904-256-1693
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: S
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 06-APR-2022
 Biology Request Reviewed By: JASROTIA_P On: 06-APR-2022
 Sampling Kit Required: YES Ship on: 19-APR-2022
 Sampling Kit Shipped: YES On: 18-APR-2022
 Sampling Kit Packed By: ROBERTS_Z
 Preservatives Needed: NO
 Date to Receive Samples: 25-APR-2022
 Logged In By:

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A:(Surf.-Salt) Nutrients
 Group B:(Surf.Fresh)Nutrients, Metals, Ecoli
 Group C:(Surf.Fresh) Nutrients,Metals
 Group D:(blank)Nutrients, Metals
 Group E:(Surf.Fresh)ct-clr,pcltqr,pcbtqr

Packing Notes:**Bottle Group A (Water) With 1 Samples:**

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group C (Water) With 1 Samples:

Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO3/L
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2011) Total Dissolved Solids in aqueous matrices

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
*Boron		
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group E (Water) With 2 Samples:

Bottle Type: BG-500ML FILL TO	Number of Bottles: 6	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.