

Login Checklist

RQ- **2022 - 08 - 29 - 13**

Login Station ID: **# 2**

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: **08/30/2022 / 09:34**

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
2.4C				14		X			5225 5548 4425

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO₃ and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes **X** No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes **X** No **ABS**

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA **X**

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes **X** No _____

Caps on tight?

Yes **X** No _____

Sufficient Sample Volume:

Yes **X** No _____

If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes **X** No _____ Init: **ABS /**

Chlorine detected? (One container checked per Field ID)

Yes _____ No **X** Init: **ABS /**

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes **X** No _____ NA _____ Init: **ABS /**

W-1-4-DIOX (Green dot on container) preserved to pH<4.0? Yes _____ No _____ NA **X Init: **ABS /**

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: **ABS /**

Date: **08/30/2022**

Event contains NCRs (Y/N)? **N**

Event ID: **NE**
NEW - DIST - 2022 - 08 - 30 - 01

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IRTG # 3 EXP:10/19/2022
pH Test Strips 0 – 6	LOT#223819AV EXP:8/30/2023
pH Paper 0 – 3	N/A
pH Paper 0 – 13	LOT 207621 EXP:3/15/2024
Chlorine Test Strips	0064/10000005 abs 0070 / 2023 / 02

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: ABS

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 8/29/2022 **Collected By:** Jeremy Parrish, Katherine Halbert
Customer: NE-DIST **Sampling Agency:** FDEP
RQ: RQ-2022-08-29-13 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 2 **Delivery Method:** FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Bull Creek Ditch 1 mile upstream of Dead lake	 G2NE1194
Bull Creek East ditch at Water Oak	 G2NE1213
BULL CREEK DITCH downstream of East ditch	 G2NE1212
Bulow CR @ Walter Boardman Ln.	 27010967
ST JOE CANAL @ CLUB HOUSE DR N	 27010209

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 8/29/2022
1630 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: A.SHAIK

Rec'd/Inspected
Date: 08/30/2022 / 09:34



RQ-2022-08-29-13	Collected By: Jeremy Parrish, Katherine Halbert
Customer: FDEP	Field Report Prepared by: Jeremy Parrish, Katherine Halbert
Project ID: SMP	Sampling Agency: FDEP
WIN ID/Field ID:  WIN ID/Field ID: G2NE1194	Comments: Microbiology sample(s) submitted to contract laboratory.

Location:
Location: Bull Creek Ditch 1 mile upstream of Dead lake

Matrix: Fresh (Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
8/29/2022 1015	3.13	40.8	28.3	6.8	0.3	0.6	2.317	621	0.3

Parameter Suite	Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*	<2				
	Lot #		Exp:								
Metals (P-500ML)	W-ICPMS		W-ICP	W-HARD	Ice	HNO3*	<2				
	Lot #		Exp:								
Filtered Nutrients (P125ML)	W-PO4-F	Field Filtered 0.45 um Filter			Ice						
	Filter Lot #	Syringe Lot #									
BOD (BP-1L)	OV-BODUNIN				Ice						
Chlorophyll (BP-1L)	CHLSUITE-W				Ice						
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W- TDS				Ice						
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice						
Macroinvertebrates (PJ-2L)	MI-FW-QLDC	Formalin Lot #			Formalin						
Periphyton (PT-50ML)	ALGAE-ID				Ice						
Microbiology (Contract Lab Bottle)	X	E Coli	F Coli	Enterococci	x	Ice				1	A
	X	Contract Lab									
Molecular (QPCR-P-500ML)	X	PCR-HF183	X	PCR-BacR	X	PCR-DG3	x	Ice	1	A	
		PCR-GULL2		PCR-GFD							
Tracers (BG-500ML)	X	W-E8321-DI	X	W-E8321-MS	x	Ice	1	A			
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI	W-E8321-MS	W-PSNP-TQ			MCAA Exp.	MCAA Lot#	4	A	
	X	W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA Buffer Solution)	X	W-CT-CL-R	x	Ice				
		W-TR-SQ-R	X	W-PCB-TQ-R							
Phytoplankton		DTY-QN-C		PKD-QN-BV		Ice					



RQ-2022-08-29-13	Collected By: Jeremy Parrish, Katherine Halbert
Customer: FDEP	Field Report Prepared by: Jeremy Parrish, Katherine Halbert
Project ID: SMP	Sampling Agency: FDEP
WIN ID/Field ID:  WIN ID/Field ID: G2NE1213	Comments: Microbiology sample(s) submitted to contract laboratory.

Location:

Location: Bull Creek East ditch at Water Oak

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. ©	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
8/29/2022 1100	4.04	50.3	26.3	6.94	0.3	0.3	0.56	790	0.38

Parameter Suite	Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*	<2				
	Lot #		Exp:								
Metals (P-500ML)	W-ICPMS		W-ICP	W-HARD	Ice	HNO3*	<2				
	Lot #		Exp:								
Filtered Nutrients (P125ML)	W-PO4-F	Field Filtered 0.45 µm Filter			Ice						
	Filter Lot #	Syringe Lot #									
BOD (BP-1L)	OV-BODUNIN				Ice						
Chlorophyll (BP-1L)	CHLSUITE-W				Ice						
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W- TDS				Ice						
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice						
Macroinvertebrates (PJ-2L)	MI-FW-QLDC	Formalin Lot #			Formalin						
Periphyton (PT-50ML)	ALGAE-ID				Ice						
Microbiology (Contract Lab Bottle)	X E Coli		F Coli	Enterococci	X Ice					1	C
	Contract Lab										
Molecular (QPCR-P-500ML)	PCR-HF183	PCR-BacR		PCR-DG3	Ice						
	PCR-GULL2	PCR-GFD									
Tracers (BG-500ML)	W-E8321-DI	W-E8321-MS			Ice						
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI	W-E8321-MS		W-PSNP-TQ	X Ice	MCAA Exp.	MCAA Lot#	2	C		
	X W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA Buffer Solution)		X W-CT-CL-R							
	W-TR-SQ-R	X	W-PCB-TQ-R								
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice						



RQ-2022-08-29-13	Collected By: Jeremy Parrish, Katherine Halbert
Customer: FDEP	Field Report Prepared by: Jeremy Parrish, Katherine Halbert
Project ID: SMP	Sampling Agency: FDEP
WIN ID/Field ID:  WIN ID/Field ID: G2NE1212	Comments: Microbiology sample(s) submitted to contract laboratory.

Location:

Location: BULL CREEK DITCH downstream of East ditch

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
8/29/2022 1115	4.81	60.2	26.9	6.94	0.3	0.45	0.924	490	0.23

Parameter Suite	Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*	<2				
	Lot #		Exp:								
Metals (P-500ML)	W-ICPMS		W-ICP	W-HARD	Ice	HNO3*	<2				
	Lot #		Exp:								
Filtered Nutrients (P125ML)	W-PO4-F	Field Filtered 0.45 µm Filter			Ice						
	Filter Lot #	Syringe Lot #									
BOD (BP-1L)	OV-BODUNIN				Ice						
Chlorophyll (BP-1L)	CHL SUITE-W				Ice						
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				Ice						
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice						
Macroinvertebrates (PJ-2L)	MI-FW-QLDC	Formalin Lot #			Formalin						
Periphyton (PT-50ML)	ALGAE-ID				Ice						
Microbiology (Contract Lab Bottle)	X E Coli		F Coli	Enterococci	X Ice					1	B
	Contract Lab										
Molecular (QPCR-P-500ML)	PCR-HF183		PCR-BacR	PCR-DG3	Ice						
	PCR-GULL2		PCR-GFD								
Tracers (BG-500ML)	X W-E8321-DI	X	W-E8321-MS		X Ice		1	B			
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI	W-E8321-MS		W-PSNP-TQ	X Ice	MCAA Exp.	MCAA Lot#	2	B		
	X W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA Buffer Solution)		X W-CT-CL-R							
	W-TR-SQ-R	X	W-PCB-TQ-R								
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice						



RQ-2022-08-29-13	Collected By: Jeremy Parrish, Katherine Halbert
Customer: FDEP	Field Report Prepared by: Jeremy Parrish, Katherine Halbert
Project ID: SMP	Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: 27010967

Comments: Microbiology sample(s) submitted to contract laboratory.

Location:

Location: Bulow CR @ Walter Boardman Ln.

Matrix: Marine

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. ©	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT _h)								
8/29/2022 1235	5.62	81.6	30.1	7.26	0.3	0.628	0.628	27078	16.47								
Parameter Suite		Parameter Methods				Preservation		pH Check		# Bottles sent to Lab	Bottle Group						
Preserved Nutrients (P-500ML)		X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				X	Ice	X	H2SO4*	X	<2	1	D			
		Lot #		SA214160		Exp:		06/02/2023									
Metals (P-500ML)		W-ICPMS		W-ICP		W-HARD		Ice	HNO3*			<2					
		Lot #				Exp:											
Filtered Nutrients (P125ML)		W-PO4-F		Field Filtered 0.45 µm Filter				Ice									
		Filter Lot #		Syringe Lot #													
BOD (BP-1L)		OV-BODUNIN						Ice									
Chlorophyll (BP-1L)		X	CHLSUITE-W											X	Ice	1	D
Physical/Aggregate (Fresh) (P-1L)		Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W- TDS						Ice									
Physical/Aggregate (Marine) (P-1L)		X	Alkalinity / W-F / W-TSS / TURBIDITY											X	Ice	1	D
Macroinvertebrates (PJ-2L)		MI-FW-QLDC		Formalin Lot #				Formalin									
Periphyton (PT-50ML)		ALGAE-ID						Ice									
Microbiology (Contract Lab Bottle)		E Coli		F Coli		X	Enterococci	X						Ice	1	D	
		Contract Lab															
Molecular (QPCR-P-500ML)		PCR-HF183		PCR-BacR		PCR-DG3		Ice									
		PCR-GULL2		PCR-GFD													
Tracers (BG-500ML)		W-E8321-DI		W-E8321-MS				Ice									
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI		W-E8321-MS		W-PSNP-TQ		Ice	MCAA Exp.	MCAA Lot#							
		W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA Buffer Solution)		W-CT-CL-R											
		W-TR-SQ-R		W-PCB-TQ-R													
Phytoplankton		DTY-QN-C		PKD-QN-BV				Ice									



RQ-2022-08-29-13

Customer: FDEP

Project ID: SMP

Collected By: Jeremy Parrish, Katherine Halbert

Field Report Prepared by: Jeremy Parrish,
Katherine Halbert

Sampling Agency: FDEP

WIN ID/Field ID:



WIN ID/Field ID: 27010209

Comments: Microbiology sample(s) submitted to contract laboratory.

Location:

Location: ST JOE CANAL @ CLUB HOUSE DR N

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
8/29/2022 1305	5.15	65	27.2	7.24	0.085	0.19	0.19	515	0.25

Parameter Suite	Parameter Methods				Preservation	pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*		
	Lot #		Exp:				<2	
Metals (P-500ML)	W-ICPMS		W-ICP	W-HARD	Ice	HNO3*		
	Lot #		Exp:				<2	
Filtered Nutrients (P125ML)	W-PO4-F		Field Filtered 0.45 um Filter		Ice			
	Filter Lot #	Syringe Lot #						
BOD (BP-1L)	OV-BODUNIN				Ice			
Chlorophyll (BP-1L)	CHLUSITE-W				Ice			
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				Ice			
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice			
Macroinvertebrates (PJ-2L)	MI-FW-QLDC	Formalin Lot #			Formalin			
Periphyton (PT-50ML)	ALGAE-ID				Ice			
Microbiology (Contract Lab Bottle)	X E Coli		F Coli	Enterococci	X Ice		1	E
	Contract Lab							
Molecular (QPCR-P-500ML)	PCR-HF183		PCR-BacR	PCR-DG3	Ice			
	PCR-GULL2		PCR-GFD					
Tracers (BG-500ML)	W-E8321-DI		W-E8321-MS		Ice			
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI	W-E8321-MS	W-PSNP-TQ		Ice	MCAA Exp.	MCAA Lot#	
	W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA Buffer Solution)	W-CT-CL-R					
	W-TR-SQ-R	W-PCB-TQ-R						
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice			

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page

Meter ID: YSF 0252

RQ: 2022-08-29-13

Project: LSUR Palm Coast

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/11/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/29/22	730	21.8	760	8.7	8.7	100	—	1.022	P/F	L/F
ICV	J	8/29/22	731	21.7	760	8.7	8.79	100.1	—	1.022	P/F	L/F
CCV	JPamsh	8/29/22	1456	23.0	761	8.6	8.56	99.8	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/29/22	732	220225A	3/23	10000	10000	P/F	L/F
ICV	J	8/29/22	733	210911A	9/22	100	98.4	P/F	L/F
CCV	JPamsh	8/29/22	1459	220225B	3/23	50000	49233	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	JPamsh	8/29/22	735	MO74-20	3/24	7.	22.2	7.01	-17.5	P/F	L/F
Calibr.	J		737	2205F96	5/24	4.	21.9	4.00	158.0	P/F	L/F
Calibr.	J		739	2202F01	8/23	10.	22.0	10.00	196	P/F	L/F
ICV	J		742	2205F96	5/24	4.	21.9	4.05	154.5	P/F	L/F
CCV	JPamsh	8/29/22	1500	2205F01	5/24	10	22.5	10.03	-159	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 168.5, slope from 4 to 7 184.5 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/11/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air	JPamsh	8/29/22	745	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CHAIN OF CUSTODY RECORD

Project Number:

(Lab Use Only)



3653 Regent Boulevard, Suite # 509
Jacksonville, FL 32224
P. (904) 807-9625 E. info@delilab.com
F. (904) 807-9627

Page _____ of _____

Company Name		Florida Department of Environmental Protection						
Address		8800 Baymeadows Way W Ste 100						
City	Jacksonville	State	FL					
Zip	32256							
Attn	Jeremy Parrish	Jeremy.M.Parrish@floridadep.gov						
Tel.		E-mail or Fax	DEP-Overflowmicro@dep.state.fl.us					
Project Name	Palm Coast							
Project Number	RQ-2022-08-29-13							
Project Location	Palm Coast							
Sampler Name/Signature Katherine Halbert <i>Katherine Halbert</i>								
Transfer #	Sample # (Lab Use Only)	Sample Identification	Date Sampled	Time Sampled	Sample Matrix Code	Specific Cond. (Field)	Comp	Grab
1		G2NE1194	8/29/22	1015	NPW	621		X
2		G2NE1213	8/29/22	1100	NPW	790		X
3		G2NE1212	8/29/22	1115	NPW	490		X
4		27010967	8/29/22	1335	NPW	21018		X
5		27010209	8/29/22	1305	NPW	515		X
6								
7								
8								
9								
0								

Container Information (Qty, Size, & Type)		Sample Matrix Codes:	
NPW	NPW	A. Air	P. Petroleum
NPW	NPW	B. Bulk	S. Soil/Sediment
		W. Wipe	O. Other
		PW. Potable Water	
		NPW. Non-Potable Water	
Container Preservation Codes		Sample Kit or Container Lot #	
6	6		
Parameters		Preservation Codes:	
		0. None	5. Na2S2O3
		1. H2SO4	6. Crushed Ice (±C)
		2. HNO3	7. C6H8O6
		3. HCl	8. NH4Cl
		4. NaOH	9. Other (List)
		Thermal Preservation:	
		In Crushed Ice?	Yes ☒ No ☐
		Cooler Temp	0.3 °C
		Sample Cont. Temp	1.5 °C
		Turnaround Requested:	
		☒ Standard (10 Business Days)	
		☐ Semi-Rush (3-5 Business Days)	
		☐ Rush (1-2 Business Days)	
		Turnaround times are estimated & subject to change. Surcharges apply for all Rush samples.	
		Billing Information:	
		☐ Invoice with Report	
		☐ Invoice to Accounting	
		Req #:	
		P.O. #:	

Special Instructions:		Relinquished By		Date / Time		Received By		Date / Time	
		<i>Jan C. [Signature]</i>		8/29/22		<i>Negon [Signature]</i>		8/29/22 14:10	
Transfer Numbers									



Date of Request: 29-JUL-2022
 Created By: PARRISH_J On: 29-JUL-2022
 Modified By: JASROTIA_P On: 10-AUG-2022
 Customer: NE-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northeast District
 Event Description: Palm Coast

Send Coolers To:

Phone: 904-256-1700
 8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Cooler Ship EMail: Jeremy.M.Parrish@dep.state.fl.us

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 10-AUG-2022
 Biology Request Reviewed By: JASROTIA_P On: 10-AUG-2022
 Sampling Kit Required: YES Ship on: 19-AUG-2022
 Sampling Kit Shipped: YES On: 12-AUG-2022
 Sampling Kit Packed By: ROBERTS_Z
 Preservatives Needed: YES
 Date to Receive Samples: 29-AUG-2022
 Logged In By: SHAIK_A On: 30-AUG-2022

Send Final Report To:

8800 Baymeadows Way West
 Suite 100
 Jacksonville, FL 32256
 Attn: Jeremy Parrish
 Report Notification EMail: Jeremy.M.Parrish@dep.state.fl.us

Comments: Group A: (Surf-fresh) Bacteria,Tracer,PCR, ctclr,pcltqr,pcbtqr
 site G2NE1213

 Group B:(Surf.fresh)Bacteria, Tracer,ctclr,pcltqr,pcbtqr
 site G2NE1212

 Grp C:(Surf.Fresh)Bacteria,ctclr,pcltqr,pcbtqr
 G2NE1213

 Grp D:(Surf.Salt)Nutrients, bacteria
 27010967

 Grp E:(Surf.fresh)Bacteria

 NE-DIV-ECQ
 NE-DIV-ENQ
 Packing Notes:
 1 box h2S04

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-DIV-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by Diversified Environmental Laboratories in Jacksonville
Bottle Type: QPCR-P-500ML	Number of Bottles: 1	Preserved With: ICE
PCR-BACR	Template: DEFAULT	(SOP-PCR-1.4) Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction
PCR-DG3	Template: DEFAULT	(SOP-PCR-1.5) Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction
PCR-HF183	Template: DEFAULT	(SOP-PCR-1.0) Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction
Bottle Type: BG-500ML FILL TO	Number of Bottles: 4	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

Bottle Group A (Water) With 1 Samples:

Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-DIV-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by Diversified Environmental Laboratories in Jacksonville
Bottle Type: BG-500ML FILL TO	Number of Bottles: 2	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-DIV-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by Diversified Environmental Laboratories in Jacksonville
Bottle Type: BG-500ML FILL TO	Number of Bottles: 2	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

Bottle Group D (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-DIV-ENQ	Template: DEFAULT	(Enterolert/QT) Enterococci by Enterolert/QT by Diversified Environmental Laboratories in Jacksonville
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group E (Water) With 1 Samples:

Bottle Type: P-250ML-WI-THI	Number of Bottles: 1	Preserved With: ICE
NE-DIV-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert/QT by Diversified Environmental Laboratories in Jacksonville

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.