

Chemical Analysis Report

NE-DIST-2022-09-27-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Gainesville 2022**
Request ID: **RQ-2022-09-26-10**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Environmental Consultant

Date Certified: 19-OCT-2022 11:01



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Little Hatchet CR East at SR26

Collection Date/Time: 09/26/2022 10:45

Field ID: G1NE0948

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359666	EPA 200.7 Rev. 4.4	Calcium	24.1		mg/L	P420197	
		Iron	600		ug/L	P420197	
		Magnesium	4.95		mg/L	P420197	
		Potassium	0.60	I	mg/L	P420197	
		Sodium	5.9		mg/L	P420197	
		Strontium	79.9		ug/L	P420197	
		Tin	3.0	U	ug/L	P420197	
		Titanium	0.85	I	ug/L	P420197	
		Vanadium	2.0	U	ug/L	P420197	
		Zinc	5.0	U	ug/L	P420197	
	EPA 200.8 Rev. 5.4	Aluminum	100		ug/L	P420197	
		Antimony	0.084	I	ug/L	P420197	
		Arsenic	0.63		ug/L	P420197	
		Barium	11.2		ug/L	P420197	
		Beryllium	0.015	I	ug/L	P420197	
		Boron	21.5		ug/L	P420197	
		Cadmium	0.020	U	ug/L	P420197	
		Chromium	0.53	I	ug/L	P420197	
		Cobalt	0.197		ug/L	P420197	
		Copper	0.40	U	ug/L	P420197	
		Lead	0.20	U	ug/L	P420197	
		Manganese	17.1		ug/L	P420197	
		Molybdenum	0.19	I	ug/L	P420197	
		Nickel	0.50	U	ug/L	P420197	
		Selenium	0.20	U	ug/L	P420197	
		Silver	0.010	U	ug/L	P420197	
		Thallium	0.10	U	ug/L	P420197	
	SM 2340 B-2011	Hardness	80.5		mg/L	P420197	

Sample Location: Little Hatchet CR West at CR222

Collection Date/Time: 09/26/2022 11:00

Field ID: G1NE0947

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359667	EPA 200.7 Rev. 4.4	Calcium	30.3		mg/L	P420197	
		Iron	870		ug/L	P420197	
		Magnesium	6.04		mg/L	P420197	
		Potassium	0.84	I	mg/L	P420197	
		Sodium	6.2		mg/L	P420197	
		Strontium	189		ug/L	P420197	
		Tin	3.0	U	ug/L	P420197	
		Titanium	1.9	I	ug/L	P420197	
		Vanadium	2.1	I	ug/L	P420197	
		Zinc	5.0	U	ug/L	P420197	
	EPA 200.8 Rev. 5.4	Aluminum	180		ug/L	P420197	
		Antimony	0.11	I	ug/L	P420197	
		Arsenic	0.90		ug/L	P420197	
		Barium	14.2		ug/L	P420197	
		Beryllium	0.026	I	ug/L	P420197	
		Boron	23.1		ug/L	P420197	
		Cadmium	0.020	U	ug/L	P420197	
		Chromium	0.78	I	ug/L	P420197	
		Cobalt	0.147		ug/L	P420197	
		Copper	0.55	I	ug/L	P420197	
		Lead	0.41		ug/L	P420197	
		Manganese	15.5		ug/L	P420197	
		Molybdenum	0.36	I	ug/L	P420197	
		Nickel	0.54	I	ug/L	P420197	
		Selenium	0.20	U	ug/L	P420197	
		Silver	0.010	U	ug/L	P420197	
		Thallium	0.10	U	ug/L	P420197	
	SM 2340 B-2011	Hardness	100		mg/L	P420197	

Sample Location: Upper Little Hatchet at Airport field

Collection Date/Time: 09/26/2022 11:30

Field ID: G1NE0949

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359639	DEP SOP: NU-094-1	Color (true)	140		PCU	P420178	
	EPA 180.1 Rev. 2.0	Turbidity	4.8		NTU	P420183	
	EPA 300.0 Rev. 2.1	Chloride	12	A	mg Cl/L	P420332	
		Sulfate	11	A	mg SO4/L	P420335	
	SM 2320 B-2011	Total Alkalinity	57		mg CaCO3/L	P420554	
	SM 2540 C-2011	TDS	132		mg/L	P420666	
	SM 2540 D-2011	TSS	2	I	mg/L	P420671	
	SM 4500-F- C-2011	Fluoride	0.090	I	mg F/L	P420558	MS
2359641	EPA 350.1 Rev. 2.0	Ammonia-N	0.030		mg N/L	P420516	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.64		mg N/L	P420271	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.12		mg N/L	P420499	
	EPA 365.1 Rev. 2.0	Total-P	0.050		mg P/L	P420287	
	SM 5310 B-2011	Organic Carbon	18		mg C/L	P420538	
2359664	EPA 200.7 Rev. 4.4	Calcium	23.8		mg/L	P420197	
		Iron	990		ug/L	P420197	
		Magnesium	3.64		mg/L	P420197	
		Potassium	1.3		mg/L	P420197	
		Sodium	6.9		mg/L	P420197	
		Strontium	48.2		ug/L	P420197	
		Tin	3.0	U	ug/L	P420197	
		Titanium	2.2	I	ug/L	P420197	
	EPA 200.8 Rev. 5.4	Vanadium	2.6	I	ug/L	P420197	
		Zinc	5.0	U	ug/L	P420197	
		Aluminum	298		ug/L	P420197	
		Antimony	0.085	I	ug/L	P420197	
		Arsenic	1.20		ug/L	P420197	
		Barium	18.1		ug/L	P420197	
		Beryllium	0.049		ug/L	P420197	
		Boron	26.6		ug/L	P420197	
		Cadmium	0.020	U	ug/L	P420197	
		Chromium	1.3	I	ug/L	P420197	
		Cobalt	0.210		ug/L	P420197	
		Copper	0.54	I	ug/L	P420197	
		Lead	0.44		ug/L	P420197	
		Manganese	16.4		ug/L	P420197	
		Molybdenum	0.21	I	ug/L	P420197	
		Nickel	0.76	I	ug/L	P420197	
		Selenium	0.20	U	ug/L	P420197	
		Silver	0.010	U	ug/L	P420197	
		Thallium	0.10	U	ug/L	P420197	
	SM 2340 B-2011	Hardness	74.4		mg/L	P420197	

Ref. Method and Comment:

SM 2320 B-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Bivens Arm lake

Collection Date/Time: 09/26/2022 12:25

Field ID: 20020097

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359640	DEP SOP: NU-094-1	Color (true)	14		PCU	P420181	
	EPA 180.1 Rev. 2.0	Turbidity	35		NTU	P420183	
	EPA 300.0 Rev. 2.1	Chloride	10		mg Cl/L	P420332	
		Sulfate	11		mg SO4/L	P420335	
	SM 2320 B-2011	Total Alkalinity	59		mg CaCO3/L	P420554	
	SM 2540 C-2011	TDS	119		mg/L	P420666	
	SM 2540 D-2011	TSS	22		mg/L	P420671	
	SM 4500-F- C-2011	Fluoride	0.17		mg F/L	P420558	MS
2359642	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P420516	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	3.6		mg N/L	P420273	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P420499	
	EPA 365.1 Rev. 2.0	Total-P	0.13		mg P/L	P420287	
	SM 5310 B-2011	Organic Carbon	15		mg C/L	P420538	

Ref. Method and Comment:

DEP SOP: NU-094-1: Batch relative percent difference is unavailable because of analytical difficulties.

SM 2320 B-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Glen Spring

Collection Date/Time: 09/26/2022 13:05

Field ID: G1NE0952

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359643	DEP SOP: NU-094-1	Color (true)	2.5	U	PCU	P420178	
	EPA 180.1 Rev. 2.0	Turbidity	0.15	I	NTU	P420183	
	EPA 300.0 Rev. 2.1	Chloride	24		mg Cl/L	P420332	
		Sulfate	19		mg SO4/L	P420335	
	SM 2320 B-2011	Total Alkalinity	134		mg CaCO3/L	P420552	
	SM 2540 C-2011	TDS	187		mg/L	P420666	
	SM 2540 D-2011	TSS	2	U	mg/L	P420671	
	SM 4500-F- C-2011	Fluoride	0.30		mg F/L	P420555	
2359644	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P420516	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P420273	
	EPA 353.2 Rev. 2.0	NO2NO3-N	1.4		mg N/L	P420499	
	EPA 365.1 Rev. 2.0	Total-P	0.37		mg P/L	P420287	
	SM 5310 B-2011	Organic Carbon	0.56	I	mg C/L	P420538	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Little Hatchet CR 200, downstream of NE 15th St. **Collection Date/Time: 09/26/2022 11:50**

Field ID: G1NE0950

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2359668	EPA 200.7 Rev. 4.4	Calcium	23.4		mg/L	P420197	
		Iron	3.08E+03		ug/L	P420197	
		Magnesium	3.13		mg/L	P420197	
		Potassium	0.55	I	mg/L	P420197	
		Sodium	5.3		mg/L	P420197	
		Strontium	42.3		ug/L	P420197	
		Tin	3.0	U	ug/L	P420197	
		Titanium	1.6	I	ug/L	P420197	
		Vanadium	2.8	I	ug/L	P420197	
		Zinc	5.0	U	ug/L	P420197	
	EPA 200.8 Rev. 5.4	Aluminum	450		ug/L	P420197	
		Antimony	0.080	I	ug/L	P420197	
		Arsenic	2.12		ug/L	P420197	
		Barium	16.5		ug/L	P420197	
		Beryllium	0.060		ug/L	P420197	
		Boron	21.9		ug/L	P420197	
		Cadmium	0.020	U	ug/L	P420197	
		Chromium	1.4	I	ug/L	P420197	
		Cobalt	0.849		ug/L	P420197	
		Copper	0.40	U	ug/L	P420197	
		Lead	0.26	I	ug/L	P420197	
		Manganese	103		ug/L	P420197	
		Molybdenum	0.21	I	ug/L	P420197	
		Nickel	1.1	I	ug/L	P420197	
		Selenium	0.28	I	ug/L	P420197	
		Silver	0.010	U	ug/L	P420197	
		Thallium	0.10	U	ug/L	P420197	
	SM 2340 B-2011	Hardness	71.3		mg/L	P420197	

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P420178

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: DEP SOP: NU-094-1
Batch ID: P420181

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420183

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P420197

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Strontium	2.0	U	ug/L
Tin	3.0	U	ug/L
Titanium	0.75	U	ug/L
Vanadium	2.0	U	ug/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P420197

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Barium	0.20	U	ug/L
Beryllium	0.010	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Cobalt	0.020	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Manganese	0.10	U	ug/L
Molybdenum	0.15	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420332

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420335

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420516

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420271

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420273

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420499

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420287

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P420552

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2320 B-2011
Batch ID: P420554

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2011
Batch ID: P420666

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P420671

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P420555

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 4500-F- C-2011
Batch ID: P420558

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P420538

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P420178

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	90 - 110

Reference Method: DEP SOP: NU-094-1
Batch ID: P420181

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420183

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	101		P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P420197

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	103		P	85 - 115
Iron	105		P	85 - 115
Magnesium	102		P	85 - 115
Potassium	98.6		P	85 - 115
Sodium	104		P	85 - 115
Strontium	101		P	85 - 115
Tin	99.2		P	85 - 115
Titanium	101		P	85 - 115
Vanadium	99.4		P	85 - 115
Zinc	96.2		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P420197

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	98.3		P	85 - 115
Antimony	101		P	85 - 115
Arsenic	102		P	85 - 115
Barium	102		P	85 - 115
Beryllium	94.3		P	85 - 115
Boron	100		P	85 - 115
Cadmium	99.6		P	85 - 115
Chromium	98.5		P	85 - 115
Cobalt	101		P	85 - 115
Copper	99.3		P	85 - 115
Lead	94.5		P	85 - 115
Manganese	98.8		P	85 - 115
Molybdenum	97.7		P	85 - 115
Nickel	99.6		P	85 - 115
Selenium	100		P	85 - 115
Silver	113		P	85 - 115
Thallium	95.6		P	85 - 115

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420332

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420335

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420516

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420271

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420273

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	109		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420499

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420287

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	107		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P420552

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	94 - 106

Reference Method: SM 2320 B-2011
Batch ID: P420554

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	94 - 106

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500-F- C-2011
Batch ID: P420555

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.2		P	96 - 104

Reference Method: SM 4500-F- C-2011
Batch ID: P420558

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P420538

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P420197

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359311	Calcium	105		P	80 - 120
2359311	Iron	108	107	P/P	80 - 120
2359311	Magnesium	106	105	P/P	80 - 120
2359311	Potassium	103	102	P/P	80 - 120
2359311	Sodium	105		P	80 - 120
2359311	Strontium	104	104	P/P	80 - 120
2359311	Tin	104	103	P/P	80 - 120
2359311	Titanium	108	108	P/P	80 - 120
2359311	Vanadium	105	104	P/P	80 - 120
2359311	Zinc	98.4	98.4	P/P	80 - 120
2359312	Calcium	99.1		P	80 - 120
2359312	Iron	104		P	80 - 120
2359312	Magnesium	99.2		P	80 - 120
2359312	Potassium	98.8		P	80 - 120
2359312	Sodium	101		P	80 - 120
2359312	Strontium	101		P	80 - 120
2359312	Tin	100		P	80 - 120
2359312	Titanium	103		P	80 - 120
2359312	Vanadium	100		P	80 - 120
2359312	Zinc	96.9		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P420197

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359311	Aluminum	98.9	100	P/P	80 - 120
2359311	Antimony	103	103	P/P	80 - 120
2359311	Arsenic	103	103	P/P	80 - 120
2359311	Barium	103	104	P/P	80 - 120
2359311	Beryllium	96.2	93.8	P/P	80 - 120
2359311	Boron	105	107	P/P	80 - 120
2359311	Cadmium	102	99.6	P/P	80 - 120
2359311	Chromium	103	104	P/P	80 - 120
2359311	Cobalt	100	101	P/P	80 - 120
2359311	Copper	101	99.4	P/P	80 - 120
2359311	Lead	95.8	96.3	P/P	80 - 120
2359311	Manganese	102	101	P/P	80 - 120
2359311	Molybdenum	102	102	P/P	80 - 120
2359311	Nickel	101	101	P/P	80 - 120
2359311	Selenium	100	99.5	P/P	80 - 120
2359311	Silver	113	115	P/P	80 - 120
2359311	Thallium	97.0	98.9	P/P	80 - 120
2359312	Aluminum	96.6		P	80 - 120
2359312	Antimony	105		P	80 - 120
2359312	Arsenic	103		P	80 - 120
2359312	Barium	104		P	80 - 120
2359312	Beryllium	92.8		P	80 - 120
2359312	Boron	102		P	80 - 120
2359312	Cadmium	103		P	80 - 120
2359312	Chromium	102		P	80 - 120
2359312	Cobalt	101		P	80 - 120
2359312	Copper	98.5		P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P420197

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359312	Lead	95.2		P	80 - 120
2359312	Manganese	102		P	80 - 120
2359312	Molybdenum	99.5		P	80 - 120
2359312	Nickel	100		P	80 - 120
2359312	Selenium	101		P	80 - 120
2359312	Silver	116		P	80 - 120
2359312	Thallium	97.2		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420332

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359639	Chloride	99.3		P	90 - 110
2359640	Chloride	99.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420335

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359639	Sulfate	96.6		P	90 - 110
2359640	Sulfate	96.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420516

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359644	Ammonia-N	102	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420499

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359623	NO2NO3-N	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420287

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359644	Total-P	105	102	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P420555

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2360064	Fluoride	99.0	97.8	P/P	94 - 104

Reference Method: SM 4500-F- C-2011
Batch ID: P420558

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359458	Fluoride	105	104	F/F	94 - 104

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P420538

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2359642	Organic Carbon	102	103	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P420178

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359622	Color (true)	0.0348	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420183

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359597	Turbidity	7.38	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P420197

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359311	Calcium	2.42	Spike	P	0 - 20
2359311	Iron	1.03	Spike	P	0 - 20
2359311	Magnesium	0.704	Spike	P	0 - 20
2359311	Potassium	0.209	Spike	P	0 - 20
2359311	Sodium	0.394	Spike	P	0 - 20
2359311	Strontium	0.116	Spike	P	0 - 20
2359311	Tin	1.38	Spike	P	0 - 20
2359311	Titanium	0.391	Spike	P	0 - 20
2359311	Vanadium	0.914	Spike	P	0 - 20
2359311	Zinc	0.0564	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P420197

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359311	Aluminum	0.927	Spike	P	0 - 20
2359311	Antimony	0.218	Spike	P	0 - 20
2359311	Arsenic	0.00455	Spike	P	0 - 20
2359311	Barium	0.309	Spike	P	0 - 20
2359311	Beryllium	2.59	Spike	P	0 - 20
2359311	Boron	2.47	Spike	P	0 - 20
2359311	Cadmium	1.93	Spike	P	0 - 20
2359311	Chromium	0.342	Spike	P	0 - 20
2359311	Cobalt	0.376	Spike	P	0 - 20
2359311	Copper	2.07	Spike	P	0 - 20
2359311	Lead	0.569	Spike	P	0 - 20
2359311	Manganese	0.391	Spike	P	0 - 20
2359311	Molybdenum	0.0696	Spike	P	0 - 20
2359311	Nickel	0.570	Spike	P	0 - 20
2359311	Selenium	0.886	Spike	P	0 - 20
2359311	Silver	2.18	Spike	P	0 - 20
2359311	Thallium	1.83	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420332

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359639	Chloride	0.329	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420335

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359639	Sulfate	0.132	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420516

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359644	Ammonia-N	0.974	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420271

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359612	Kjeldahl Nitrogen	4.46	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420273

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359745	Kjeldahl Nitrogen	5.63	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420499

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359623	NO2NO3-N	0.758	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420287

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359644	Total-P	1.21	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P420552

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360064	Total Alkalinity	2.06	Sample	P	0 - 3.8

Quality Assurance Report Precision

Reference Method: SM 2540 C-2011
Batch ID: P420666

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359597	TDS	7.64	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P420555

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360064	Fluoride	0.953	Spike	P	0 - 1.6

Reference Method: SM 4500-F- C-2011
Batch ID: P420558

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359458	Fluoride	0.948	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P420538

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359642	Organic Carbon	0.395	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A114965

Included Lab Sample IDs: 2359639, 2359640, 2359643

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	102	P/P	90 - 110
Color (true)	103	103	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A114967

Included Lab Sample IDs: 2359639, 2359640, 2359643

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A114999

Included Lab Sample IDs: 2359639, 2359640, 2359643

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Chloride	101	102	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A115038

Included Lab Sample IDs: 2359664, 2359666, 2359667, 2359668

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	102	101	P/P	90 - 110
Antimony	101	99.4	P/P	90 - 110
Arsenic	102	101	P/P	90 - 110
Barium	100	100	P/P	90 - 110
Beryllium	97.1	96.8	P/P	90 - 110
Boron	106	102	P/P	90 - 110
Cadmium	101	101	P/P	90 - 110
Chromium	99.4	97.7	P/P	90 - 110
Cobalt	98.8	99.2	P/P	90 - 110
Copper	98.2	98.3	P/P	90 - 110
Lead	95.5	95.1	P/P	90 - 110
Manganese	99.1	97.4	P/P	90 - 110
Molybdenum	99.6	99.3	P/P	90 - 110
Nickel	98.6	98.8	P/P	90 - 110
Selenium	101	100	P/P	90 - 110
Silver	100	100	P/P	90 - 110
Thallium	96.1	95.9	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A115060

Included Lab Sample IDs: 2359664, 2359666, 2359667, 2359668

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	100	103	P/P	90 - 110
Iron	101	103	P/P	90 - 110
Magnesium	98.8	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A115060

Included Lab Sample IDs: 2359664, 2359666, 2359667, 2359668

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Potassium	98.9	98.8	P/P	90 - 110
Sodium	102	103	P/P	90 - 110
Strontium	104	98.9	P/P	90 - 110
Tin	104	99.6	P/P	90 - 110
Titanium	105	99.8	P/P	90 - 110
Vanadium	104	98.9	P/P	90 - 110
Zinc	100	97.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A115077

Included Lab Sample IDs: 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	102	P/P	90 - 110
Total-P	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A115084

Included Lab Sample IDs: 2359641

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.8	99.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A115124

Included Lab Sample IDs: 2359641, 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	105	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A115137

Included Lab Sample IDs: 2359641, 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A115139

Included Lab Sample IDs: 2359641, 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.7	98.0	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A115145

Included Lab Sample IDs: 2359641, 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.8	99.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-2011

Run ID: A115145

Included Lab Sample IDs: 2359641, 2359642, 2359644

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.2	99.4	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A115150

Included Lab Sample IDs: 2359639, 2359640, 2359643

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	95.7	102	P/P	90 - 110
Total Alkalinity	97.7	94.5	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A115150

Included Lab Sample IDs: 2359639, 2359640, 2359643

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	102	P/P	90 - 110
Fluoride	103	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
DEP SOP: NU-094-1	Color (true)	102				0.0348	
	Color (true)	101					
EPA 180.1 Rev. 2.0	Turbidity	101				7.38	
EPA 200.7 Rev. 4.4	Calcium	103	105	99.1			2.42
	Iron	105	108	107	104		1.03
	Magnesium	102	106	105	99.2		0.704
	Potassium	98.6	103	102	98.8		0.209
	Sodium	104	105	101			0.394
	Strontium	101	104	104	101		0.116
	Tin	99.2	104	103	100		1.38
	Titanium	101	108	108	103		0.391
	Vanadium	99.4	105	104	100		0.914
	Zinc	96.2	98.4	98.4	96.9		0.0564
EPA 200.8 Rev. 5.4	Aluminum	98.3	98.9	100	96.6		0.927
	Antimony	101	103	103	105		0.218
	Arsenic	102	103	103	103		0.0045
	Barium	102	103	104	104		0.309
	Beryllium	94.3	96.2	93.8	92.8		2.59
	Boron	100	105	107	102		2.47
	Cadmium	99.6	102	99.6	103		1.93
	Chromium	98.5	103	104	102		0.342
	Cobalt	101	100	101	101		0.376
	Copper	99.3	101	99.4	98.5		2.07
	Lead	94.5	95.2	95.8	96.3		0.569
	Manganese	98.8	102	101	102		0.391
	Molybdenum	97.7	102	102	99.5		0.0696
	Nickel	99.6	101	101	100		0.570
	Selenium	100	100	99.5	101		0.886
	Silver	113	113	115	116		2.18
	Thallium	95.6	97.0	98.9	97.2		1.83
EPA 300.0 Rev. 2.1	Chloride	103	99.3	99.6		0.329	
	Sulfate	102	96.6	96.7		0.132	
EPA 350.1 Rev. 2.0	Ammonia-N	97.6	102	103			0.974
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104					4.46
	Kjeldahl Nitrogen	109					5.63
EPA 353.2 Rev. 2.0	NO2NO3-N	102	100	101			0.758
EPA 365.1 Rev. 2.0	Total-P	107	105	102			1.21
SM 2320 B-2011	Total Alkalinity	100				2.06	
	Total Alkalinity	100					
SM 2540 C-2011	TDS					7.64	
SM 4500-F- C-2011	Fluoride	99.2	99.0	97.8			0.953
	Fluoride	101	105	104			0.948
SM 5310 B-2011	Organic Carbon	98.6	102	103			0.395

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2359639, 2359640, 2359643
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2359639, 2359640, 2359643
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2359664, 2359666, 2359667, 2359668
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2359664, 2359666, 2359667, 2359668

Reference Method Descriptions

Method	Description	Associated Samples
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2359639, 2359640, 2359643
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2359639, 2359640, 2359643
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2359641, 2359642, 2359644
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2359641, 2359642, 2359644
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2359641, 2359642, 2359644
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2359641, 2359642, 2359644
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2359639, 2359640, 2359643
SM 2340 B-2011	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃	2359664, 2359666, 2359667, 2359668
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2359639, 2359640, 2359643
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2359639, 2359640, 2359643
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2359639, 2359640, 2359643
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2359641, 2359642, 2359644

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	09/27/2022			09/27/2022 15:38	Anna M. Plaviak	2359639
	09/27/2022			09/27/2022 15:39	Anna M. Plaviak	2359643
	09/27/2022			09/27/2022 16:02	Anna M. Plaviak	2359640
EPA 180.1 Rev. 2.0	09/27/2022			09/27/2022 15:39	Chandler E Cox	2359639
	09/27/2022			09/27/2022 15:40	Chandler E Cox	2359640
	09/27/2022			09/27/2022 15:43	Chandler E Cox	2359643
EPA 200.7 Rev. 4.4	09/27/2022	09/28/2022 10:45	Megan Whitefoot	10/01/2022 01:50	McKinley C Carter	2359664
	09/27/2022	09/28/2022 10:45	Megan Whitefoot	10/01/2022 01:56	McKinley C Carter	2359666
	09/27/2022	09/28/2022 10:45	Megan Whitefoot	10/01/2022 02:04	McKinley C Carter	2359667
EPA 200.8 Rev. 5.4	09/27/2022	09/28/2022 10:45	Megan Whitefoot	10/01/2022 02:11	McKinley C Carter	2359668
	09/27/2022	09/28/2022 10:45	Megan Whitefoot	09/29/2022 19:41	Emily M Philpot	2359664
	09/27/2022	09/28/2022 10:45	Megan Whitefoot	09/29/2022 19:50	Emily M Philpot	2359666
EPA 300.0 Rev. 2.1	09/27/2022	09/28/2022 10:45	Megan Whitefoot	09/29/2022 19:59	Emily M Philpot	2359667
	09/27/2022			09/29/2022 20:08	Emily M Philpot	2359668
	09/27/2022			09/30/2022 04:03	Anna M. Plaviak	2359639
EPA 350.1 Rev. 2.0	09/27/2022			09/30/2022 04:33	Anna M. Plaviak	2359640
	09/27/2022			09/30/2022 07:49	Anna M. Plaviak	2359643
	09/27/2022			10/05/2022 12:23	Ping Hua	2359644
EPA 351.2 Rev. 2.0	09/27/2022			10/05/2022 12:32	Ping Hua	2359641
	09/27/2022			10/05/2022 12:33	Ping Hua	2359642
	09/27/2022	09/29/2022 11:16	Simonne.V. Calhoun	10/04/2022 15:11	Samantha L Bates	2359641
EPA 353.2 Rev. 2.0	09/27/2022	09/29/2022 11:16	Simonne.V. Calhoun	10/04/2022 15:16	Samantha L Bates	2359642
	09/27/2022	09/29/2022 11:16	Simonne.V. Calhoun	10/04/2022 15:18	Samantha L Bates	2359644
	09/27/2022			10/05/2022 10:00	Beverly Y. Sanders	2359641
EPA 365.1 Rev. 2.0	09/27/2022			10/05/2022 10:02	Beverly Y. Sanders	2359642
	09/27/2022			10/05/2022 11:02	Beverly Y. Sanders	2359644
	09/27/2022	09/30/2022 12:23	Adam P Silver	10/03/2022 10:11	James L Waggeberby	2359644
SM 2320 B-2011	09/27/2022	09/30/2022 12:23	Adam P Silver	10/03/2022 10:24	James L Waggeberby	2359642
	09/27/2022	09/30/2022 12:23	Adam P Silver	10/03/2022 12:13	James L Waggeberby	2359641
	09/27/2022			10/05/2022 19:36	Dale L. Simmons	2359643
SM 2340 B-2011	09/27/2022			10/06/2022 05:23	Dale L. Simmons	2359639
	09/27/2022			10/06/2022 05:35	Dale L. Simmons	2359640
	09/27/2022			10/01/2022 01:50	McKinley C Carter	2359664
SM 2540 C-2011	09/27/2022			10/01/2022 01:56	McKinley C Carter	2359666
	09/27/2022			10/01/2022 02:04	McKinley C Carter	2359667
	09/27/2022			10/01/2022 02:11	McKinley C Carter	2359668
SM 2540 D-2011	09/27/2022			09/30/2022 16:11	Yijie Li	2359639, 2359640, 2359643
SM 4500-F- C-2011	09/27/2022			10/05/2022 19:36	Dale L. Simmons	2359643
	09/27/2022			10/06/2022 05:23	Dale L. Simmons	2359639

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 4500-F- C-2011	09/27/2022			10/06/2022 05:35	Dale L. Simmons	2359640
SM 5310 B-2011	09/27/2022			10/05/2022 04:46	Touraj Touran	2359642
	09/27/2022			10/05/2022 07:48	Touraj Touran	2359641
	09/27/2022			10/05/2022 08:01	Touraj Touran	2359644