

Chemical Analysis Report

NE-DIST-2022-10-06-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **LSJR Lake**
Request ID: **RQ-2022-10-10-09**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 25-OCT-2022 11:38

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line extending from the end of the name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Silver lake Center

Collection Date/Time: 10/05/2022 10:45

Field ID: G2NE1132

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2360720	DEP SOP: NU-094-1	Color (true)	310		PCU	P420579	
	EPA 180.1 Rev. 2.0	Turbidity	1.0		NTU	P420583	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P421101	
		Sulfate	2.6		mg SO4/L	P421104	
	SM 2320 B-2011	Total Alkalinity	0.65	U	mg CaCO3/L	P420770	
	SM 2540 C-2015	TDS	83		mg/L	P421275	
	SM 2540 D-2015	TSS	3	I	mg/L	P420916	
	SM 4500-F- C-2011	Fluoride	0.035	U	mg F/L	P420772	
2360721	EPA 350.1 Rev. 2.0	Ammonia-N	0.21		mg N/L	P420798	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P420653	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.023		mg N/L	P420661	
	EPA 365.1 Rev. 2.0	Total-P	0.054		mg P/L	P420761	
	SM 5310 B-2011	Organic Carbon	24		mg C/L	P420780	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P420579

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420583

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421101

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421104

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420798

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420653

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420661

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420761

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P420770

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2015
Batch ID: P421275

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P420916

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P420772

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P420780

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P420579

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420583

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421101

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421104

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420798

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420653

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420661

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420761

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P420770

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	94 - 106

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2540 C-2015
Batch ID: P421275

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	92.2		P	90 - 110

Reference Method: SM 2540 D-2015
Batch ID: P420916

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	89.0		P	85 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P420772

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	102		P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P420780

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421101

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2360307	Chloride	95.9		P	90 - 110
2360786	Chloride	96.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421104

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2360307	Sulfate	96.8		P	90 - 110
2360786	Sulfate	96.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420661

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2360770	NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420761

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2361337	Total-P	102	103	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P420772

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2361114	Fluoride	102	101	P/P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P420780

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2360686	Organic Carbon	99.2	97.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P420579

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360769	Color (true)	0.119	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420583

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360681	Turbidity	13.1	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421101

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360307	Chloride	0.505	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P421104

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360307	Sulfate	1.17	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420798

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360686	Ammonia-N	0.651	Sample	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420653

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360847	Kjeldahl Nitrogen	4.84	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420661

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360770	NO2NO3-N	0.482	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420761

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2361337	Total-P	0.541	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P420770

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2361114	Total Alkalinity	0.728	Sample	P	0 - 3.8

Reference Method: SM 2540 C-2015
Batch ID: P421275

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360786	TDS	5.41	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P420772

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2361114	Fluoride	0.476	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P420780

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2360686	Organic Carbon	1.09	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A115157

Included Lab Sample IDs: 2360720

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A115159

Included Lab Sample IDs: 2360720

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A115199

Included Lab Sample IDs: 2360721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A115218

Included Lab Sample IDs: 2360721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.4	101	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A115243

Included Lab Sample IDs: 2360720

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.6	99.7	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A115243

Included Lab Sample IDs: 2360720

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A115247

Included Lab Sample IDs: 2360721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.6	98.8	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A115253

Included Lab Sample IDs: 2360721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	96.8	97.4	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A115285

Included Lab Sample IDs: 2360721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A115395

Included Lab Sample IDs: 2360720

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.8	98.5	P/P	90 - 110
Sulfate	99.1	99.7	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
DEP SOP: NU-094-1	Color (true)	100			0.119	
EPA 180.1 Rev. 2.0	Turbidity	100			13.1	
EPA 300.0 Rev. 2.1	Chloride	97.0	95.9	96.5	0.505	
	Sulfate	98.4	96.8	96.0	1.17	
EPA 350.1 Rev. 2.0	Ammonia-N	97.6			0.651	
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.0				4.84
EPA 353.2 Rev. 2.0	NO2NO3-N	102	104	104		0.482
EPA 365.1 Rev. 2.0	Total-P	102	102	103		0.541
SM 2320 B-2011	Total Alkalinity	102			0.728	
SM 2540 C-2015	TDS	92.2			5.41	
SM 2540 D-2015	TSS	89.0				
SM 4500-F- C-2011	Fluoride	102	102	101		0.476
SM 5310 B-2011	Organic Carbon	98.6	99.2	97.4		1.09

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2360720
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2360720
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2360720
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2360720
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2360721
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2360721
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2360721
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2360721
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2360720
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2360720
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2360720
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2360720
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2360721

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	10/06/2022			10/06/2022 14:22	Anna M. Plaviak	2360720
EPA 180.1 Rev. 2.0	10/06/2022			10/06/2022 14:13	Chandler E Cox	2360720
EPA 300.0 Rev. 2.1	10/06/2022			10/18/2022 20:44	Anna M. Plaviak	2360720
EPA 350.1 Rev. 2.0	10/06/2022			10/12/2022 13:59	Ping Hua	2360721
EPA 351.2 Rev. 2.0	10/06/2022	10/10/2022 13:58	Samantha L Bates	10/11/2022 10:20	Samantha L Bates	2360721
EPA 353.2 Rev. 2.0	10/06/2022			10/10/2022 09:53	Beverly Y. Sanders	2360721
EPA 365.1 Rev. 2.0	10/06/2022	10/12/2022 14:20	Simonne.V. Calhoun	10/13/2022 13:51	Simonne.V. Calhoun	2360721
SM 2320 B-2011	10/06/2022			10/12/2022 00:06	Adam P Silver	2360720
SM 2540 C-2015	10/06/2022			10/11/2022 15:26	Yijie Li	2360720
SM 2540 D-2015	10/06/2022			10/11/2022 15:26	Yijie Li	2360720
SM 4500-F- C-2011	10/06/2022			10/12/2022 00:06	Adam P Silver	2360720
SM 5310 B-2011	10/06/2022			10/11/2022 17:59	Khloe J Berg	2360721