

CALIBRATION AND VERIFICATION LOG (DEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Craw Daddy

RQ-

2022-06-06-29

Project:

HAB/SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

6/1/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	6/20/22	0735	23.3	756.7	8.5	8.50	99.6	/	0.956	P/F	L/F
ICV	1	1	0736	23.3	756.9	8.5	8.50	99.6	/	/	P/F	L/F
CCV	UnBates	6/20/22	15:00	26.0	757.3	8.1	8.20	101.0	/	/	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	6/20/22	0745	4203H74	03/24	20,000	20,000	P/F	L/F
ICV	1	1	0748	2112E28	12/23	100	98.2	P/F	L/F
CCV	UnBates	6/20/22	15:04	2112E28	12/23	100	99.0	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	6/20/22	0753	210792	02/23	7.00	23.9	7.00	-24.0	P/F	L/F
Calibr.	1	1	0758	206704	11/22	4.00	23.9	4.00	152.4	P/F	L/F
Calibr.	1	1	0801	207370	03/23	10.01	23.9	10.01	-201.0	P/F	L/F
ICV	1	1	0803	207370	03/23	10.0	24.0	10.01	-201.0	P/F	L/F
CCV	UnBates	6/20/22	1508	206704	11/22	4.00	23.7	4.02	151.2	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 177.0, slope from 4 to 7 176.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 6/1/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: knotheat; Date of last verification: 6/1/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Terry Riordan	6/20/22	0803	0.272	0.274	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Lake Bryan LVI

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Shrimp Trap

RQ: 2022-03-21-23

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/1/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	6/20/22	0732	23.5	756.8	8.5	8.46	99.6	/	1.054	P/F	L/F
ICV	1	1	0733	23.5	756.8	8.5	8.45	99.6	/	/	P/F	L/F
CCV	Terry Riordan	16/4	23.2	757.0	8.5	8.60	100.7	/	/	/	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75. Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO change 1%/A; Steady									
Spec. Cond.	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
FT 1200									
Calibr.	Terry Riordan	6/20/22	0742	4203H74	03/24	20,000	20,000	P/F	L/F
ICV	1	1	0746	2112E28	12/23	100	99.0	P/F	L/F
CCV	Terry Riordan	1	1617	2112E28	12/23	100	97.1	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

Conductivity Acceptance criteria ± 5%											
pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	6/20/22	0752	210792	02/23	7.00	24.0	7.00	-26.4	P/F	L/F
Calibr.	1	1	0756	206704	11/22	4.00	23.9	4.00	150.9	P/F	L/F
Calibr.	1	1	0800	207370	03/23	10.01	24.2	10.01	-203.3	P/F	L/F
ICV	1	1	0803	207370	03/23	10.01	24.2	10.00	-202.7	P/F	L/F
CCV	Terry Riordan	1	1616	206704	11/22	4.00	23.6	4.12	144.0	P/F	L/F
CCV										P/F	L/F

0.55

mV pH 4 Range ±180 ±50;

mV pH 10 Range -180 ±50;

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ; If mV are recorded: slope from 7 to 10 176.9, slope from 4 to 7 177.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 6/1/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: High 4; Date of last verification: 5/31/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Terry Riordan	6/20/22	0803	0.272	0.272	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: