

Chemical Analysis Report

CEN-DIST-2022-03-18-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Run 8**
Request ID: **RQ-2022-02-28-17**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Matthew Bearden

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 11-APR-2022 10:36

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line extending from the top of the "T".

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Addison Creek @ Enchanted Forest Nature Preserve

Collection Date/Time: 03/17/2022 10:18

Field ID: G5CE0112

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2313335	EPA 180.1 Rev. 2.0	Turbidity	19		NTU	P411314	
	SM 2320 B-2011	Total Alkalinity	103		mg CaCO3/L	P411602	
	SM 2540 D-2011	TSS	9.8	IA	mg/L	P411586	
	SM 4500 F-C-2011	Fluoride	0.077	I	mg F/L	P411605	
2313336	EPA 350.1 Rev. 2.0	Ammonia-N	0.20		mg N/L	P411554	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P411768	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.16		mg N/L	P411383	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P411461	
	SM 5310 B-2011	Organic Carbon	18		mg C/L	P411474	
2313337	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.043		mg P/L	P411311	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P411314

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P411554

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P411768

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P411383

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P411461

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P411311

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P411602

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-2011
Batch ID: P411586

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P411605

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P411474

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P411314

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	96.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P411554

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P411768

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	110		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P411383

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P411461

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	99.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P411311

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.3		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P411602

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.4		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P411605

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.2		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P411474

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P411554

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2314156	Ammonia-N	96.8	99.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P411768

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313275	Kjeldahl Nitrogen	97.7	96.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P411383

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313374	NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P411461

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313339	Total-P	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P411311

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313277	O-Phosphate-P	100	100	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P411605

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313867	Fluoride	96.0	94.1	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P411474

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2313160	Organic Carbon	95.3	98.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P411314

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313357	Turbidity	5.80	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P411554

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2314156	Ammonia-N	2.19	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P411768

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313275	Kjeldahl Nitrogen	1.41	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P411383

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313374	NO2NO3-N	0.939	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P411461

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313339	Total-P	0.449	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P411311

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313277	O-Phosphate-P	0.399	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P411602

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313867	Total Alkalinity	0.862	Sample	P	0 - 5.2

Reference Method: SM 4500 F-C-2011
Batch ID: P411605

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313867	Fluoride	0.973	Spike	P	0 - 2.1

Quality Assurance Report Precision

Reference Method: SM 5310 B-2011
Batch ID: P411474

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2313160	Organic Carbon	2.51	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A111033

Included Lab Sample IDs: 2313337

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.6	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A111034

Included Lab Sample IDs: 2313335

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A111075

Included Lab Sample IDs: 2313336

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A111102

Included Lab Sample IDs: 2313336

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.0	95.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A111133

Included Lab Sample IDs: 2313336

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.4	97.8	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A111156

Included Lab Sample IDs: 2313335

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.1	97.3	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A111156

Included Lab Sample IDs: 2313335

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A111157

Included Lab Sample IDs: 2313336

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.6	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A111296

Included Lab Sample IDs: 2313336

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity	96.5				5.80	
EPA 350.1 Rev. 2.0	Ammonia-N	95.8	96.8	99.2			2.19
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	110	97.7	96.1			1.41
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	101	102			0.939
EPA 365.1 Rev. 2.0	Total-P	99.8	102	101			0.449
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.3	100	100			0.399
SM 2320 B-2011	Total Alkalinity	98.4				0.862	
SM 4500 F-C-2011	Fluoride	97.2	96.0	94.1			0.973
SM 5310 B-2011	Organic Carbon	96.0	95.3	98.3			2.51

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2313335
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2313336
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2313336
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2313336
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2313336
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2313337
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2313335
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2313335
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2313335
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2313336

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/18/2022			03/18/2022 16:14	Khloe J Berg	2313335
EPA 350.1 Rev. 2.0	03/18/2022			03/24/2022 13:38	Judith K. Clark	2313336
EPA 351.2 Rev. 2.0	03/18/2022	03/31/2022 11:24	Samantha L Bates	04/04/2022 12:32	Alexandra J Mattheus	2313336
EPA 353.2 Rev. 2.0	03/18/2022			03/22/2022 09:56	Beverly Y. Sanders	2313336
EPA 365.1 Rev. 2.0	03/18/2022	03/23/2022 17:20	Simonne.V. Calhoun	03/25/2022 12:57	Sarah A Nixon	2313336
EPA 365.1 Rev. 2.0 dissolved	03/18/2022			03/18/2022 17:26	James L Waggerby	2313337
SM 2320 B-2011	03/18/2022			03/25/2022 00:13	Dale L. Simmons	2313335
SM 2540 D-2011	03/18/2022			03/23/2022 14:22	Yijie Li	2313335
SM 4500 F-C-2011	03/18/2022			03/25/2022 00:13	Dale L. Simmons	2313335
SM 5310 B-2011	03/18/2022			03/22/2022 22:41	Khloe J Berg	2313336