

Chemical Analysis Report

CEN-DIST-2022-06-30-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Addison Creek retrieval and other Addison**
Request ID: **RQ-2022-06-20-17**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:


FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Matthew Bearden

For additional information please contact

Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 19-JUL-2022 10:26

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the first name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Addison Creek @ Enchanted Forest Nature Preserve

Collection Date/Time: 06/29/2022 10:10

Field ID: G5CE0112

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2338371	EPA 180.1 Rev. 2.0	Turbidity	40		NTU	P416063	
	SM 2320 B-2011	Total Alkalinity	70		mg CaCO3/L	P416322	
	SM 2540 D-2011	TSS	14		mg/L	P416413	
	SM 4500-F- C-2011	Fluoride	0.054	I	mg F/L	P416326	
2338372	EPA 350.1 Rev. 2.0	Ammonia-N	0.17		mg N/L	P416266	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.5		mg N/L	P416067	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.037		mg N/L	P416430	
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P416168	
	SM 5310 B-2011	Organic Carbon	23		mg C/L	P416124	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P416063

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P416266

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P416067

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P416430

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P416168

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P416322

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-2011
Batch ID: P416413

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P416326

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P416124

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P416063

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	98.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P416266

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P416067

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.6		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P416430

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P416168

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P416322

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	94 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P416326

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P416124

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P416266

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338418	Ammonia-N	100	98.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P416067

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338272	Kjeldahl Nitrogen	101	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P416430

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338015	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P416168

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338015	Total-P	104	103	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P416326

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338595	Fluoride	102	102	P/P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P416124

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2338420	Organic Carbon	98.5	99.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P416063

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338380	Turbidity	1.36	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P416266

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338418	Ammonia-N	1.35	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P416067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338272	Kjeldahl Nitrogen	3.00	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P416430

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338015	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P416168

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338015	Total-P	1.37	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P416322

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338595	Total Alkalinity	1.22	Sample	P	0 - 3.8

Reference Method: SM 4500-F- C-2011
Batch ID: P416326

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338595	Fluoride	0.500	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P416124

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2338420	Organic Carbon	0.568	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A113222

Included Lab Sample IDs: 2338371

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: SM 5310 B-2011

Run ID: A113260

Included Lab Sample IDs: 2338372

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	100	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A113312

Included Lab Sample IDs: 2338372

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.0	95.3	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A113332

Included Lab Sample IDs: 2338372

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.8	98.7	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A113365

Included Lab Sample IDs: 2338371

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	98.7	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A113365

Included Lab Sample IDs: 2338371

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A113380

Included Lab Sample IDs: 2338372

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	98.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A113420

Included Lab Sample IDs: 2338372

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	103	103	P/P	90 - 110

Quality Assurance Report

Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity	98.0			1.36	
EPA 350.1 Rev. 2.0	Ammonia-N	95.0	100	98.6		1.35
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.6	101	105		3.00
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	102	102		0.0
EPA 365.1 Rev. 2.0	Total-P	104	104	103		1.37
SM 2320 B-2011	Total Alkalinity	101			1.22	
SM 4500-F- C-2011	Fluoride	101	102	102		0.500
SM 5310 B-2011	Organic Carbon	99.0	98.5	99.4		0.568

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2338371
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2338372
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2338372
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2338372
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2338372
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2338371
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2338371
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2338371
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2338372

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/30/2022			06/30/2022 11:38	Khloe J Berg	2338371
EPA 350.1 Rev. 2.0	06/30/2022			07/07/2022 12:25	Ping Hua	2338372
EPA 351.2 Rev. 2.0	06/30/2022	07/05/2022 13:58	Samantha L Bates	07/06/2022 12:36	Alexandra J Mattheus	2338372
EPA 353.2 Rev. 2.0	06/30/2022			07/11/2022 13:38	Judith K. Clark	2338372
EPA 365.1 Rev. 2.0	06/30/2022	07/07/2022 16:00	Simonne.V. Calhoun	07/08/2022 10:44	James L Waggerby	2338372
SM 2320 B-2011	06/30/2022			07/07/2022 02:07	Dale L. Simmons	2338371
SM 2540 D-2011	06/30/2022			07/05/2022 16:02	Yijie Li	2338371
SM 4500-F- C-2011	06/30/2022			07/07/2022 02:07	Dale L. Simmons	2338371
SM 5310 B-2011	06/30/2022			07/01/2022 20:50	Khloe J Berg	2338372