

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Ves. Van + Lamm

RQ: 2022-01-17-31

Project: Desoto Creek

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/05/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	01/31/22	0740	24.2	768.3	8.5	8.49	101.1		120359	P/F	L/F
ICV	Thomas Baker	01/31/22	0742	24.1	768.3	8.5	8.49	101.2			P/F	L/F
CCV	Thomas Baker	01/31/22	1343	21.9	766.6	8.8	8.86	101.0			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	01/31/22	0743	1108C98	08/23	2500	2500	P/F	L/F
ICV	Thomas Baker	01/31/22	0746	2109K87	09/23	100	98.5	P/F	L/F
CCV	Thomas Baker	01/31/22	101333	1108C98	08/23	2500	2498	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	01/31/22	0749	2108B23	7/12	7.00	24.2	7.01	40.5	P/F	L/F
Calibr.			0751	2111H30	11/23	4.01	24.2	4.00	179.9	P/F	L/F
Calibr.			0754	2110G52	4/123	10.02	24.3	10.01	470.8	P/F	L/F
ICV	Thomas Baker	01/31/22	0755	2111H30	11/23	4.01	24.2	3.98		P/F	L/F
CCV	Thomas Baker	01/31/22	1335	2110G52	4/123	10.01	24.4	10.05		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 -170.3, slope from 4 to 7 180.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: DEGREER ALG. CR 33124 Date of last verification: 08/11/21

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: