

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Pro Max

RQ: 2022-05-09-47

Project: Rookery Bay

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 04/14/22

DO DEP SOP FT 1400	Name	Date	Time CT-ET	Temp °C	Baro meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	05/09/22	0656	25.7	755.1	8.2	8.1	99.4		1.0584	(P) / F	(L) / F
ICV	Thomas Baker	05/09/22	0658	25.7	755.1	8.2	8.1	99.4			(P) / F	(L) / F
CCV	William Miller	05/09/22	1529	24.5	756.0	8.3	8.29	99.6			(P) / F	(L) / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	05/09/22	0702	1107981	7/23	10000	10000	(P) / F	(L) / F
ICV	Thomas Baker	05/09/22	0704	220309	3/23	50000	49900	(P) / F	(L) / F
CCV	William Miller	05/09/22	1536	1107981	7/23	10,000	10,200	(P) / F	(L) / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	05/09/22	0708	1112468	11/23	7.00	25.4	7.00	-12.7	(P) / F	(L) / F
Calibr.	Thomas Baker	05/09/22	0711	2203624	03/24	4.01	25.4	4.01	188.3	(P) / F	(L) / F
Calibr.	Thomas Baker	05/09/22	0714	4202048	7/23	10.01	25.1	10.00	-187	(P) / F	(L) / F
ICV	Thomas Baker	05/09/22	0716	2203624	3/24	4.01	25.2	4.00		(P) / F	(L) / F
CCV	William Miller	05/09/22	1541	4202048	7/23	10.02	23.8	10.02		(P) / F	(L) / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: NEEPBLUE Date of last verification: 02/24/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: