

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

My Boy Blue

RQ:

2022-08-01-70

Project:

SMP: Cont. Haddett

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 07/13/22

DO FDEP SOP FT 1400	Name	Date	Time CET	Temp °C	Baro meter mmHg	DO Gain mg/L	Meter DO mg/L	% DO	Probe Charge	Probe Gain	Pass	Lab
Calibr.	<u>William Miller</u>	<u>08/02/22</u>	<u>0708</u>	<u>26.5</u>	<u>763.4</u>	<u>8.1</u>	<u>8.10</u>	<u>100.7</u>		<u>1.3377</u>	<u>P/F</u>	<u>L/F</u>
ICV	<u>William Miller</u>	<u>08/02/22</u>	<u>0711</u>	<u>26.5</u>	<u>763.6</u>	<u>8.1</u>	<u>8.17</u>	<u>101.7</u>			<u>P/F</u>	<u>L/F</u>
CCV	<u>William Miller</u>	<u>08/18/22</u>	<u>1502</u>	<u>23.2</u>	<u>759.9</u>	<u>8.5</u>	<u>9.96</u>	<u>116.6</u>			<u>P/F</u>	<u>L/F</u>
CCV											<u>P/F</u>	<u>L/F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Conductivity FDEP SOP	Name	Date	Time (CET)	Lot #	Exp Date	Standard mhos/cm	Meter mhos/cm	Pass	Lab
Calibr.	William Miller	08/02/22	0714	220309	03/23	50,000	50,000	P/F	L/F
ICV	William Miller	08/02/22		1107981	07/23	10,000	10,000	P/F	L/F
CCV	Gray Suh	08/17/22	1129	2204020	4/24	100	98.1	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

PH FDEP SOP FT 1400	Name	Date	Time CET	Temp °C	Temp °C	pH Buffer SU	Temp °C	pH Buffer SU	Temp °C	Pass	Lab
Calibr.	William Miller	08/02/22	0720	22.02637	02/24	7.00	23.4	7.00	-428	P/F	L/F
Calibr.	William Miller	08/02/22	0728	22.03624	03/24	4.01	22.9	4.01	136.9	P/F	L/F
Calibr.	William Miller	08/02/22	0733	22.02465	07/23	10.03	22.8	10.03	-212.5	P/F	L/F
ICV	William Miller	08/02/22	0737	22.03624	03/24	4.01	22.8	4.00		P/F	L/F
CCV	Gray Suh	8/17/22	0733	22.02465	7/23	10.0	23.9	9.96		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.7, slope from 4 to 7 179.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES (NO) / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 3312B; Date of last verification: 02/24/22

Depth Sensor Daily Calibration (mV)	Name	Date	Time CET	Calibrated Value (0.00 at 0 m) meters	ICV Value meters	Pass	Lab
Pressure mode in air						<u>P/F</u>	<u>L/F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: 1 quality DO