

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: FRIDA

RQ- 2022-10-03-63

Project: SMP L&B CRICKS

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/10/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	10/24/22	0720	22.0	758.8	8.7	8.73	99.8	-	1.06	P/F	L/F
ICV	↓	↓	0722	22.0	758.7	8.7	8.73	99.8	-	-	P/F	L/F
CCV	William Miller	10/24/22	1322	24.0	758.3	8.4	8.46	100.4			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	10/24/22	0724	2201A88	7/24	2500	2500	P/F	D/F
ICV	↓	↓	0726	2112628	12/23	100	1000	P/F	D/F
CCV	William Miller	10/24/22	1324	220309	03/23	50,000	49,800	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	BARRINGTON	10/24/22	0728	2205654	5/24	7. -	21.6	7.02	-32.1	P/F	D/F
Calibr.	↓	↓	0730	2206A61	5/24	4. -	21.3	4.00	144.8	P/F	D/F
Calibr.	↓	↓	0732	2204172	10/23	10. -	21.6	10.04	-199.9	P/F	D/F
ICV	↓	↓	0734	2206A61	5/24	4. -	21.3	4.04	-	P/F	D/F
CCV	William Miller	10/24/22	1327	2204172	10/23	10.04	21.8	10.01		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 174.8, slope from 4 to 7 180.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/29/22

If NO, what will be used? (circle one) Secchi-Disk Line / Sonar Unique ID: 3312B; Date of last verification: 8/29/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: