

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: IPresidentCaracho

RQ: 2022-12-03-37

Project: Myakka

12-2022-12-12-23

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/10/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	12/13/22	0705	23.2	762.4	8.6	8.58	100.4		1.021	P/F	L/F
ICV	Thomas Baker	12/13/22	0706	23.2	762.5	8.6	8.57	100.3			P/F	L/F
CCV	Monica Jaeger	12/13/22	1445	22.3	761.3	8.7	8.67	99.8			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	12/13/22	0714	2208AS8	7/24	2500	2500	P/F	L/F
ICV	Thomas Baker	12/13/22	0716	2112E28	12/23	100	98.5	P/F	L/F
CCV	Monica Jaeger	12/13/22	1447	2208AS8	7/24	2500	2500	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Thomas Baker	12/13/22	0714	2205E54	5/24	7.01	22.5	7.01	—	P/F	L/F
Calibr.			0721	2206A61	5/24	4.00	23.0	4.00	—	P/F	L/F
Calibr.			0722	4208J57	2/24	10.03	22.8	10.03	—	P/F	L/F
ICV	Thomas Baker	12/13/22	0723	2206A61	5/24	4.01	23.1	4.03		P/F	L/F
CCV	Monica Jaeger	12/13/22	1451	4208J57	2/24	10.05	21.0	10.05		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **NO** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) **Secchi Disk Line** / Sonar Unique ID: 3312A; Date of last verification: 08/29/21

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

REMARKS: