


BENCHMARK EA SOUTH



NELAC Certification #B85086

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAC STANDARDS

Submission Number : S22040246

Florida Dept Of Env Protection
2600 Blair Stone Rd
Tallahassee, FL 32399

Project Name : RQ-2022-04-11-32
Date Received : 04/11/2022
Time Received : 15:31

Submission Number: 22040246
Sample Number: 001
Sample Description: G3SD0194

Sample Date: 04/11/2022
Sample Time: 14:45
Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
ENTEROCOCCI	414	#/100 ML	10	10	ENTEROLERT	04/11/2022 16:23	04/12/2022 16:58	MM

Melinda Merchant

04/15/2022

Dale D. Dixon / Laboratory Director

Date

Melinda Merchant/ Laboratory Manager

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J6 = Estimated value. Data questionable due to improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis lost or not performed.

Q = Sample held beyond accepted hold time.
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are usable.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.
! = Data deviate from historically established concentration ranges.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
* = Not reported due to interference.

NOTES:

JO = All Quanti-Tray wells were positive. Result may be higher than reported.
PQL = 4xMDL.
ND = Not detected at or above the adjusted reporting limit.
X = Value exceeds MCL.
1 dilution (10 mL) on each Enterococcol and E.coli sample and 3 dilutions (1 mL, 10 mL, & 100 mL) on each Fecal Coliform sample.

COMMENTS:

For questions or comments regarding these results, please contact us at (941) 625-3137.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc.
1001 Corporate Ave. Suite 102
North Port, FL 34289
(941) 625-3137 / (941) 240-3071 fax

BEAS Sample Receipt Temp. 5.7 °C

Checked at Benchmark EA South E85088
With Temperature Gun ID #7

Client Name: FDEP

Address: 2600 Blair Stone Rd.
Tallahassee, FL, 32399

Phone: (239) 344-5719

Email: dep-overflowmicro@dep.state.fl.us

Bill to: Frances Moore
P.O. #: Lab B95341

Project Name: RQ-2021- ²⁰²²⁻ 54-11-32
Project Run/Location: Antebell Creek

Laboratory Submission #: 522040246

Station ID	Surface Water Sample Matrix* (Fresh / Marine) Sample Type	Enterococci E.coli (SM9223B Quantitray)	Laboratory Sample #
		E.Coli = 250 mL ; Entero = 120 mL ; Fecal Coliform = 120 mL	
63SD 01941	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time: 04/11/22 1445	1
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	
	F / M / Grab	E. coli / Entero / Fecal coliform Date & Time:	

*Sample Type is used to indicate whether the sample was a grab (G) or whether it was a composite (C).

*Sample Matrix is used to indicate whether the sample is being discharged to drinking water (DW), ground water (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil sediment (SDA/NT), or sludge (SLDG).

*Container Type is used to indicate whether the container is plastic (P) or glass (G).

Sample must be refrigerated or stored in wet ice after collection. The maximum temperature during storage should be 6°C (42°F). Under "Preservative," list any preservatives that were added to the sample container.

Instructions:

- Each bottle has a label identifying sample ID, preservative (preservative contained in the bottle, sample type, client ID, and parameters for analysis).
- The following information should be added to each bottle label after collection with permanent black ink: date and time of collection, sample's name or initials, and any field number or ID.
- All bottles not containing preservative may be rinsed with appropriate sample prior to collection.
- The client is responsible for documentation of the sampling event. Please note special sampling events on the sample custody form.

Laboratory Sample Acceptability:

pH < 8 BEAS Temperature: 5.7°C

1. Collector: <u>William Miller</u>	Date: <u>04/11/22</u>	Time: <u>1531</u>	Received By: <u>Melinda Merchant</u>	Date: <u>4/11/22</u>	Time: <u>1531</u>
2. Relinquished by:	Date:	Time:	Received By:	Date:	Time:
3. Relinquished by:	Date:	Time:	Received By:	Date:	Time:
4. Relinquished by:	Date:	Time:	Received By:	Date:	Time:
5. Relinquished by:	Date:	Time:	Received By:	Date:	Time:

Page 9 of 9