

From: katie.sanderslabs.net
To: DEP-Overflowmicro
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022
Date: Wednesday, August 17, 2022 11:47:34 AM
Attachments: [image001.png](#)

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

Correct. When we apply dilution factor ie one plate was in range and we use DF to calculate value it is same but when we have add 3 plates together then no DF would apply. Yes report is correct. Thanks!

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 9:00 am - 4:00 pm
Friday: 9:00 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Service is contact-less. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Kit requests are typically 5 days out. We cannot provide kits without prior request. As always our hearts with all of you.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Wednesday, August 17, 2022 11:45 AM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Katie,
Oh, I didn't know that. So the report is correct then? It seems like most of the time the MDL and the dilution factor are the same, but in the case of those two samples it was different.
Thanks,
Lisa

From: katie.sanderslabs.net <katie@sanderslabs.net>
Sent: Wednesday, August 17, 2022 11:15 AM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

DF do not get applied when calculation is done manually otherwise it would double value that has already been calculated. Thanks

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C

239-590-0337 O
Monday - Thursday : 9:00 am - 4:00 pm
Friday: 9:00 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Service is contact-less. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Kit requests are typically 5 days out. We cannot provide kits without prior request. As always our hearts with all of you.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Wednesday, August 17, 2022 11:06 AM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Hi Katie,
Can you please revise the report attached so that all of the fecal coliforms have a dilution factor of 2? As it is right now, only one has the right dilution factor.
Thank you,
Lisa

From: katie.sanderslabs.net <katie@sanderslabs.net>
Sent: Friday, August 12, 2022 9:27 AM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

Thanks have a great weekend

Get [Outlook for iOS](#)

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Friday, August 12, 2022 9:16:41 AM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Hi Katie,

Thank you for clarifying. I went back and read this email chain and I think I understand everything and it makes sense.

Thanks again,
Kara

From: katie.sanderslabs.net <katie@sanderslabs.net>
Sent: Tuesday, August 9, 2022 5:35 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

I think I am confusing myself because there are multiple samples on this email. So subject is 2207668 with picture 3a but then in email body you also mention 2207813 which I had FC inquiring- that report hadn't been finalized until I sent other day. You are correct that is typo on DF on Enterobacter it was 10. I will revise but we are on same sample and page so to speak?

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 9:00 am - 4:00 pm
Friday: 9:00 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Service is contact-less. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Kit requests are typically 5 days out. We cannot provide kits without prior request. As always our hearts with all of you.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: katie sanderslabs.net <katie@sanderslabs.net>
Sent: Tuesday, August 9, 2022 4:11 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

Can you clarify because the coc from your staff for 3a has matrix as fresh and was ran as that.

Get [Outlook for iOS](#)

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Tuesday, August 9, 2022 3:37:27 PM
To: katie sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Hi Katie,

Lisa should have asked about the entero sample not being diluted on report 2207813. It is marine.

I do see that a revision was sent for the fecal though as well??

Thanks,
Kara

From: katie sanderslabs.net <katie@sanderslabs.net>
Sent: Friday, August 5, 2022 1:17 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

after my next interview. Thanks . It should be a 2 if it was combined plate or low value but I'll look at it

Get [Outlook for iOS](#)

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>

Sent: Friday, August 5, 2022 1:15:04 PM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Thanks Katie,
I think we got that one figured out. The dilution factor issue was for Report #2207813. The FC has a dilution factor of 1. The COC says it's a marine sample, so it should be at least a DF of 10.
Thanks,
Lisa

From: katie.sanderslabs.net <katie@sanderslabs.net>
Sent: Friday, August 5, 2022 12:53 PM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

OK I reviewed this. Sample 3 according to COC was marked F for fresh. It was ran 100mL. The EDD was typo I was trying correct FC to 2 and wrote on both lines I must have accidentally highlighted both when I copied. It is MDL 1 and I will adjust EDD to match. Thanks

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 9:00 am - 4:00 pm
Friday: 9:00 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Service is contact-less. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Kit requests are typically 5 days out. We cannot provide kits without prior request. As always our hearts with all of you.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Friday, August 5, 2022 11:20 AM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: RE: Analytical Report: DEP Fecals,2207668,7/18/2022

Katie,
My question (about the MDL) was for the entero sample collected on 7/18/22 (G1SD0125). I'm not sure if we're talking about the same thing:

CLIENT: DEP-Bureau or Labs-Biology Section				Lab Order: 2207668		
Project: DEP Fecals						
Lab ID: 2207668-003		Collection Date: 7/18/2022 12:05:00 PM				
Client Sample ID: G1SD0125		Matrix: SURFACE WATER				
Analyses	Result	MDL	Qual	Units	DF	Date Analyzed
ENTEROCOCCUS				ENTEROLERT		Analyst: KS
Enterococcus	40.5	1.00		MPN/100mL	1	7/18/2022 3:38:00 PM

From: katie.sanderslabs.net <katie@sanderslabs.net>
Sent: Friday, August 5, 2022 10:50 AM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Re: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

Well it was dup so ran 50/50 added together so could be either/or. Plates were 50 mls which is MDL 2 but you can add plates so added its 1. Let me know and I will fix. Thanks

Katie Strothman
Laboratory Director/ QCM
Sanders Laboratories, Inc.
10090 Bavaria Rd
Ft Myers, FL 33913
941-915-7447 C
239-590-0337 O
Monday - Thursday : 9:00 am - 4:00 pm
Friday: 9:00 am - 1:00 pm
Saturday / Sunday: By appointment only
Closed Holidays

COVID-19 update: In order to do our part to reduce the spread we will have Bay/Suite 2 open and main office will be closed. Service is contact-less. Please leave your samples in the black fridge and COC in the bin on the cart. Make a copy of the COC for your records at your facility, get copies from client portal, or take picture with your phone we will not be making on site. Call 239-590-0337 if you need one of our team to assist you with a mask. Call kit pickups in ahead and we will put your bottles in Bay 2 on the cart. Kit requests are typically 5 days out. We cannot provide kits without prior request. As always our hearts with all of you.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

From: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Sent: Friday, August 5, 2022 10:46 AM
To: katie.sanderslabs.net <katie@sanderslabs.net>
Subject: FW: Analytical Report: DEP Fecals,2207668,7/18/2022

Hi Katie,
One question on this..which is the correct MDL for the Entero test for G1SD0125? Report says 1.00, the EDD says 2.

-----Original Message-----

From: Sanders Laboratories <sanderslabsreporting@gmail.com>
Sent: Friday, August 5, 2022 9:29 AM
To: DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Analytical Report: DEP Fecals,2207668,7/18/2022

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

Please note this a NO reply outgoing emailer only. If you have questions in response to this email. Send your email to Katie@sanderslabs.net or Jeff@sanderslabs.net.

Your sample(s) has (have) been tested in accordance with NELAP requirements unless stated otherwise.

Your report has been emailed by electronic format. A hard copy of the report can be obtained by contacting our office for a fee of \$25 per initial report.

Best Regards,

Katie Strothman

Laboratory Director

katie@sanderslabs.net

10090 Bavaria Rd.

Fort Myers, FL 33913

http://secure-web.cisco.com/1S2DFevkBCI_-

[HUTHfDDv5TN2kJF0Tg2Tc66gmxAGKikiTyTrYeSyh2LNBNFetmmCsPjINm5T0kiX1gg1xC1oYC95qeSRxdJf26JosiZ8b55T866mn3PuNE7bWDQsIR-](http://secure-web.cisco.com/1S2DFevkBCI_-HUTHfDDv5TN2kJF0Tg2Tc66gmxAGKikiTyTrYeSyh2LNBNFetmmCsPjINm5T0kiX1gg1xC1oYC95qeSRxdJf26JosiZ8b55T866mn3PuNE7bWDQsIR-6YMoqHGaQtu-GF-2rE-)

[6YMoqHGaQtu-GF-2rE-](http://secure-web.cisco.com/1S2DFevkBCI_-6YMoqHGaQtu-GF-2rE-)

[ntGmAFKj7mBoj28ZFpvQOf6gLBCXxRiSiZWPsHecGO93Uf5Nwi0MhxKkv3St1_pzHYsz6mMhYSJ6JzipMCwMQz4uliR860jaL9wBOrr800u3qrLbNuqq-](http://secure-web.cisco.com/1S2DFevkBCI_-ntGmAFKj7mBoj28ZFpvQOf6gLBCXxRiSiZWPsHecGO93Uf5Nwi0MhxKkv3St1_pzHYsz6mMhYSJ6JzipMCwMQz4uliR860jaL9wBOrr800u3qrLbNuqq-)

[lisLO8AuMsA-zgFnB8X44UqjPwQXQyICKRROmg_N51BefEwRieyCRWaeJmBDTKR1njllE1toEpXw/http%3A%2F%2Fwww.sanderslabs.net](http://secure-web.cisco.com/1S2DFevkBCI_-lisLO8AuMsA-zgFnB8X44UqjPwQXQyICKRROmg_N51BefEwRieyCRWaeJmBDTKR1njllE1toEpXw/http%3A%2F%2Fwww.sanderslabs.net)

Sanders is dedicated to always putting our clients first. We pride ourselves on having quality data, great customer service, and competitive pricing. Sanders always strives to match or beat any local competitor's pricing. As always thanks for choosing Sanders Laboratories.

The information contained in this email is privileged and confidential information, intended only for the use of the individual (s) distributed to. If the reader of this message is not the intended recipient, you are hereby notified that any distribution or copying of this is strictly prohibited. If you have received this in error, please notify promptly.

[Dep Customer Survey]<<http://survey.dep.state.fl.us/?refemail=DEP-Overflowmicro@dep.state.fl.us>>

[Redacted]

[Redacted]

[Redacted]
[Redacted]
[Redacted]
[Redacted]