


BENCHMARK EA SOUTH



NELAC Certification #E85086

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAC STANDARDS

Submission Number : S22120651

Florida Dept Of Env Protection
2600 Blair Stone Rd
Tallahassee, FL 32399

Project Name : RQ-2022-12-05-51
Date Received : 12/21/2022
Time Received : 16:55

Submission Number: 22120651
Sample Number: 001
Sample Description: G4SD0216

Sample Date: 12/21/2022
Sample Time: 09:30
Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	393	#/100 ML	10	10	SM9223B	12/21/2022 16:27	12/22/2022 16:46	BK/MM

Melinda Merchant

12/27/2022

Dale D. Dixon / Laboratory Director

Date

Melinda Merchant / Laboratory Manager

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J5 = Estimated value. Data questionable due to improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis test not performed.

Q = Sample held beyond accepted hold time.
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are usable.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.
! = Data deviate from historically established concentration ranges.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
* = Not reported due to interference.

NOTES:

JO = All Quanti-Tray wells were positive. Result may be higher than reported.
PQL = 4xMDL.
ND = Not detected at or above the adjusted reporting limit.
X = Value exceeds MCL.
1 dilution (10 mL) on each Enterococci and E.coli sample and 3 dilutions (1 mL, 10 mL, & 100 mL) on each Fecal Coliform sample.

COMMENTS:

For questions or comments regarding these results, please contact us at (941) 625-3137.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc.
1001 Corporate Ave. Suite 102
North Port, FL 34289
(941) 625-3137 / (941) 240-3071 fax

BEAS Sample Receipt Temp. 18
Checked at Benchmark EA South E85088
With Temperature Gun ID #7

Client Name: FDEP
Address: 2600 Blair Stone Rd.
Tallahassee, FL 32399
Phone: (239) 344-5719
Email: dep-overflowmicro@dep.state.fl.us

Bill to: Frances Moore
P.O. #: Lab 855341

COEFES

Project Name: RQ-2021-
2022-12-05-51

Project Run/Location: SMP, Grophner

Laboratory Submission #:

582190651

Station ID

Surface Water
Sample Matrix
(Fresh / Marine)
Sample Type

Enterococci
E.coli (SM9223B Quantitray)

Laboratory
Sample #

E.coli = 250 ml ; Enteroc = 120 ml ; Fecal Coliform = 120 ml

64SD0216

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

F / M / Grab

E.coli

Enteroc

F.coliform

Date & Time

12/21/22

930

1

Laboratory Sample Acceptability:
pH < 10 BEAS Temperature: 18°C

1. The client is responsible for documentation of the sampling event. Please note: correct sampling events on the sample receipt form.
2. Sample Matrix is used to indicate whether the sample is being subjected to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), salt surface water (SSW), or surface water (SW).
3. "Container Type" is used to indicate whether the container is plastic (P) or glass (G).
4. Sample must be refrigerated or stored in wet ice after collection. The maximum temperature during storage should be 0°C (32°F). Under "Preservation," list any preservatives that were added to the sample container.
5. "Date" is the date the sample was collected. The date and time of collection, sample ID, and any field number or ID.
6. "Time" is the time the sample was collected. The date and time of collection, sample ID, and any field number or ID.
7. "Time" is the time the sample was collected. The date and time of collection, sample ID, and any field number or ID.
8. "Time" is the time the sample was collected. The date and time of collection, sample ID, and any field number or ID.
9. "Time" is the time the sample was collected. The date and time of collection, sample ID, and any field number or ID.

Collector: NADIA JACOB

Date: 12/21/22

Time: 1555

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Received By: M. Elind

Redeemed By:

Date:

Time:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Redeemed By:

Date:

Time:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Redeemed By:

Date:

Time:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Redeemed By:

Date:

Time:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By:

Received By: