

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: **1550 77**

RQ: **2022-07-25-44**

Project: **SMH**

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification **6/27/22**

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<i>Hank Hing</i>	7/27	0711	23.7	763	8.5	8.55	100.2	—	1.0689	P/F	L/F
ICV	<i>Hank Hing</i>	7/27	0720	23.4	763	8.5	8.51	100.1	—	—	P/F	L/F
CCV	<i>Hank Hing</i>	7/27	1504	23.2	762.4	7.7	7.93	102.2	—	—	P/F	L/F
CCV				25.6							P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	<i>Hank Hing</i>	7/27	0727	210330B	9/2022	5,000	5002	P/F	L/F
ICV	<i>Hank Hing</i>	7/27	0732	220413C	4/2023	1,000	985	P/F	L/F
CCV	<i>Hank Hing</i>	7/27	1513	220415D	4/2023	100	98.2	P/F	L/F
CCV								P/F	L

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<i>Hank Hing</i>	7/27	0735	213754	6/2023	7.	24.3	7.01	-319	P/F	L/F
Calibr.	<i>Hank Hing</i>	7/27	0740	215478	7/2023	4.	21.4	4.00	140.6	P/F	L/F
Calibr.	<i>Hank Hing</i>	7/27	743	205284	10/2022	10.	21.3	10.04	-203.8	P/F	L/F
ICV	<i>Hank Hing</i>	7/27	0749	215478	7/2023	4	24.4	3.90	148	P/F	L/F
CCV	<i>Hank Hing</i>	7/27	1515	213754	6/2023	7	22.1	6.99	30.20	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 **171.9**, slope from 4 to 7 **172.5** (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? **YES** / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: **6/27/22**

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air	<i>Hank Hing</i>	7/27	0751	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

COMMENTS: