

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: **TRENDYST**

RQ: **2022-09-1239**

Project: **SMP**

~~Holdly "X" this box if there are qualified data on this page.~~

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO FEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Hannah H	9/19/22	729	27.0	761.9	8.0	8.6	100		1.172	P/F	L/F
ICV	Hannah H	9/19/22	730	26.9	761.9	8.0	8.02	100.2			P/F	L/F
CCV	Hannah H.	9/19	1531	23.7	760.1	8.5	9.84	116.3		138.1	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Hannah H	9/19	733	2103308	9/30	5,000	5003	P/F	L/F
ICV	Hannah H	9/19	734	220413D	9/23	100	100.5	P/F	L/F
CCV	Hannah H.	9/19	1619	220413D	4/23	100	100.2	P/F	L/F
CCV								P/F	L

Conductivity Acceptance criteria $\pm 5\%$

pH FEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Hannah H	9/19	734	213754	6/23	7.	25.5	7.01	-6.1	P/F	L/F
Calibr.		9/19	736	1LE0260	6/23	4.	27.4	4.02	168.4	P/F	L/F
Calibr.		9/19	738	205284	10/22	10.	24.9	9.98	-179.3	P/F	L/F
ICV	Hannah H	9/19	746	1LE0260	6/23	4	27.0	4.04	165.2	P/F	L/F
CCV	Hannah H.	9/19	1622	213754	6/23	7	23.2	7.10	-16.3	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air	Hannah H	9/19	751	0	0	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

COMMENTS: