

CALIBRATION AND VERIFICATION LOG (DSS SOP FT 1000-FT 1500, FD 1000-FD 4000)

RQ-2022-19-23

Boldly "X" this box if there are qualified data on this page.

Meter ID: Trend 451

RQ: 2022-12-12-07

Project: Trend /SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 9/27/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	L. McNally	12/14	650	21.4	758.1	8.8	8.81	100	—	108	P/F	L/F
ICV			652	21.5	758.0	8.8	8.82	100.2	—		P/F	L/F
CCV	S. Shale	12/14	1655	23.5	754.9	8.5	8.54	101.1	—		P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass / Fail	Lab / Field
Calibr.	L. McNally	12/14	653	220413B	4/23	10000	10000	P/F	L/F
ICV			655	220413D	4/23	100	99.8	P/F	L/F
CCV	S. Shale	12/14	1706	2204130	4/23	100	99.8	P/F	L/F
CCV	S. Shale	12/14	1708	208413L	4/23	1000	994	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	L. McNally	12/14	656	1LD0599	8/23	7.	21.7	7.00	-12.8	P/F	L/F
Calibr.			657	1LE0260	6/23	4.	21.5	4.00	158.5	P/F	L/F
Calibr.			659	1LF0738	8/23	10.	21.6	10.00	-182.5	P/F	L/F
ICV			701	1LE0260	6/23	4.00	21.5	4.05	155.7	P/F	L/F
CCV	S. Shale	12/14	1710	1LF0738	08/23	10.6	20.6	9.92	-122.4	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.7, slope from 4 to 7 171.3 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/27/22

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	L. McNally	12/14	700	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

COMMENTS: