From:	Jeremy Bosso
To:	Slade, Christopher "Stopher"; DEP-Overflowmicro; King, Michael; Butler, Myranda K
Subject:	microbiology report for RQ-2022-04-18-27 samples collected 04/21/2022
Date:	Friday, April 22, 2022 5:24:31 PM
Attachments:	FF_2022_04_21A.txt DEP_Reports_2022.xlsx STF_FDEP_2022_04_21A.pdf

---

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.

Good afternoon. I've attached the microbiology report for RQ-2022-04-18-27 samples collected 04/21/2022.

Samples analyzed for *E. coli* by **SM 9223B (Colilert-18 QT)** were incubated from 04/21/2022 (15:30) to 04/22/2022 (12:16).

Note: The "J2" qualifier on sample 22-49509 (G4NW0382) is due to our lab duplicate for that sample being outside of our internal precision acceptance criteria. I believe this is most likely due to random error (the duplicate results were 4.1 versus 13.5 MPN/100 mL before multiplying the dilution factor of 10). We always report the higher number of duplicate analyses.

Thank you,
Jeremy Bosso, MPH
Technical Director, Wetlands Research Laboratory
An Analytical Arm of the Center for Environmental Diagnostics & Bioremediation
University of West Florida
11000 University Pkwy, Pensacola, FL 32514
(850) 857-6096

Confidentiality Notice: This e-mail communication and any attachments may contain confidential and privileged information for the use of the designated recipient(s) named above. If you are not the intended recipient, you are hereby notified that you have received this communication in error and that any review, disclosure, dissemination, distribution or copying of it or its contents is prohibited.