

Klug, Kyle

From: Klug, Kyle
Sent: Wednesday, November 30, 2022 13:54
To: Olin, Brook
Cc: Henning, Hannah; Slade, Christopher "Stopher"
Subject: FW: Improperly labeled sample
Attachments: 20221130_115710.jpg

Importance: High

Good afternoon,

If you are repurposing old containers, please do not forget to write the correct RQ on the containers. The cooler with these samples arrived today and without an RQ or collection date written on the bottles, we did not notice that the cooler was supposed to have arrived yesterday.

Everything arrived in temp, however the short hold samples were expired and will not be logged in as per the SMP protocol.

Thank you,
Kyle

From: Klug, Kyle <Kyle.Klug@FloridaDEP.gov>
Sent: Wednesday, November 30, 2022 12:18
To: Klug, Kyle <Kyle.Klug@FloridaDEP.gov>
Subject: Improperly labeled sample

Sent from my Verizon, Samsung Galaxy smartphone
Get [Outlook for Android](#)

Login Checklist

RQ- 2022-11-21-06

Login Station ID: #1

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 11/ 30 /22 @ 10:04

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
		YES	HNO ₃		Formalin	Present?		*Intact?	
	Yes					No	Yes	No	
0.5				5		✓			5859 4728 9040

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. **NCR#** _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. **NCR#** _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. **NCR#** _____
- If coolers or samples are received late; identify the affected samples in an NCR. **NCR#** _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No ☒ NA

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA ☒ Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____ Containers intact? Yes ☒ No _____

Caps on tight?

Yes ☒ No _____

Sufficient Sample Volume:

Yes ☒ No _____

If **NO** is checked above, generate an NCR listing affected sample IDs in its appropriate category. **NCR#** _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes ☒ No _____ Init: NA

Chlorine detected? (One container checked per Field ID)

Yes _____ No ☒ Init: NA

If chlorine is detected, generate an NCR listing affected sample IDs . **NCR#** _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes ☒ No _____ NA _____ Init: NA

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes _____ No _____ NA ☒ Init: NA

If samples were not preserved correctly, generate an NCR listing affected sample IDs . **NCR#** _____

Coolers Unpacked/Checked by: NA

Date: 11/ 30 /22 Event contains NCRs(Y/N)? N

EventID: Holmes Run
NW-DIST-2022-11-30-01 (SMP)
Date Received: 30-NOV-2022 10:04

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #1 Exp:10/17/2023
pH Test Strips 0 – 6	Lot# 223819AV Exp:8/30/23
pH Paper 0 – 3	
pH Paper 0 – 13	Lot# 207621 Exp:3/15/24
Chlorine Test Strips	Lot# 0070 Exp:2/23

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 11/28/2022 **Collected By:** Brook Olin, Hannah Henning
Customer: NW-DIST **Sampling Agency:** FDEP
RQ: RQ-2022-11-21-06 **Project ID:** SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 1

Delivery Method: FedEx

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Sikes Creek upstream of Bonifay Gritney Rd	 G3NW0159

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 11/28/2022

1600 CEN

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: 

Rec'd/Inspected

Date: 11-30-22 / 1004



RQ-2022-11-21-06	Collected By: Brook Olin, Hannah Henning
Customer: FDEP	Field Report Prepared by: Brook Olin, Hannah Henning
Project ID: SMP	Sampling Agency: FDEP
WIN ID/Field ID: WIN ID/Field ID: G3NW0159	Comments: Microbiology sample(s) submitted to contract laboratory.

Location: Sikes Creek upstream of Bonifay Gritney Rd

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {CEN}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT _h)					
11/28/2022 1101	6.4	62.6	14.4	6.16	0.05	0.1	0.1	107.7	0.05					
Parameter Suite	Parameter Methods				Preservation		pH Check		# Bottles sent to Lab	Bottle Group				
Preserved Nutrients (P-500ML)	X	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP			X	Ice	X	H2SO4*	X	<2	1	A		
		Lot #	2145160		Exp:	6/02/2023								
Metals (P-500ML)	x	W-ICPMS	X	W-ICP	X	W-HARD	X	Ice	X	HNO3*		<2	1	A
		Lot #	2126030		Exp:	6/02/2023								
Filtered Nutrients (P125ML)		W-PO4-F		Field Filtered 0.45 um Filter			Ice							
		Filter Lot #		Syringe Lot #										
BOD (BP-1L)		OV-BODUNIN					Ice							
Chlorophyll (BP-1L)	x	CHLSUITE-W					Ice				1	A		
Physical/Aggregate (Fresh) (P-1L)	X	Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				x	Ice				1	A		
Physical/Aggregate (Marine) (P-1L)		Alkalinity / W-F / W-TSS / TURBIDITY					Ice							
Macroinvertebrates (PJ-2L)		MI-FW-QLDC		Formalin Lot #			Formalin							
Periphyton (PT-50ML)		ALGAE-ID					Ice							
Microbiology (Contract Lab Bottle)	X	E Coli		F Coli		Enterococci	Ice			1	A			
	X	Contract Lab												
Molecular (QPCR-P-500ML)		PCR-HF183		PCR-BacR		PCR-DG3	Ice							
		PCR-GULL2		PCR-GFD										
Tracers (BG-500ML)		W-E8321-DI		W-E8321-MS			Ice							
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI		W-E8321-MS		W-PSNP-TQ	x	Ice	MCAA Exp.	MCAA Lot#	1	A		
	X	W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA Buffer Solution)	X	W-CT-CL-R								
		W-TR-SQ-R	X	W-PCB-TQ-R										
Phytoplankton		DTY-QN-C		PKD-QN-BV			Ice							
DOC		W-DOC		Field Filtered 0.45 um Filter			Ice		H2SO 4*		<2			
			Acid Preservative Expiration Date											
			Acid Preservative Lot Number											

Document Reference # 05.1.0

Laboratory: Wetlands Research Laboratory

Sampler's initials : HH/DO

Page 1 of 1

Date/Time: 11/28/22 1 1334

Date/Time: 11/28/22 1334

CST

[illegible]

Comments: **Sample Container lot#** 621347# PO # C01214 Surface water, no chlorine expected

RQ - 2022-11-21-06

Contact name: **Mike King**

Phone: 850-595-8300

Sampler's name:

¹Matrix examples: SW = surface water, PW = pore water, GW = ground water, FW = Freshwater, MW = Marine Water, BW = Brackish Water, S = soil/sediment, Other X =

³ Recorded using IRT-2³ Recorded using IRT-2

Date of Request: 17-OCT-2022
 Created By: SLADE_C On: 17-OCT-2022
 Modified By: JASROTIA_P On: 03-NOV-2022
 Customer: NW-DIST
 Project: SMP
 Division: Division of Environmental Assessment and Restoration
 District: Northwest District
 Event Description: Holmes Run

Send Coolers To:

Phone: 850-595-0567
 160 W Government Street
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Stopher Slade
 Cooler Ship EMail: Christopher.slade@FloridaDEP.gov

Priority: 3
 Event Status: A
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 03-NOV-2022
 Biology Request Reviewed By: JASROTIA_P On: 03-NOV-2022
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped:
 Sampling Kit Packed By:
 Preservatives Needed: NO
 Date to Receive Samples: 21-NOV-2022
 Logged In By:

Send Final Report To:

160 W Government Street
 Suite 208
 Pensacola, FL 32502-5794
 Attn: Myranda K Butler
 Report Notification EMail:
 Christopher.Slade@FloridaDEP.gov;myranda.k.butler@dep.state.fl.us

Comments: Group A: (WBID 142) Freshwater Nutrients, Metals, Limited Pest, Contract Lab Bacteria
 Group B: (WBID 142, 94) Metals, Contract Lab Bacteria
 Group C: (WBID 94) Metals, Tracers, Markers, Contract Lab Bacteria

Packing notes;

FedEx return packing slips

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-1L	Number of Bottles: 1	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2015) Total Dissolved Solids in aqueous matrices
W-TSS	Template: DEFAULT	(SM 2540 D-2015) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 1	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: TEST HAS NO	Number of Bottles: 1	Preserved With: ICE
NW-UWF-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert-18/QT by UWF in Pensacola
Bottle Type: BG-500ML FILL TO	Number of Bottles: 4	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS

Bottle Group A (Water) With 1 Samples:

Bottle Type: BG-500ML FILL TO	Number of Bottles: 4	Preserved With: ICE
W-PCL-TQ-R	Template:	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Boron		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: ICE And H2SO4
W-NH3	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group B (Water) With 2 Samples:

Bottle Type: TEST HAS NO	Number of Bottles: 2	Preserved With: ICE
NW-UWF-ECQ	Template: DEFAULT	(SM 9223 B Quanti-Tray) E. coli by Colilert-18/QT by UWF in Pensacola
Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		

Bottle Group B (Water) With 2 Samples:

Bottle Type: P-500ML	Number of Bottles: 2	Preserved With: HNO3
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Aluminum
Antimony
Arsenic
Barium
Beryllium
Boron
Cadmium
Chromium
Cobalt
Copper
Lead
Manganese
Molybdenum
Nickel
Selenium
Silver
Thallium

Bottle Group C (Water) With 1 Samples:

Bottle Type: QPCR-P-500ML	Number of Bottles: 1	Preserved With: ICE
PCR-BACR	Template: DEFAULT	(SOP-PCR-1.4) Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction
PCR-DG3	Template: DEFAULT	(SOP-PCR-1.5) Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction
PCR-HF183	Template: DEFAULT	(SOP-PCR-1.0) Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction

Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO3
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Calcium
Iron
Magnesium
Potassium
Sodium
Strontium
Tin
Titanium
Vanadium
Zinc

W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
---------	-------------------	---

Aluminum
Antimony
Arsenic
Barium
Beryllium
Boron
Cadmium
Chromium
Cobalt
Copper
Lead
Manganese
Molybdenum

Bottle Group C (Water) With 1 Samples:

Bottle Type: P-500ML Number of Bottles: 1 Preserved With: HNO3
W-ICPMS Template: DEFAULT (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples
supporting Clean Water Act Projects

Nickel
Selenium
Silver
Thallium

Bottle Type: TEST HAS NO Number of Bottles: 1 Preserved With: ICE
NW-UWF-ECQ Template: DEFAULT (SM 9223 B Quanti-Tray) E. coli by Colilert-18/QT by UWF in Pensacola

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.