

Chemical Analysis Report

TLH-ROC-2022-01-31-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Run 11**
Request ID: **RQ-2022-01-10-34**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 23-FEB-2022 11:56

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line above the "T" and a long horizontal stroke at the end.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Sandy Creek Upstream Boat Ramp

Collection Date/Time: 01/31/2022 11:15

Field ID: G3TLHR0037

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2302628	EPA 180.1 Rev. 2.0	Turbidity	2.6	A	NTU	P409273	
	SM 2320 B-2011	Total Alkalinity	6.0		mg CaCO3/L	P409433	
	SM 2540 D-2011	TSS	2	I	mg/L	P409808	
	SM 4500 F-C-2011	Fluoride	0.035	U	mg F/L	P409436	
2302629	EPA 350.1 Rev. 2.0	Ammonia-N	0.014		mg N/L	P409636	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.21		mg N/L	P409303	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.074		mg N/L	P409451	
	EPA 365.1 Rev. 2.0	Total-P	0.005		mg P/L	P409295	
	SM 5310 B-2011	Organic Carbon	4.2		mg C/L	P409975	
2302630	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P409260	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P409273

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409636

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409303

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409451

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409295

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409260

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P409433

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-2011
Batch ID: P409808

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-2011
Batch ID: P409436

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 5310 B-2011
Batch ID: P409975

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P409273

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	95.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409636

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409451

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409295

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409260

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.1		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P409433

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.6		P	94 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P409436

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.7		P	93 - 104

Reference Method: SM 5310 B-2011
Batch ID: P409975

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409636

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302395	Ammonia-N	97.6	98.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302640	Kjeldahl Nitrogen	100	96.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409451

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302605	NO2NO3-N	98.9	97.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409295

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302604	Total-P	103	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409260

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302643	O-Phosphate-P	97.3	99.1	P/P	90 - 110

Reference Method: SM 4500 F-C-2011
Batch ID: P409436

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302600	Fluoride	99.2	98.5	P/P	92 - 107

Reference Method: SM 5310 B-2011
Batch ID: P409975

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2302623	Organic Carbon	97.6	94.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P409273

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302628	Turbidity	1.93	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P409636

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302395	Ammonia-N	1.11	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P409303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302640	Kjeldahl Nitrogen	3.61	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P409451

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302605	NO2NO3-N	0.885	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P409295

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302604	Total-P	1.87	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P409260

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302643	O-Phosphate-P	1.63	Spike	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P409433

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302600	Total Alkalinity	0.518	Sample	P	0 - 5.2

Reference Method: SM 4500 F-C-2011
Batch ID: P409436

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302600	Fluoride	0.480	Spike	P	0 - 2.1

Quality Assurance Report Precision

Reference Method: SM 5310 B-2011
Batch ID: P409975

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2302623	Organic Carbon	2.02	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A110151

Included Lab Sample IDs: 2302630

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.7	99.0	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A110157

Included Lab Sample IDs: 2302628

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.0	P/P	95 - 105

Reference Method: SM 2320 B-2011

Run ID: A110225

Included Lab Sample IDs: 2302628

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.3	97.5	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A110225

Included Lab Sample IDs: 2302628

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A110265

Included Lab Sample IDs: 2302629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.3	96.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A110277

Included Lab Sample IDs: 2302629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	104	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A110313

Included Lab Sample IDs: 2302629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.2	96.4	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A110463

Included Lab Sample IDs: 2302629

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.5	93.4	P/P	90 - 110

Quality Assurance Report

Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity	95.9				1.93	
EPA 350.1 Rev. 2.0	Ammonia-N	95.6	97.6	98.8			1.11
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	100	96.0			3.61
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	98.9	97.9			0.885
EPA 365.1 Rev. 2.0	Total-P	102	103	101			1.87
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.1	97.3	99.1			1.63
SM 2320 B-2011	Total Alkalinity	98.6				0.518	
SM 4500 F-C-2011	Fluoride	97.7	99.2	98.5			0.480
SM 5310 B-2011	Organic Carbon	96.4	97.6	94.9			2.02

Reference Method Descriptions

Method	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2302628
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2302629
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2302629
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2302629
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2302629
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2302630
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2302628
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2302628
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2302628
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2302629

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	01/31/2022			02/01/2022 15:00	Sarah A Nixon	2302628
EPA 350.1 Rev. 2.0	01/31/2022			02/09/2022 13:30	Roberta Tatti	2302629
EPA 351.2 Rev. 2.0	01/31/2022	02/03/2022 10:35	Samantha L Bates	02/04/2022 12:32	Alexandra J Mattheus	2302629
EPA 353.2 Rev. 2.0	01/31/2022			02/04/2022 09:52	Beverly Y. Sanders	2302629
EPA 365.1 Rev. 2.0	01/31/2022	02/03/2022 14:00	Simonne.V. Calhoun	02/07/2022 14:16	Shengyu Ding	2302629
EPA 365.1 Rev. 2.0 dissolved	01/31/2022			02/01/2022 14:27	James L Waggerby	2302630
SM 2320 B-2011	01/31/2022			02/04/2022 02:37	Dale L. Simmons	2302628
SM 2540 D-2011	01/31/2022			02/04/2022 14:55	Yijie Li	2302628
SM 4500 F-C-2011	01/31/2022			02/04/2022 02:37	Dale L. Simmons	2302628
SM 5310 B-2011	01/31/2022			02/16/2022 09:08	Nathaniel J Jones	2302629