

Chemical Analysis Report

TLH-ROC-2022-08-04-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Run 15**
Request ID: **RQ-2022-07-18-48**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 26-AUG-2022 09:49



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Camp Branch @ CR173

Collection Date/Time: 08/04/2022 12:20

Field ID: G3TLHR0004

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2346403	DEP SOP: NU-094-1	Color (true)	13		PCU	P417737	
	EPA 180.1 Rev. 2.0	Turbidity	9.4		NTU	P417742	
	EPA 300.0 Rev. 2.1	Chloride	4.4		mg Cl/L	P418021	
		Sulfate	1.3		mg SO4/L	P418026	
	SM 2320 B-2011	Total Alkalinity	105		mg CaCO3/L	P418047	
	SM 2540 C-2011	TDS	119		mg/L	P418184	
	SM 2540 D-2011	TSS	18		mg/L	P418228	
	SM 4500-F- C-2011	Fluoride	0.042	I	mg F/L	P418051	
2346404	EPA 350.1 Rev. 2.0	Ammonia-N	0.080		mg N/L	P417974	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.31		mg N/L	P417861	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.070		mg N/L	P417757	
	EPA 365.1 Rev. 2.0	Total-P	0.061		mg P/L	P417925	
	SM 5310 B-2011	Organic Carbon	2.5		mg C/L	P417814	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P417737

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P417742

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418021

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418026

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P417974

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P417861

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P417757

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417925

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P418047

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2011
Batch ID: P418184

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P418228

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P418051

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P417814

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P417737

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P417742

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	96.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418021

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418026

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P417974

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P417861

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.1		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P417757

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417925

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P418047

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	94 - 106

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500-F- C-2011
Batch ID: P418051

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P417814

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418021

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346156	Chloride	96.8		P	90 - 110
2346288	Chloride	96.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418026

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346156	Sulfate	91.9		P	90 - 110
2346288	Sulfate	93.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P417974

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346644	Ammonia-N	98.4	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P417757

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346289	NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417925

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346375	Total-P	104	103	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P418051

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346639	Fluoride	100	100	P/P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P417814

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2346159	Organic Carbon	94.6	94.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P417737

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346639	Color (true)	1.43	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P417742

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346489	Turbidity	0.314	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418021

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346156	Chloride	0.269	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P418026

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346156	Sulfate	0.0297	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P417974

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346644	Ammonia-N	2.61	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P417861

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346469	Kjeldahl Nitrogen	5.59	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P417757

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346289	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P417925

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346375	Total-P	0.812	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P418047

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346639	Total Alkalinity	0.772	Sample	P	0 - 3.8

Reference Method: SM 2540 C-2011
Batch ID: P418184

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346363	TDS	0.980	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P418051

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346639	Fluoride	0.0	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P417814

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2346159	Organic Carbon	0.207	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A113939

Included Lab Sample IDs: 2346403

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	103	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A113940

Included Lab Sample IDs: 2346403

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A113955

Included Lab Sample IDs: 2346404

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A113980

Included Lab Sample IDs: 2346404

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.8	95.8	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A114030

Included Lab Sample IDs: 2346404

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A114034

Included Lab Sample IDs: 2346404

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A114048

Included Lab Sample IDs: 2346403

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	102	P/P	90 - 110
Sulfate	102	99.5	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A114059

Included Lab Sample IDs: 2346403

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.5	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500-F- C-2011

Run ID: A114059

Included Lab Sample IDs: 2346403

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A114118

Included Lab Sample IDs: 2346404

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	90.5	91.7	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
DEP SOP: NU-094-1	Color (true)	102			1.43	
EPA 180.1 Rev. 2.0	Turbidity	96.3			0.314	
EPA 300.0 Rev. 2.1	Chloride	102	96.8	96.8	0.269	
	Sulfate	98.8	91.9	93.7	0.0297	
EPA 350.1 Rev. 2.0	Ammonia-N	96.8	98.4	101		2.61
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.1				5.59
EPA 353.2 Rev. 2.0	NO2NO3-N	100	100	100		0.0
EPA 365.1 Rev. 2.0	Total-P	103	104	103		0.812
SM 2320 B-2011	Total Alkalinity	100			0.772	
SM 2540 C-2011	TDS				0.980	
SM 4500-F- C-2011	Fluoride	101	100	100		0.0
SM 5310 B-2011	Organic Carbon	99.0	94.6	94.2		0.207

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2346403
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2346403
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2346403
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2346403
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2346404
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2346404
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2346404
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2346404
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2346403
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2346403
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2346403
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2346403
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2346404

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	08/04/2022			08/05/2022 15:22	Samantha L Bates	2346403
EPA 180.1 Rev. 2.0	08/04/2022			08/05/2022 15:05	Khloe J Berg	2346403
EPA 300.0 Rev. 2.1	08/04/2022			08/11/2022 23:54	Anna M. Plaviak	2346403
EPA 350.1 Rev. 2.0	08/04/2022			08/11/2022 11:57	Ping Hua	2346404
EPA 351.2 Rev. 2.0	08/04/2022	08/10/2022 13:15	Simonne.V. Calhoun	08/15/2022 17:37	Samantha L Bates	2346404
EPA 353.2 Rev. 2.0	08/04/2022			08/08/2022 11:20	Beverly Y. Sanders	2346404
EPA 365.1 Rev. 2.0	08/04/2022	08/10/2022 16:53	Adam P Silver	08/11/2022 11:21	James L Waggerby	2346404
SM 2320 B-2011	08/04/2022			08/12/2022 07:57	Dale L. Simmons	2346403
SM 2540 C-2011	08/04/2022			08/09/2022 15:22	Yijie Li	2346403
SM 2540 D-2011	08/04/2022			08/09/2022 15:22	Yijie Li	2346403
SM 4500-F- C-2011	08/04/2022			08/12/2022 07:57	Dale L. Simmons	2346403
SM 5310 B-2011	08/04/2022			08/09/2022 06:16	Khloe J Berg	2346404