

Pendleton, Christopher

From: Blair, Paul
Sent: Monday, October 31, 2022 2:43 PM
To: Lab.Receiving
Subject: SAMPLES ARRIVING AT 16:20 RQ-2022-08-15-32
Attachments: RQ-2022-08-15-32_CoverSheetComplete_10312022.pdf

Having some technical issues. Disregard the 10:45am entry for G1TLHR0065. I've tried deleting and resubmitting and it won't update the custody sheet.

Get [Outlook for iOS](#)

Login Checklist

RQ- 2022-08-15-32

Login Station ID: #1

Shipping Method: FedEx | UPS | FD | Greyhound

Date/Time of Receipt: 10/31/22 @ 16:20

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
0.1				2		✓			

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ☒ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes _____ No ☒

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA ☒ Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes ☒ No _____

Caps on tight?

Yes ☒ No _____

Sufficient Sample Volume:

Yes _____ No _____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes _____ No ☒ Init: KK

Chlorine detected? (One container checked per Field ID)

Yes _____ No _____ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH< 2.0?

Yes ☒ No _____ NA _____ Init: KK

**W-1-4-DIOX (Green dot on container) preserved to pH<4.0?

Yes _____ No _____ NA ☒ Init: KK

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: KK/BK/ZR

Date: 10/31/22 Event contains NCRs(Y/N)? N

EventID: TLH-ROC-2022-10-31-02 (SMP)

Date Received: 31-OCT-2022 16:20

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	IR TEMP GUN #1 Exp:10/17/2023
pH Test Strips 0 – 6	Lot# 223819AV Exp:8/30/23
pH Paper 0 – 3	
pH Paper 0 – 13	Lot# 207621 Exp:3/15/24
Chlorine Test Strips	Lot# 0070 Exp:2/23

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 10/31/2022 **Collected By:** Paul Blair, Trey Fitzgerald
Customer: TLH-DIST **Sampling:** FDEP
RQ: RQ-2022-08-15-32 **Agency:**
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB

Number of Coolers Shipped: 3 **Delivery Method:** HandDelivered

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Steinhatchee River 40m N of Yearling Rd	 G1TLHR0065
Steinhatchee River 40m N of Yearling Rd	 G1TLHR0065

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 10/31/2022

1620 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By: C.P

Rec'd/Inspected

Date: 10/31/2022



RQ-2022-08-15-32

Customer: FDEP

Project ID: SMP

Collected By: Paul Blair, Trey Fitzgerald

Field Report Prepared by: Paul Blair, Trey Fitzgerald

Sampling Agency: FDEP

Comments:

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0065

Location: Steinhatchee River 40m N of Yearling Rd

Matrix: Fresh

(Check One)

☒ Grab☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos /cm)	Salinity (PPT)
10/31/2022 1045	4	44.5	20.4	7.4	0.1	0.2	0.2	542	0.26

Parameter Suite	Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*		<2	
	Lot #		Exp:						
Metals (P-500ML)	x	W-ICPMS	x	W-ICP	x	Ice	x	HNO3*	
	Lot #		1105100		Exp: 6-7-23		X	<2	1
Filtered Nutrients (P125ML)	W-PO4-F		Field Filtered 0.45 µm Filter		Ice				
	Filter Lot #		Syringe Lot #						
BOD (BP-1L)	OV-80DUNIN				Ice				
Chlorophyll (BP-1L)	CHLSUITE-W				Ice				
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-CHC / W-COLOR / W-SO4-IC / W-TDS				Ice				
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice				
Macroinvertebrates (PJ-2L)	MI-FW-QLDC		Formalin Lot #		Formalin				
Periphyton (PT-50ML)	ALGAE-ID				Ice				
Microbiology (Contract Lab Bottle)	x	E Coli		F Coli		Enterococci	x	Ice	1
	Contract Lab								
Molecular (QPCR-P-500ML)	PCR-HF183		PCR-BacR		PCR-DG3			Ice	
	PCR-GULL2		PCR-GFD						
Tracers (BG-500ML)	W-E8321-DI		W-E8321-MS		Ice				
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI		W-E8321-MS		W-PSNP-TQ			MCAA Exp.	MCAA Lot#
	W-PCL-TQ-R		W-CARB-AA (5 ml of MCAA Buffer Solution)		W-CT-CL-R			Ice	
	W-TR-SQ-R		W-PCB-TQ-R						
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice				
DOC	W-DOC		Field Filtered 0.45 µm Filter		Ice			H2SO4*	
			Acid Preservative Expiration Date						<2
			Acid Preservative Lot Number						



RQ-2022-08-15-32

Collected By: Paul Blair, Trey Fitzgerald

Customer: FDEP

Field Report Prepared by: Paul Blair, Trey Fitzgerald

Project ID: SMP

Sampling Agency: FDEP

Comments:

WIN ID/Field ID:



WIN ID/Field ID: G1TLHR0065

Location: Steinhatchee River 40m N of Yearling Rd

Matrix: Fresh

(Check One) ☒ Grab ☐ Composite

Date and Time {EST}	DO (mg/L)	DO (%SAT)	Temp. °C	pH	Sample Depth (m)	Secchi Depth	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)
10/31/2022 1035	4	44.5	20.4	7.4	0.1	0.2	0.2	542	0.26

Parameter Suite	Parameter Methods				Preservation	pH Check	# Bottles sent to Lab	Bottle Group						
Preserved Nutrients (P-500ML)	W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				Ice	H2SO4*								
	Lot #		Exp:											
Metals (P-500ML)	x	W-ICPMS	x	W-ICP	x	W-HARD	x	Ice	x	HNO3*	X	<2	1	D
	Lot #		1105100		Exp:		6-7-23							
Filtered Nutrients (P125ML)	W-PO4-F		Field Filtered 0.45 µm Filter		Ice									
	Filter Lot #	Syringe Lot #												
BOD (BP-1L)	OV-BODUNIN				Ice									
Chlorophyll (BP-1L)	CHLSUITE-W				Ice									
Physical/Aggregate (Fresh) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY / W-D-IC / W-COLOR / W-SO4-IC / W-TDS				Ice									
Physical/Aggregate (Marine) (P-1L)	Alkalinity / W-F / W-TSS / TURBIDITY				Ice									
Macroinvertebrates (PJ-2L)	MI-FW-QLDC	Formalin Lot #				Formalin								
Periphyton (PT-50ML)	ALGAE-ID				Ice									
Microbiology (Contract Lab Bottle)	x	E Coli		F Coli		Enterococci	x	Ice				1	D	
	Contract Lab													
Molecular (QPCR-P-500ML)		PCR-HF183		PCR-BacR		PCR-DG3		Ice						
		PCR-GULL2		PCR-GFD										
Tracers (BG-500ML)	W-E8321-DI		W-E8321-MS		Ice									
Pesticides (BG-500ML) (BG-1L)	W-E8321-DI	W-E8321-MS		W-PSNP-TQ		Ice	MCAA Exp.	MCAA Lot#						
	W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA Buffer Solution)		W-CT-CL-R										
	W-TR-SQ-R	W-PCB-TQ-R												
Phytoplankton	DTY-QN-C		PKD-QN-BV		Ice									
DOC	W-DOC	Field Filtered 0.45 µm Filter		Ice		H2SO4*								
		Add Preservative Expiration Date												
		Add Preservative Lot Number												

Date of Request: 20-JUL-2022
 Created By: GREEN_C On: 20-JUL-2022
 Modified By: JASROTIA_P On: 28-JUL-2022
 Customer: TLH-ROC
 Project: SMP
 Division:
 District: Not Applicable
 Event Description: June-Aug. North Run

Send Coolers To:

Phone: 850-245-8413
 2600 Blair Stone Rd

 Tallahassee, FL 32399
 Attn: Chris Green
 Cooler Ship EMail: Terry.Green@FloridaDEP.gov

Priority: 3
 Event Status: R
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 27-JUL-2022
 Biology Request Reviewed By: JASROTIA_P On: 28-JUL-2022
 Sampling Kit Required: YES Ship on: 12-AUG-2022
 Sampling Kit Shipped: YES On: 11-AUG-2022
 Sampling Kit Packed By: GIVENS_G
 Preservatives Needed: YES
 Date to Receive Samples: 15-AUG-2022
 Logged In By: KLUG_K On: 15-SEP-2022

Send Final Report To:

2600 Blair Stone Rd

 Tallahassee, FL 32399
 Attn: Chris Green
 Report Notification EMail: Terry.Green@FloridaDEP.gov

Comments: bottle A (G2/31)
 bottle B (G2/15)
 bottle C (G3/48 - G3/49)
 bottle D (G1/163 - G3/50)
 bottle E (G2/37 - G1/148)

#####

Buffer Solution for W-CARB-AA Bottle Needed

#####

W-TR-SQ-R (sunscreens)

Wash your hands thoroughly and wear gloves. Refrain from using sunscreens on the day of sample collection.

#####

Bottle Group A (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: BG-500ML FILL TO	Number of Bottles: 4	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

Bottle Group B (Water) With 1 Samples:

Bottle Type: P-250ML-WO-THI	Number of Bottles: 1	Preserved With: ICE
ECOLI-18QT	Template: DEFAULT	(SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: QPCR-P-500ML	Number of Bottles: 1	Preserved With: ICE
PCR-BACR	Template: DEFAULT	(SOP-PCR-1.4) Detection and quantification of Ruminant specific Bacteroidetes BacR genetic marker with quantitative polymerase chain reaction
PCR-DG3	Template: DEFAULT	(SOP-PCR-1.5) Detection and quantification of canine-specific DG3 Bacteroides genetic marker with quantitative polymerase chain reaction

Bottle Group B (Water) With 1 Samples:

Bottle Type: QPCR-P-500ML	Number of Bottles: 1	Preserved With: ICE
PCR-HF183	Template: DEFAULT	(SOP-PCR-1.0) Detection and quantification of human-specific HF183 Bacteroides genetic marker with quantitative polymerase chain reaction
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: CH ₂ ClCOOH and Ice
W-CARB-AA	Template: DEFAULT	(EPA 8321B) Carbamates in water matrices by HPLC/MS/MS
Bottle Type: BG-500ML FILL TO	Number of Bottles: 2	Preserved With: ICE
W-CT-CI-R	Template: DEFAULT	(EPA 8276) Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization
W-PCB-TQ-R	Template: DEFAULT	(EPA 8270E) PCBs in water matrices by GC/MS/MS
W-PCL-TQ-R	Template: DEFAULT	(EPA 8270E) Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS
Bottle Type: BG-500ML	Number of Bottles: 1	Preserved With: ICE
W-E8321-DI	Template: DEFAULT	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
W-E8321-MS	Template: SMP	(EPA 8321B) Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS
Bottle Type: P-500ML	Number of Bottles: 1	Preserved With: HNO ₃
W-HARD	Template: DEFAULT	(SM 2340 B-2011) Hardness by calculation, in aqueous matrices, reported as mg/L CaCO ₃
Hardness		
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects
Calcium		
Iron		
Magnesium		
Potassium		
Sodium		
Strontium		
Tin		
Titanium		
Vanadium		
Zinc		
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects
Aluminum		
Antimony		
Arsenic		
Barium		
Beryllium		
Boron		
Cadmium		
Chromium		
Cobalt		
Copper		
Lead		
Manganese		
Molybdenum		
Nickel		
Selenium		
Silver		
Thallium		
Bottle Type: BG-1L	Number of Bottles: 4	Preserved With: ICE
W-PSNP-TQ	Template: DEFAULT	(EPA 8270E) Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS
W-TR-SQ-R	Template:	(EPA 8270E) Analysis of Sunscreen Compounds and Flame Retardants in Water by GC/MS-SIM

Bottle Group C (Water) With 2 Samples:

Bottle Type: P-1L	Number of Bottles: 2	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
TURBIDITY	Template: DEFAULT	(EPA 180.1 Rev. 2.0) Turbidity in aqueous matrices
W-CL-IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Chloride in aqueous matrices
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
Color (true)		
W-F	Template: DEFAULT	(SM 4500-F- C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO ₄ -IC	Template: DEFAULT	(EPA 300.0 Rev. 2.1) Sulfate in aqueous matrices
W-TDS	Template: DEFAULT	(SM 2540 C-2015) Total Dissolved Solids in aqueous matrices

Bottle Group C (Water) With 2 Samples:

Bottle Type: P-1L W-TSS	Template: DEFAULT	Number of Bottles: 2 (SM 2540 D-2015)	Preserved With: ICE Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L CHLSUITE-W	Template: DEFAULT	Number of Bottles: 2 (SM 10200 H (mod.))	Preserved With: ICE Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-250ML-WO-THI ECOLI-18QT	Template: DEFAULT	Number of Bottles: 2 (SM 9223 B Quanti-Tray)	Preserved With: ICE Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML W-HARD	Template: DEFAULT	Number of Bottles: 2 (SM 2340 B-2011)	Preserved With: HNO3 Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness			
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	
Calcium			
Iron			
Magnesium			
Potassium			
Sodium			
Strontium			
Tin			
Titanium			
Vanadium			
Zinc			
W-ICPMS	Template: DEFAULT	(EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	
Aluminum			
Antimony			
Arsenic			
Barium			
Beryllium			
Boron			
Cadmium			
Chromium			
Cobalt			
Copper			
Lead			
Manganese			
Molybdenum			
Nickel			
Selenium			
Silver			
Thallium			
Bottle Type: P-500ML W-NH3	Template: DEFAULT	Number of Bottles: 2 (EPA 350.1 Rev. 2.0)	Preserved With: ICE And H2SO4 Ammonia in aqueous matrices as mg N/L
W-NO2NO3	Template: DEFAULT	(EPA 353.2 Rev. 2.0)	Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0)	Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0)	Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011)	Nonpurgeable Organic Carbon in aqueous matrices

Bottle Group D (Water) With 2 Samples:

Bottle Type: P-250ML-WO-THI ECOLI-18QT	Template: DEFAULT	Number of Bottles: 2 (SM 9223 B Quanti-Tray)	Preserved With: ICE Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems
Bottle Type: P-500ML W-HARD	Template: DEFAULT	Number of Bottles: 2 (SM 2340 B-2011)	Preserved With: HNO3 Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3
Hardness			
W-ICP	Template: SMP-FRESH	(EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	
Calcium			
Iron			
Magnesium			

Bottle Group D (Water) With 2 Samples:

Bottle Type: P-500ML Number of Bottles: 2 Preserved With: HNO3
W-ICP Template: SMP-FRESH (EPA 200.7 Rev. 4.4) Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Potassium
Sodium
Strontium
Tin
Titanium
Vanadium
Zinc

W-ICPMS Template: DEFAULT (EPA 200.8 Rev. 5.4) Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Aluminum
Antimony
Arsenic
Barium
Beryllium
Boron
Cadmium
Chromium
Cobalt
Copper
Lead
Manganese
Molybdenum
Nickel
Selenium
Silver
Thallium

Bottle Group E (Water) With 2 Samples:

Bottle Type: P-250ML-WO-THI Number of Bottles: 2 Preserved With: ICE
ECOLI-18QT Template: DEFAULT (SM 9223 B Quanti-Tray) Escherichia coli by Colilert 18 Quanti-Tray method - non-chlorinated water systems

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.