

Chemical Analysis Report

TLH-ROC-2022-09-22-02

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Run 15**
Request ID: **RQ-2022-09-19-32**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 13-OCT-2022 16:53



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Indian Springs

Collection Date/Time: 09/22/2022 09:20

Field ID: G1TLHR0154

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2358935	DEP SOP: NU-094-1	Color (true)	39	A	PCU	P420030	
	EPA 180.1 Rev. 2.0	Turbidity	0.75	I	NTU	P420048	
	EPA 300.0 Rev. 2.1	Chloride	6.0		mg Cl/L	P420330	
		Sulfate	6.6		mg SO4/L	P420333	
	SM 2320 B-2011	Total Alkalinity	149		mg CaCO3/L	P420379	
	SM 2540 C-2011	TDS	176		mg/L	P420487	
	SM 2540 D-2011	TSS	2	U	mg/L	P420476	
	SM 4500-F- C-2011	Fluoride	0.14		mg F/L	P420383	
2358936	EPA 350.1 Rev. 2.0	Ammonia-N	0.034		mg N/L	P420735	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.44		mg N/L	P420203	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.023		mg N/L	P420412	
	EPA 365.1 Rev. 2.0	Total-P	0.022		mg P/L	P420231	
	SM 5310 B-2011	Organic Carbon	6.5		mg C/L	P420387	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P420030

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420048

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420330

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420333

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420735

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420203

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420412

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420231

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P420379

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2011
Batch ID: P420487

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P420476

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P420383

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P420387

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P420030

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.2		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420330

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420333

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420735

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420203

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420412

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420231

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P420379

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.6		P	94 - 106

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500-F- C-2011
Batch ID: P420383

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	102		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P420387

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420330

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358748	Chloride	93.5		P	90 - 110
2359074	Chloride	95.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420333

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358748	Sulfate	92.1		P	90 - 110
2359074	Sulfate	93.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420735

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358910	Ammonia-N	96.6	95.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420412

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358803	NO2NO3-N	95.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420231

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358977	Total-P	107	105	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P420383

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358674	Fluoride	100	100	P/P	94 - 104

Reference Method: SM 5310 B-2011
Batch ID: P420387

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2358936	Organic Carbon	100	97.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P420030

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358935	Color (true)	2.45	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P420048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359072	Turbidity	5.66	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420330

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358748	Chloride	0.236	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P420333

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358748	Sulfate	0.0504	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P420735

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358910	Ammonia-N	0.922	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P420203

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2359128	Kjeldahl Nitrogen	1.26	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P420412

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358803	NO2NO3-N	0.175	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P420231

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358977	Total-P	1.63	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P420379

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358674	Total Alkalinity	1.87	Sample	P	0 - 3.8

Reference Method: SM 2540 C-2011
Batch ID: P420487

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358674	TDS	3.48	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P420383

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358674	Fluoride	0.0	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011
Batch ID: P420387

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2358936	Organic Carbon	1.71	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A114911

Included Lab Sample IDs: 2358935

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.6	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A114918

Included Lab Sample IDs: 2358935

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A114999

Included Lab Sample IDs: 2358935

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.7	99.9	P/P	90 - 110
Sulfate	99.7	99.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A115024

Included Lab Sample IDs: 2358936

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.8	97.0	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A115070

Included Lab Sample IDs: 2358935

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	96.5	102	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A115070

Included Lab Sample IDs: 2358935

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	102	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A115074

Included Lab Sample IDs: 2358936

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.7	96.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A115093

Included Lab Sample IDs: 2358936

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A115115

Included Lab Sample IDs: 2358936

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	105	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A115220

Included Lab Sample IDs: 2358936

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.5	98.3	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
DEP SOP: NU-094-1	Color (true)	98.2				2.45	
EPA 180.1 Rev. 2.0	Turbidity	101				5.66	
EPA 300.0 Rev. 2.1	Chloride	98.7	93.5	95.1		0.236	
	Sulfate	96.3	92.1	93.9		0.0504	
EPA 350.1 Rev. 2.0	Ammonia-N	98.8	96.6	95.6			0.922
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100					1.26
EPA 353.2 Rev. 2.0	NO2NO3-N	104	95.5				0.175
EPA 365.1 Rev. 2.0	Total-P	104	107	105			1.63
SM 2320 B-2011	Total Alkalinity	98.6				1.87	
SM 2540 C-2011	TDS					3.48	
SM 4500-F- C-2011	Fluoride	102	100	100			0.0
SM 5310 B-2011	Organic Carbon	97.5	100	97.9			1.71

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2358935
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2358935
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2358935
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2358935
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2358936
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2358936
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2358936
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2358936
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2358935
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2358935
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2358935
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2358935
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2358936

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	09/22/2022			09/23/2022 10:35	Samantha L Bates	2358935
EPA 180.1 Rev. 2.0	09/22/2022			09/23/2022 10:49	Chandler E Cox	2358935
EPA 300.0 Rev. 2.1	09/22/2022			09/29/2022 16:45	Anna M. Plaviak	2358935
EPA 350.1 Rev. 2.0	09/22/2022			10/11/2022 14:07	Judith K. Clark	2358936
EPA 351.2 Rev. 2.0	09/22/2022	09/28/2022 12:10	Alexandra J Mattheus	10/04/2022 09:40	Samantha L Bates	2358936
EPA 353.2 Rev. 2.0	09/22/2022			10/03/2022 12:40	Beverly Y. Sanders	2358936
EPA 365.1 Rev. 2.0	09/22/2022	09/28/2022 14:15	Simonne.V. Calhoun	09/29/2022 12:40	James L Waggerby	2358936
SM 2320 B-2011	09/22/2022			10/01/2022 15:25	Dale L. Simmons	2358935
SM 2540 C-2011	09/22/2022			09/27/2022 15:36	Yijie Li	2358935
SM 2540 D-2011	09/22/2022			09/27/2022 15:36	Yijie Li	2358935
SM 4500-F- C-2011	09/22/2022			10/01/2022 15:25	Dale L. Simmons	2358935
SM 5310 B-2011	09/22/2022			09/29/2022 22:11	Touraj Touran	2358936