


BENCHMARK



EnviroAnalytical, Inc.

NELAP Certification #E84167

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAP STANDARDS

Submission Number : 22111725

F D E P/S W D
13051 N Telecom Pkwy.
Tampa, FL 33637

Project Name : RQ-2022-11-28-16
Date Received : 11/30/2022
Time Received : 13:30

Erica Peck

Submission Number: 22111725
Sample Number: 001
Sample Description: G1SW0127

Sample Date: 11/30/2022
Sample Time: 11:45
Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	10	#/100 ML	10	10	SM9223B	11/30/2022 13:57	12/01/2022 14:29	AT/DJ

Haley Richardson
Dale D. Dixon Laboratory Director

12/05/2022

Date

Tülay Tannrisever / Haley Richardson - QC Officers

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J5 = Estimated value. Data questionable due to improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis lost or not performed.

Q = Sample held beyond accepted hold time.
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are useable.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.
! = Data deviate from historically established concentration ranges.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
* = Not reported due to interference.
Oil & Grease - If client does not send sufficient sample quantity for spike evaluation surface water samples are supplied by the laboratory.

NOTES:

MBAS calculated as LAS; molecular weight = 340.
PQL = 4xMDL.
ND = Not detected at or above the adjusted reporting limit.
X = Value exceeds MCL.
G1 = Accuracy standard does not meet method control limits, but does meet lab control limits that are in agreement with USEPA generated data. USEPA letter available upon request.

COMMENTS:

1 (10 mL) dilution on each E. coli sample.

For questions or comments regarding these results, please contact us at (941) 723-9986.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc.
E841671711 Twelfth Street East
Palmetto, FL 34221

(941) 723-9986 / (941) 723-6061 fax

Sample Temperature checked upon receipt at BEA with
Temperature Gun ID #258**Benchmark EA South E85086**

1001 Corporate Avenue, Suite 102

North Port, FL 34289

(941) 625-3137

(941) 423-7336 fax

Sample Temperature checked upon receipt at BEAS
with Temperature Gun ID #7**Client Information: FDEP**Address: 13051 N. Telecom Parkway
Tampa FL 33604

Phone # 815-470-5878

Report to: benjamin.paswater@dep.state.fl.us

Invoice to: FDEP Micro Overflow

PO# CODE FSProject Name: RA-2022-11-28-1C

Laboratory Submission #

2211725

Sample Name	Sample Type ¹ / Sample Matrix ²	Collection		Container			Preservative ⁴	Parameters for Analysis	Laboratory Sample #
		Date	Time	Qty	Capacity	Type ³			
G1SW0127	SW	11-30-22	11:45	1	250mL	P	Plain (On Ice)	E. coli	1
	SW				250mL	P	Plain (On Ice)		
	SW				250mL	P	Plain (On Ice)		
	SW				250mL	P	Plain (On Ice)		
	SW				250mL	P	Plain (On Ice)		
	SW				250mL	P	Plain (On Ice)		
	SW				250mL	P	Plain (On Ice)		

1. "Sample Type" is used to indicate whether the sample was a grab (G) or whether it was a composite (C).

2. "Sample Matrix" is used to indicate whether the sample is being discharged to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil, sediment (SDMNT), or sludge (SLDG).

3. "Container Type" is used to indicate whether the container is plastic (P) or glass (G).

4. Sample must be refrigerated or stored in wet ice after collection. The temperature during storage should be less than or equal to 6°C (42.8°F).

5. Under "Preservative," list any preservatives that were added to the sample container. Lot Number of preservative used is specific to the bottles included in the kit. NaH₂PO₄, H₂SO₄, and HNO₃ do not have expiration dates per the manufacturer. Micro bottles are pre-preserved at manufacturing stage. 40mL vials are pre-preserved at manufacturing stage.**Instructions:**

1. Each bottle has a label identifying sample ID, premeasured preservative contained in the bottle, sample type, client ID, and parameters for analysis.

2. The following information should be added to each bottle label after collection with permanent black ink: date and time of collection, sampler's name or initials, and any field number or ID.

3. All bottles not containing preservative may be rinsed with appropriate sample prior to collection.

4. The client is responsible for documentation of the sampling event. Please note special sampling events on the sample custody form.

5. Sample kit has been created by BEA using new, certified bottles unless otherwise noted.

Laboratory Sample Acceptability:pH <: ☒ BEA Temperature: 1.0°C
BEAS Temperature:

1	Collected / Relinquished By & Affiliation: (Print & Sign) <u>Megan Maloney</u>	Date <u>11-30-22</u>	Time <u>13:30</u>	Received By & Affiliation: (Print & Sign) <u>Nathan Hadse 11</u>	Date <u>11-30-22</u>	Time <u>1330</u>
2	Relinquished By & Affiliation: (Print & Sign)	Date	Time	Received By & Affiliation: (Print & Sign)	Date	Time

Bea