

Chemical Analysis Report

SW-DIST-2022-12-15-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Run 14**
Request ID: **RQ-2022-11-07-30**
Customer: **SW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Ben Paswater

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 06-JAN-2023 17:22



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Beulah SW

Collection Date/Time: 12/14/2022 11:45

Field ID: G2SW0053

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2375619	DEP SOP: NU-094-1	Color (true)	10		PCU	P423605	
	EPA 180.1 Rev. 2.0	Turbidity	1.0		NTU	P423623	
	EPA 300.0 Rev. 2.1	Chloride	3.2		mg Cl/L	P423805	
		Sulfate	1.9		mg SO4/L	P423808	
	SM 2320 B-2011	Total Alkalinity	40		mg CaCO3/L	P423946	
	SM 2540 C-2015	TDS	57		mg/L	P424055	
	SM 2540 D-2015	TSS	2	I	mg/L	P423855	
	SM 4500-F- C-2011	Fluoride	0.24		mg F/L	P423951	
2375621	EPA 350.1 Rev. 2.0	Ammonia-N	0.008	Y	mg N/L	P424075	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.48	Y	mg N/L	P423749	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	IY	mg N/L	P423830	
	EPA 365.1 Rev. 2.0	Total-P	0.020	Y	mg P/L	P424020	
	SM 5310 B-2011	Organic Carbon	5.0	Y	mg C/L	P423699	
2375623	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P423622	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 351.2 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 353.2 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 365.1 Rev. 2.0: Total-P: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

SM 5310 B-2011: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

Sample Location: Crystal Lake @ Center

Collection Date/Time: 12/14/2022 12:30

Field ID: G3SW0076

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2375620	DEP SOP: NU-094-1	Color (true)	23		PCU	P423606	
	EPA 180.1 Rev. 2.0	Turbidity	7.8		NTU	P423623	
	EPA 300.0 Rev. 2.1	Chloride	9.4		mg Cl/L	P423805	
		Sulfate	5.4		mg SO4/L	P423808	
	SM 2320 B-2011	Total Alkalinity	34		mg CaCO3/L	P423946	
	SM 2540 C-2015	TDS	75		mg/L	P424055	
	SM 2540 D-2015	TSS	6	I	mg/L	P423855	
	SM 4500-F- C-2011	Fluoride	0.20		mg F/L	P423951	
2375622	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P424075	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P423749	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P423830	
	EPA 365.1 Rev. 2.0	Total-P	0.056		mg P/L	P423660	
	SM 5310 B-2011	Organic Carbon	8.8		mg C/L	P423699	
2375624	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P423622	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Non-Conformance Report

NCR ID: 9668

Event(s)

SW-DIST-2022-12-15-01

Job(s)

TLH-2022-12-15-20

Sample(s)

2375621

Test(s)

NCR Type: SHIPPING/RECEIVING

NCR Category: Preservation Not Intact

Observation: Samples not preserved to pH < 2.

Resolution: Sample preserved in the laboratory.

H2SO4 Lot# SA2229010 | Expiration Date: 08/30/2023

Authorized by/Date: Kyle Klug 12/19/2022

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
Please address questions to:

Chemistry	Colin Wright	(850) 245-8085
Biology	Cheryl Swanson	(850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P423605

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: DEP SOP: NU-094-1
Batch ID: P423606

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P423623

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423805

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423808

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P424075

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P423749

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P423830

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P423660

Component	Result	Code	Units
Total-P	0.010	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P424020

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P423622

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P423946

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2015
Batch ID: P424055

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P423855

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P423951

Component	Result	Code	Units
Fluoride	0.025	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P423699

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P423605

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	90 - 110

Reference Method: DEP SOP: NU-094-1
Batch ID: P423606

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	103		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P423623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423805

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423808

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	105		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P424075

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P423749

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.3		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P423830

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	108		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P423660

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P424020

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P423622

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.3		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P423946

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.9		P	94 - 106

Reference Method: SM 2540 C-2015
Batch ID: P424055

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	102		P	90 - 110

Reference Method: SM 2540 D-2015
Batch ID: P423855

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	98.1		P	85 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P423951

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	102		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P423699

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423805

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375700	Chloride	103		P	90 - 110
2375701	Chloride	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423808

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375700	Sulfate	103		P	90 - 110
2375701	Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P424075

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375621	Ammonia-N	96.6	97.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P423749

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375615	Kjeldahl Nitrogen	98.3	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P423830

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375563	NO2NO3-N	108	108	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P423660

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375704	Total-P	104	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P424020

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375704	Total-P	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P423622

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375623	O-Phosphate-P	100	98.1	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P423951

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375528	Fluoride	100	99.5	P/P	94 - 104

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: SM 5310 B-2011
Batch ID: P423699

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2375622	Organic Carbon	97.3	97.2	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P423605

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375416	Color (true)	0.0809	Sample	P	0 - 20

Reference Method: DEP SOP: NU-094-1
Batch ID: P423606

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375701	Color (true)	1.30	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P423623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375595	Turbidity	0.250	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423805

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375700	Chloride	0.251	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P423808

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375700	Sulfate	0.488	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P424075

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375621	Ammonia-N	1.06	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P423749

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375615	Kjeldahl Nitrogen	2.46	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P423830

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375563	NO2NO3-N	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P423660

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375704	Total-P	1.40	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P424020

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375704	Total-P	0.386	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P423622

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375623	O-Phosphate-P	2.12	Spike	P	0 - 20

Reference Method: SM 2320 B-2011

Batch ID: P423946

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375528	Total Alkalinity	0.302	Sample	P	0 - 3.8

Reference Method: SM 2540 C-2015

Batch ID: P424055

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375492	TDS	5.56	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011

Batch ID: P423951

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375528	Fluoride	0.477	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011

Batch ID: P423699

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2375622	Organic Carbon	0.0708	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1

Run ID: A116468

Included Lab Sample IDs: 2375619, 2375620

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 110
Color (true)	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A116477

Included Lab Sample IDs: 2375623, 2375624

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.5	97.8	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A116478

Included Lab Sample IDs: 2375619, 2375620

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 5310 B-2011

Run ID: A116527

Included Lab Sample IDs: 2375621, 2375622

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.4	98.8	P/P	90 - 110
Organic Carbon	98.7	98.4	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A116559

Included Lab Sample IDs: 2375619, 2375620

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	105	P/P	90 - 110
Sulfate	104	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A116591

Included Lab Sample IDs: 2375622

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A116597

Included Lab Sample IDs: 2375621, 2375622

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	106	108	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A116600

Included Lab Sample IDs: 2375621, 2375622

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.0	101	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A116652

Included Lab Sample IDs: 2375619, 2375620

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.0	99.6	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A116652

Included Lab Sample IDs: 2375619, 2375620

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	107	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A116709

Included Lab Sample IDs: 2375621

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A116711

Included Lab Sample IDs: 2375621, 2375622

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.3	97.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
DEP SOP: NU-094-1	Color (true)	101				0.0809	
	Color (true)	103				1.30	
EPA 180.1 Rev. 2.0	Turbidity	100				0.250	
EPA 300.0 Rev. 2.1	Chloride	105	103	104		0.251	
	Sulfate	105	103	102		0.488	
EPA 350.1 Rev. 2.0	Ammonia-N	97.6	96.6	97.8			1.06
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.3	98.3	102			2.46
EPA 353.2 Rev. 2.0	NO2NO3-N	108	108	108			0.0
EPA 365.1 Rev. 2.0	Total-P	104	104	102			1.40
	Total-P	104	104	104			0.386
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.3	100	98.1			2.12
SM 2320 B-2011	Total Alkalinity	98.9				0.302	
SM 2540 C-2015	TDS	102				5.56	
SM 2540 D-2015	TSS	98.1					
SM 4500-F- C-2011	Fluoride	102	100	99.5			0.477
SM 5310 B-2011	Organic Carbon	95.6	97.3	97.2			0.0708

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2375619, 2375620
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2375619, 2375620
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2375619, 2375620
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2375619, 2375620
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2375621, 2375622
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2375621, 2375622
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2375621, 2375622
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2375621, 2375622
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2375623, 2375624
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2375619, 2375620
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2375619, 2375620
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2375619, 2375620
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2375619, 2375620
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2375621, 2375622

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	12/15/2022			12/15/2022 14:21	Samantha L Bates	2375619
	12/15/2022			12/15/2022 14:24	Samantha L Bates	2375620
EPA 180.1 Rev. 2.0	12/15/2022			12/15/2022 13:35	Chandler E Cox	2375619
	12/15/2022			12/15/2022 13:36	Chandler E Cox	2375620
EPA 300.0 Rev. 2.1	12/15/2022			12/21/2022 07:44	Anna M. Plaviak	2375619
	12/15/2022			12/21/2022 07:59	Anna M. Plaviak	2375620
EPA 350.1 Rev. 2.0	12/15/2022			01/03/2023 16:29	Ping Hua	2375621
	12/15/2022			01/03/2023 16:36	Ping Hua	2375622
EPA 351.2 Rev. 2.0	12/15/2022	12/20/2022 13:35	Simonne.V. Calhoun	12/21/2022 13:45	Samantha L Bates	2375621
	12/15/2022	12/20/2022 13:35	Simonne.V. Calhoun	12/21/2022 13:46	Samantha L Bates	2375622
EPA 353.2 Rev. 2.0	12/15/2022			12/21/2022 10:23	Beverly Y. Sanders	2375621
	12/15/2022			12/21/2022 10:24	Beverly Y. Sanders	2375622
EPA 365.1 Rev. 2.0	12/15/2022	12/20/2022 16:15	Simonne.V. Calhoun	12/21/2022 10:52	James L Waggeberby	2375622
	12/15/2022	12/29/2022 14:13	Adam P Silver	01/03/2023 13:56	James L Waggeberby	2375621
EPA 365.1 Rev. 2.0 dissolved	12/15/2022			12/15/2022 16:05	Elise M. Gillis	2375623
	12/15/2022			12/15/2022 16:08	Elise M. Gillis	2375624
SM 2320 B-2011	12/15/2022			12/23/2022 10:41	Dale L. Simmons	2375619
	12/15/2022			12/23/2022 10:55	Dale L. Simmons	2375620
SM 2540 C-2015	12/15/2022			12/19/2022 16:08	Yijie Li	2375619, 2375620
SM 2540 D-2015	12/15/2022			12/19/2022 16:08	Yijie Li	2375619, 2375620
SM 4500-F- C-2011	12/15/2022			12/23/2022 10:41	Dale L. Simmons	2375619
	12/15/2022			12/23/2022 10:55	Dale L. Simmons	2375620
SM 5310 B-2011	12/15/2022			12/16/2022 21:24	Khloe J Berg	2375622
	12/15/2022			12/17/2022 01:02	Khloe J Berg	2375621