

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodor

RQ: 2022-01-31-61

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/18/2022

DO DEP SOP FD 1500	Name	Date	Time GT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	2/3/22	740	21.0	760	8.9	8.90	99.9	—	1.08	P/F	L/F
ICV	"	"	741	21.1	760	8.9	8.90	100.1	—	1.08	P/F	L/F
CCV	"	"	1401	23.2	760	8.5	8.52	99.8	—	1.08	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FD 1200	Name	Date	Time GT-ET	Lot#	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	2/3/22	743	1108FS1	8/2023	1000	1000	P/F	L/F
ICV	"	"	746	1105CS7	5/2023	100	97.8	P/F	L/F
CCV	"	"	1463	1105CS1	5/2023	50,000	49,638	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FD 1100	Name	Date	Time GT-ET	Lot#	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Andrew Luering	2/3/22	749	213013	4/2023	7.0	22.4	7.01	-30.2	P/F	L/F
Calibr.	"	"	751	2110D46	9/2023	4.0	21.5	4.00	146.2	P/F	L/F
Calibr.	"	"	754	1106829	11/2022	10.0	22.5	10.03	-119.5	P/F	L/F
ICV	"	"	756	1106829	11/2022	10.0	22.3	10.01	-119.4	P/F	L/F
CCV	"	"	1404	2110D40	9/2023	4.0	22.0	4.09	146.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 169, slope from 4 to 7 176.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 01/18/2022

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____ Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time GT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Andrew Luering	2/3/22	758	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: