

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Hodor

RQ: 2022-02-07-52

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/18/2022

DO	Name	Date	Time	Temp	Baro	DO	Meter	% DO	Probe	Probe	Pass	Lab
FDEP SOP			CT-ET	°C	mmHg	mg/L	D.O.	mg/L	Charge	Gain	Fail	Field
FT1500	Andrew Laering	2/8/22	706	20.9	757	8.9	8.89	99.6	-	1.09	P/F	P/F
Calibr.	Drew Liddick	"	"	709	757	8.9	8.94	100.4	-	1.09	P/F	P/F
ICV	"	"	"	1258	757	8.7	8.63	99.5	-	1.09	P/F	P/F
CCV	"	"	"								P/F	P/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec.	Name	Date	Time	Lot#	Expir.	Standard	Meter	Pass	Lab
Cond.			CT-ET		Date	Amhos/cm	Reading	Fail	Field
FT1200	Andrew Laering	2/8/22	711	1108FS1	8/2023	1000	1600	P/F	P/F
Calibr.	Drew Liddick	"	"	1102792	11/2023	100	98.4	P/F	P/F
ICV	"	"	"	1105CS2	5/2023	5000	4969	P/F	P/F
CCV	"	"	"					P/F	P/F

Conductivity Acceptance criteria $\pm 5\%$

pH	Name	Date	Time	Lot#	Expir.	pH	Temp	Meter	mV	Pass	Lab
FDEP SOP			CT-ET		Date	Buffer	°C	reading		Fail	Field
FT1100	Andrew Laering	2/8/22	719	213013	4/2023	7.0	22.9	7.01	-39.1	P/F	P/F
Calibr.	Drew Liddick	"	"	2110D46	9/2023	4.0	22.8	4.00	137.7	P/F	P/F
Calibr.	"	"	"	1106829	11/2022	10.0	22.7	10.03	-200.7	P/F	P/F
ICV	"	"	"	1106829	11/2022	10.0	22.7	9.99	-201.0	P/F	P/F
CCV	"	"	"	2110D46	9/2023	4.0	22.8	3.99	139.3	P/F	P/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 161.6, slope from 4 to 7 176.8 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 01/18/2022
 If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor	Name	Date	Time	Calibrated	ICV Value,	Pass /	Lab /
(Daily Calibration & ICV)			CT-ET	Value (0.00 or Offset), meters	meters	Fail	Field
Pressure mode in air	Andrew Laering	2/8/22	729	0.00	0.00	P/F	P/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: