

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: YSE 908

RQ: 2022-05-02-35

Project: 6 a m v i l l e

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 5/2/22

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J Pansh	5/9/22	730	19.6	761	9.2	9.2	100.1	—	1.02	(P) / F	(L) / F
ICV	J	5/9/22	731	19.6	761	9.2	9.2	100.1	—	1.02	(P) / F	(L) / F
CCV	J Pansh	5/9/22	1530	20.7	762	9.0	9.07	101.2	—	1.02	(P) / F	(L) / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Explr. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J Pansh	5/9/22	732	220225A	3/23	1000	1000	(P) / F	(L) / F
ICV	J	5/9/22	733	210417A	9/22	100	98.5	(P) / F	(L) / F
CCV	J Pansh	5/9/22	1533	211225	7/22	1413	1423	(P) / F	(L) / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Explr. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J Pansh	5/9/22	736	2112819	12/23	7.	19.6	7.02	-43.2	(P) / F	(L) / F
Calibr.	J		738	2201088	1/24	4.	19.9	4.00	137.2	(P) / F	(L) / F
Calibr.	J		740	2202163	7/23	10.	20.0	10.06	-207.3	(P) / F	(L) / F
ICV	J		742	2201088	1/24	4	20.0	3.98	136	(P) / F	(L) / F
CCV	J Pansh	5/9/22	1535	2202163	7/23	10	20.8	10.09	213	(P) / F	(L) / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.3, slope from 4 to 7 180 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J Pansh	5/9/22	747	0.00	0.00	(P) / F	(L) / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



DIVERSIFIED

ENVIRONMENTAL LABORATORIES, INC.

CHAIN OF CUSTODY RECORD

Project Number:

(Lab Use Only)

3653 Regent Boulevard, Suite # 509

Jacksonville, FL 32224

P. (904) 807-9625 E. info@delilab.com

F. (904) 807-9627

Page _____ of _____

Company Name Florida Department of Environmental Protection										Container Information (Qty, Size, & Type)								Sample Matrix Codes:			
Address 8800 Baymeadows Way W Ste 100										NPW NPW NPW								A. Air P. Petroleum B. Bulk S. Soil/Sludge W. Wipe O. Other _____ PW. Potable Water NPW. Non-Potable Water			
City Jacksonville State FL Zip 32256																					
Attn. Jeremy Parrish										Container Preservation Codes								Sample Kit or Container Lot #			
Tel. E-mail or Fax DEP-Overflowmicro@dep.state.fl.us										6 6 6											
Project Name Gainesville										Parameters								Preservation Codes:			
Project Number RQ-2022-05-02-35																		0. None 5. Na2S2O3 1. H2SO4 6. Crushed Ice (4C) 2. HNO3 7. C6H8O6 3. HCl 8. NH4Cl 4. NaOH 9. Other: (List)			
Project Location																		Thermal Preservation:			
Sampler Name/Signature Katrin Villinger																		In Crushed Ice? ☒ Yes ☐ No Cooler Temp: 0.3 °C Sample Cont. Temp: 0.9 °C			
Transfer #	Sample # (Lab Use Only)	Sample Identification	Date Sampled	Time Sampled	Sample Matrix Code	Specific Cond. (Field)	Comp	Grab		E. Coli - Quantitray	Enterococci - Quantitray	Coliform, Fecal - MF 3 Dilutions									
1		G1NE0948	5/9/22	1020	NPW	164		✓	1												
2		G1NE0949	5/9/22	1100	NPW	183		✓	1												
3		20020109	5/9/22	1125	NPW	223		✓	1												
4		G1NE0941	5/9/22	1345	NPW	180		✓	1												
5																					
6																					
7																					
8																					
9																					
0																					

 Certification #E821059	Special Instructions:	Transfer Numbers	Relinquished By	Date / Time	Received By	Date / Time
		1-4	Katrin Villinger	5/9/22 1500		05/09/22 1500