

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Shrimp Trap

RQ: 2023-08-21-31

Project: SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/3/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Barometer mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0729	23.6	759.5	8.5	8.49	99.9	/	0.994	P/F	L/F
ICV			0731	23.6	759.8	8.5	8.49	100.0	/	/	P/F	L/F
CCV	Victoria Schwartz	8/21/23	14:27	27.5	758.7	7.9	7.93	100.4	/	/	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0738	4306L79	06/25	1000	1000	P/F	L/F
ICV			0740	2207G95	07/24	100	100.3	P/F	L/F
CCV	Victoria Schwartz	8/21/23	14:29	2207G95	07/24	100	99.9	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0746	216233	08/23	7.00	24.1	7.00	-46.7	P/F	L/F
Calibr.			0750	231208	03/25	4.00	24.1	4.00	131.3	P/F	L/F
Calibr.			0753	222368	10/24	10.01	24.1	10.01	-225.4	P/F	L/F
ICV			0757	222368	10/24	10.01	24.2	9.99	-224.5	P/F	L/F
CCV	Victoria Schwartz	8/21/23	14:32	231208	03/25	4.00	23.8	4.11	125.3	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 179.7, slope from 4 to 7 178.0 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 7/3/23

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Light; Date of last verification: 5/17/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Terry Riordan	8/21/23	0757	0.272	0.271	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

2023-06-26-38

Boldly "X" this box if there are qualified data on this page.

Meter ID: Craw Daddy

RQ: 2023-07-03-50

Project: HAB

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification: 7/3/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0727	23.5	759.3	8.5	8.50	99.9	/	0.952	P/F	L/F
ICV			0730	23.5	759.6	8.5	8.50	100.0	/	/	P/F	L/F
CCV	Terry Riordan		1215	25.8	759.6	8.1	8.24	101.3	/	/	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75. Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0738	4306179	06/25	1000	1000	P/F	L/F
ICV			0741	2207695	07/24	100	100.7	P/F	L/F
CCV	Terry Riordan		1217	2207695	07/24	100	100.8	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Terry Riordan	8/21/23	0747	216233	08/23	7.00	24.0	7.00	-41.4	P/F	L/F
Calibr.			0751	231208	03/25	4.00	24.1	4.00	137.5	P/F	L/F
Calibr.			0755	222368	10/24	10.01	24.2	10.01	-221.2	P/F	L/F
ICV			0800	222368	10/24	10.01	24.2	9.99	-220.1	P/F	L/F
CCV	Terry Riordan		1220	231208	03/25	4.00	23.9	4.03	135.7	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ; If mV are recorded: slope from 7 to 10 179.8, slope from 4 to 7 178.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 7/3/23

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Knobhead; Date of last verification: 5/17/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Terry Riordan	8/21/23	0800	0.272	0.270	P/F	P/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: