

TREND KISS PARK

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Crawdad RQ- 2023-11-06-22 Project: Trend & Swamp

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0719	22.4	756.7	8.6	8.45	99.6		0.989	P/F	L/F
ICV			0722	22.4	756.7	8.6	8.45	99.8			P/F	L/F
CCV			1514	25.3	754.7	8.2	8.55	104			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec: Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0724	4306	6/25	1000	1000	P/F	L/F
ICV			0727	2207	7/24	100	100	P/F	L/F
CCV			1518			100	100.1	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0731	2310	5/25	7.01	23.4	7.01	-21.2	P/F	L/F
Calibr.			0734	2312	3/25	4.00	23.4	4.00	158.7	P/F	L/F
Calibr.			0739	2323	10/24	10.02	23.0	10.02	-194.9	P/F	L/F
ICV			0743	2323	10/24	10.0	22.9	9.97	-192.6	P/F	L/F
CCV			1522	2312	3/25	4.00	23.2	4.09	152.7	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 173.7 slope from 4 to 7 179.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/28/23

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: N/A; Date of last verification: N/A

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Jessie Atchison	10/30/23	0748	0.272	0.275	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Shrimp-trap

RQ-

Project:

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification: 9/28/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0719	22.5	756.9	8.6	8.63	99.6	/	1.11	P/F	L/F
ICV		1	0721	22.6	756.9	8.6	8.66	100.2	/		P/F	L/F
CCV											P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0725	4306L79	6/25	1000	1000	P/F	L/F
ICV		1	0729	2207095	7/24	100	100	P/F	L/F
CCV								P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jessie Atchison	10/30/23	0732	231678	5/25	7.01	23.4	7.01	53.6	P/F	L/F
Calibr.		1	0737	231208	3/25	4.00	23.4	4.00	126.10	P/F	L/F
Calibr.		1	0742	222368	10/24	10.02	22.9	10.02	229.50	P/F	L/F
ICV		1	0745			10.02	22.9	9.99	228	P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 75.9, slope from 4 to 7 180.2 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/28/23

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: crusty; Date of last verification: 5/17/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Jessie Atchison	10/30/23	0750	0.272	0.272	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 200102974
Dr. Phil's

RQ: 2023-07-17-26

Project: HAB

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/14/2023 7/13/2023

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Denver Barossa	10/30/23	0846	23.7	757.0	8.4	8.40	99.0		110275	P/F	L/F
ICV			0848	23.7	757.1	8.4	8.40	99.2			P/F	L/F
CCV			1238	23.1	756.5	8.5	8.78	102.8			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Denver Barossa	10/30/23	0851	4306L79	6/25	1000	1000.0	P/F	L/F
ICV			0854	2207695	7/24	100	101.0	P/F	L/F
CCV			1241			100	101.3	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Denver Barossa	10/30/23	0900	231678	5/25	7.01	23.5	7.01	-71.6	P/F	L/F
Calibr.			0903	231208	3/25	4.00	23.6	4.00	104.5	P/F	L/F
Calibr.			0907	222368	10/24	10.02	23.0	10.02	-246.0	P/F	L/F
ICV			0910			10.02	23.1	10.01	-245.4	P/F	L/F
CCV			1243	231208	3/25	4.00	23.0	4.06	100.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
If mV are recorded: slope from 7 to 10 174.4, slope from 4 to 7 176.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **(NO)** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____
If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Crusty; Date of last verification: 5/17/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: