

Chemical Analysis Report

CEN-DIST-2023-10-27-02

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Citrus Blue**
Request ID: **RQ-2023-10-23-08**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Matthew Bearden

For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 13-NOV-2023 09:28



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: CITRUS BLUE SPRING

Collection Date/Time: 10/26/2023 11:33

Field ID: G4CE0231

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2447605	DEP SOP: NU-094-1	Color (true)	2.5	U	PCU	P437092	
	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P437098	
	EPA 300.0 Rev. 2.1	Chloride	7.7		mg Cl/L	P437504	
		Sulfate	19		mg SO4/L	P437508	
	SM 2320 B-2011	Total Alkalinity	146		mg CaCO3/L	P437159	
	SM 2540 C-2015	TDS	200	A	mg/L	P437566	
	SM 2540 D-2015	TSS	2	UA	mg/L	P437417	
	SM 4500-F- C-2011	Fluoride	0.055	I	mg F/L	P437162	
2447606	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P437196	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.19	I	mg N/L	P437122	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.96		mg N/L	P437366	
	EPA 365.1 Rev. 2.0	Total-P	0.032		mg P/L	P437172	
	SM 5310 B-2011	Organic Carbon	0.50	U	mg C/L	P437267	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P437092

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P437098

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437504

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437508

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P437196

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P437122

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P437366

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P437172

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P437159

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2015
Batch ID: P437566

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P437417

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P437162

Component	Result	Code	Units
Fluoride	0.025	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P437267

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P437092

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P437098

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437504

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437508

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	105		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P437196

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P437122

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.4		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P437366

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P437172

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	108		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P437159

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	97.8		P	92 - 105

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2540 C-2015
Batch ID: P437566

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	98.8		P	85 - 115

Reference Method: SM 2540 D-2015
Batch ID: P437417

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	101		P	85 - 115

Reference Method: SM 4500-F- C-2011
Batch ID: P437162

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	103		P	94 - 106

Reference Method: SM 5310 B-2011
Batch ID: P437267

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.3		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437504

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447474	Chloride	105		P	90 - 110
2447605	Chloride	104		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437508

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447474	Sulfate	105		P	90 - 110
2447605	Sulfate	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P437196

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447606	Ammonia-N	102	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P437366

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447475	NO2NO3-N	97.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P437172

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447570	Total-P	106	109	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P437162

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447414	Fluoride	103	102	P/P	94 - 106

Reference Method: SM 5310 B-2011
Batch ID: P437267

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2447539	Organic Carbon	94.8	97.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P437092

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447568	Color (true)	12.5	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P437098

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447616	Turbidity	0.851	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437504

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447474	Chloride	0.107	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P437508

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447474	Sulfate	0.377	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P437196

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447606	Ammonia-N	0.392	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P437122

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447642	Kjeldahl Nitrogen	5.39	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P437366

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447475	NO2NO3-N	2.04	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P437172

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447570	Total-P	2.98	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P437159

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447414	Total Alkalinity	2.91	Sample	P	0 - 3.9

Reference Method: SM 2540 C-2015
Batch ID: P437566

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447605	TDS	6.00	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P437162

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447414	Fluoride	0.467	Spike	P	0 - 1.9

Reference Method: SM 5310 B-2011
Batch ID: P437267

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2447539	Organic Carbon	1.67	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: DEP SOP: NU-094-1
Run ID: A122416
Included Lab Sample IDs: 2447605

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.6	99.2	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A122419
Included Lab Sample IDs: 2447605

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.5	P/P	95 - 105

Reference Method: SM 2320 B-2011
Run ID: A122444
Included Lab Sample IDs: 2447605

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.8	96.0	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Run ID: A122444
Included Lab Sample IDs: 2447605

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A122464
Included Lab Sample IDs: 2447606

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.6	99.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A122491
Included Lab Sample IDs: 2447606

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: SM 5310 B-2011
Run ID: A122507
Included Lab Sample IDs: 2447606

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.3	96.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A122558
Included Lab Sample IDs: 2447606

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A122585

Included Lab Sample IDs: 2447606

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	91.9	92.7	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A122610

Included Lab Sample IDs: 2447605

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	103	102	P/P	90 - 110
Sulfate	103	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
DEP SOP: NU-094-1	Color (true)	100			12.5	
EPA 180.1 Rev. 2.0	Turbidity	104			0.851	
EPA 300.0 Rev. 2.1	Chloride	104	105	104	0.107	
	Sulfate	105	105	104	0.377	
EPA 350.1 Rev. 2.0	Ammonia-N	97.8	102	102		0.392
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.4				5.39
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	97.0	99.0		2.04
EPA 365.1 Rev. 2.0	Total-P	108	106	109		2.98
SM 2320 B-2011	Total Alkalinity	97.8			2.91	
SM 2540 C-2015	TDS	98.8			6.00	
SM 2540 D-2015	TSS	101				
SM 4500-F- C-2011	Fluoride	103	103	102		0.467
SM 5310 B-2011	Organic Carbon	95.3	94.8	97.3		1.67

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2447605
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2447605
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2447605
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2447605
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2447606
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2447606
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2447606
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2447606
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2447605
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2447605
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2447605
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2447605
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2447606

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	10/27/2023			10/27/2023 14:49	Chandler E Cox	2447605
EPA 180.1 Rev. 2.0	10/27/2023			10/27/2023 13:46	Samantha L Bates	2447605
EPA 300.0 Rev. 2.1	10/27/2023			11/07/2023 13:25	James L Waggenerby	2447605
EPA 350.1 Rev. 2.0	10/27/2023			10/31/2023 12:50	Ping Hua	2447606
EPA 351.2 Rev. 2.0	10/27/2023	10/30/2023 13:30	Simonne.V. Calhoun	11/01/2023 15:28	Samantha L Bates	2447606
EPA 353.2 Rev. 2.0	10/27/2023			11/02/2023 14:56	Fadila Guebre	2447606
EPA 365.1 Rev. 2.0	10/27/2023	10/31/2023 12:07	Adam P Silver	11/01/2023 13:32	Chandler E Cox	2447606
SM 2320 B-2011	10/27/2023			10/31/2023 03:52	Adam P Silver	2447605
SM 2540 C-2015	10/27/2023			10/31/2023 14:50	Yijie Li	2447605
SM 2540 D-2015	10/27/2023			10/31/2023 14:50	Yijie Li	2447605
SM 4500-F- C-2011	10/27/2023			10/31/2023 03:52	Adam P Silver	2447605
SM 5310 B-2011	10/27/2023			11/01/2023 22:28	Elise M. Gillis	2447606