

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Resident Canacho RQ- 2023-01-09-41 Project: SUP: Myakka Crix

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/10/22

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	William Miller	01/11/23	0649	22.2	765.9	8.8	8.80	100.9		1.028	(P) F	(L) F
ICV	William Miller	01/11/23	0652	22.2	765.9	8.8	8.78	101.0			(P) F	(L) F
CCV	William Miller	01/11/23	1222	21.9	766.2	8.8	8.88	100.7			(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	William Miller	01/11/23	0654	1107981	7/23	10,000	10,000	(P) F	(L) F
ICV	William Miller	01/11/23	0657	220309	3/23	50,000	50,225	(P) F	(L) F
CCV	William Miller	01/11/23	1224	2207695	7/24	100	101.2	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	William Miller	01/11/23	0659	2205254	5/24	7.00	22.7	7.00	✓	(P) F	(L) F
Calibr.	William Miller	01/11/23	0703	2207815	6/24	4.01	22.8	4.01	✓	(P) F	(L) F
Calibr.	William Miller	01/11/23	0707	4208577	2/24	10.04	22.4	10.04	✓	(P) F	(L) F
ICV	William Miller	01/11/23	0710	2207815	6/24	4.01	22.9	4.02	✓	(P) F	(L) F
CCV	William Miller	01/11/23	1227	4208577	2/24	10.03	22.6	10.03		(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 ✓, slope from 4 to 7 ✓ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **(NO)** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) **Secchi Disk Line** / Sonar Unique ID: 3312; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: