

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: President Camacho RQ: 2523-01-16-20 Project: Stadium Arcadia

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 11/18/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	11/19/23	654	22.0	764.2	8.8	8.80	100.5		1.030	(P) F (L) F	
ICV	11	11	656	22.0	764.2	8.8	8.80	100.4			(P) F (L) F	
CCV	Monica Jaeger	11/19/23	1436	27.4	762.8	7.9	8.00	101.0			(P) F (L) F	
CCV											P / F L / F	

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	11/19/23	658	2208A58	7/24	2500	2500	(P) F	(L) F
ICV	"	"	700	2207G95	7/24	100	100.1	(P) F	(L) F
CCV	Monica Jaeger	11/19/23	1438	2208A58	7/24	2500	2500	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	11/19/23	704	2205E54	5/24	7.01	22.3	7.01	—	(P) F (L) F	
Calibr.	11	11	705	2207B15	6/24	4.00	22.2	4.03	—	(P) F (L) F	
Calibr.	11	11	707	4208J17	2/24	10.03	22.2	10.03	—	(P) F (L) F	
ICV	11	11	713	2207B15	6/24	4.00	22.1	4.00		(P) F (L) F	
CCV	Monica Jaeger	11/19/23	1440	4208J17	2/24	10.03	23.1	10.07		(P) F (L) F	
CCV										P / F L / F	

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **NO** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) **Secchi Disk Line** / Sonar Unique ID: 3312A; Date of last verification: 8/29/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: