

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: PRESIDENT Camacho RQ- 2023-01-30-20 Project: SMP: Sheldon J.

- Notes: (1) Always wait for meter to stabilize before recording any readings. Plankton
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 01/18/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	01/30/23	0705	21.8	766.3	8.8	8.84	100.8	—	1.021	(P) F	(L) F
ICV			0708	22.2	766.3	8.8	8.78	100.9	—	—	(P) F	(L) F
CCV	William Heller	01/30/23	1414	22.0	764.7	8.8	8.73	99.9			(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	01/30/23	0712	2208A58	07/24	2500	2500	(P) F	(L) F
ICV			0716	2207GA5	07/24	100	101.4	(P) F	(L) F
CCV	William Heller	01/30/23	1416	2208A58	07/24	2500	2488	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	01/30/23	0720	2205E54	05/24	7.01	22.2	7.01	-38.1	(P) F	(L) F
Calibr.			0724	2207815	06/24	4.01	22.1	4.00	145.3	(P) F	(L) F
Calibr.			0722	4208J77	02/24	10.04	22.2	10.03	-206.3	(P) F	(L) F
ICV			0727	2207815	06/24	4.01	21.7	3.95	—	(P) F	(L) F
CCV	William Heller	01/30/23	1418	4208J77	02/24	10.03	22.8	10.03		(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 168.2, slope from 4 to 7 183.4 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 3312A; Date of last verification: 08/29/22

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: