

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: **PRO MOVE**

RQ: **2023-04-10-32**

Project: **SMP: VENICE**

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification **01/18/23**

DO FDEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	04/17/23	0728	21.4	754.6	8.8	8.77	99.3	—	1.063	(P) F	(L) F
ICV			0729	21.5	754.6	8.8	8.77	99.3	—	—	(P) F	(L) F
CCV	NICOLE REISTAD	04/17/23	1419	22.5	753.9	8.6	8.58	99.1	—	—	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec Cond FT 1200	Name	Date	Time CT-ET	Lot #	Expir Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	04/17/23	0731	2205H08	05/24	10,000	10,000	(P) F	(L) F
ICV			0732	2209K36	09/24	50,000	49,612	(P) F	(L) F
CCV	NICOLE REISTAD	04/17/23	1412	2207G95	07/24	100	104.2	(P) F	(L) F
CCV			1413	2209K36	09/24	50,000	49,354	(P) F	(L) F

Conductivity Acceptance criteria $\pm 5\%$

pH FDEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	NICOLE REISTAD	04/17/23	0741	2205E54	05/24	7.01	20.6	7.02	8.9	(P) F	(L) F
Calibr.			0743	1210718	09/24	4.01	21.3	4.00	173.2	(P) F	(L) F
Calibr.			0744	2208K61	02/24	10.04	21.5	10.04	-154.9	(P) F	(L) F
ICV			0747	1210718	09/24	4.01	21.9	4.00	—	(P) F	(L) F
CCV	NICOLE REISTAD	04/17/23	1416	2208K61	02/24	10.05	21.0	10.01	—	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 **163.8**, slope from 4 to 7 **164.3** (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **NO** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: **—**

If NO, what will be used? (circle one) **Secchi Disk Line** / Sonar Unique ID: **DEEP BLUE**; Date of last verification: **02/01/23**

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/ Fail	Lab/ Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: