

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: The One

RQ- 2023-06-05-06

Project: Caloosa Creek

Boldly "X" this box if there a qualified data this page.

- Notes: (1) Always wait for meter to stabilize before recording any readings. 2023-05-29-66
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration. PLAB-Response

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 04/19/23

DO	Name	Date	Time	Temp	Baro	D.O.	Meter	% DO	Probe	Probe	Pass	Lab
DEP SOP			CT-ET	°C	meter	Chart	D.O.		Charge	Gain	/	/
FT 1500					mmHg	mg/L	mg/L				Fail	Field
Calibr.	Thomas Baker	06/05/23	0706	21.9	756.6	8.7	8.74	99.6		1.0282	P/F	L/F
ICV	Thomas Baker	06/05/23	0707	21.8	756.6	8.7	8.74	99.6			P/F	L/F
CCV	Thomas Baker	06/05/23	1407	26.1	757.3	8.0	7.97	99.3			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec	Name	Date	Time	Lot #	Expir.	Standard	Meter	Pass	Lab
Cond.			CT-ET		Date	µmhos/cm	Reading	/	/
FT 1200							µmhos/cm	Fail	Field
Calibr.	Thomas Baker	06/05/23	0709	4207M20	07/24	2500	2500	P/F	L/F
ICV	Thomas Baker	06/05/23	0711	2207645	07/24	100	102.3	P/F	L/F
CCV	Thomas Baker	06/05/23	1409	2205H28	5/24	210000	9948	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH	Name	Date	Time	Lot #	Expir.	pH	Temp	Meter	mV	Pass	Lab
DEP SOP			CT-ET		Date	Buffer	°C	reading		/	/
FT 1100						SU		SU		Fail	Field
Calibr.	Thomas Baker	06/05/23	0716	2208634	8/24	7.02	21.4	7.02	-44.8	P/F	L/F
Calibr.			0717	2301612	1/25	4.00	21.6	4.00	135.3	P/F	L/F
Calibr.			0718	2302171	8/24	10.04	21.5	10.04	-206.7	P/F	L/F
ICV	Thomas Baker	06/05/23	0720	2301622	7/25	4.00	21.7	3.99		P/F	L/F
CCV	Thomas Baker	06/05/23	1411	2302171	8/24	10.04	21.9	10.02		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 162.1, slope from 4 to 7 180.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: Slippery 10-25 Date of last verification: 02/01/23

Depth Sensor	Name	Date	Time	Calibrated	ICV Value	Pass	Lab
(Daily Calibration & ICV)			CT-ET	Value (0.00 or Offset), meters	meters	Fail	Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or $\pm 0.05m$, whichever is greater.

COMMENTS: