

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

President Camacho

RQ-

2023-09-25-21

Project:

SMP: Sarasota

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/17/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	9/25/23	654	20.9	760.5	8.94	8.94	100.1	—	1.024	P/F	L/F
ICV	11	11	656	21.0	760.5	8.92	8.93	100.3	—	—	P/F	L/F
CCV	Nicole Reistad	09/25/23	1610	21.4	760.3	8.85	8.87	100.3	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	9/25/23	700	22126066	11/24	10000	10000	P/F	L/F
ICV	11	11	702	2207695	7/24	100	101.3	P/F	L/F
CCV	Nicole Reistad	09/25/23	1603	230217A	02/24	50,000	49,648	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Monica Jaeger	9/25/23	704	230607B	12/24	7.02	20.6	7.02	-54.4	P/F	L/F
Calibr.	11	11	706	230607D	12/24	4.00	20.7	4.00	120.3	P/F	L/F
Calibr.	11	11	708	230607C	12/24	10.05	20.6	10.05	218.6	P/F	L/F
ICV	11	11	710	230607D	12/24	4.00	20.7	3.98	—	P/F	L/F
CCV	Nicole Reistad	09/25/23	1607	230607C	12/24	10.04	21.5	10.04	10.02	P/F	L/F
CCV								10.03	NRNR	P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 164.2 slope from 4 to 7 174.7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 3312A ; Date of last verification: 8/14/23

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: