


BENCHMARK EA SOUTH



NELAC Certification #E85086

ANALYTICAL TEST REPORT

THESE RESULTS MEET NELAC STANDARDS

Submission Number : S23050227

Florida Dept Of Env Protection
2600 Blair Stone Rd
Tallahassee, FL 32399

Project Name : RQ-2023-05-01-46
Date Received : 05/09/2023
Time Received : 11:45

Submission Number: 23050227 Sample Date: 05/09/2023
Sample Number: 001 Sample Time: 09:00
Sample Description: G3SD0100 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	88	#/100 ML	10	10	SM9223B	05/09/2023 15:47	05/10/2023 16:27	BK/MM


Submission Number: 23050227 Sample Date: 05/09/2023
Sample Number: 002 Sample Time: 09:50
Sample Description: G3SD0097 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	269	#/100 ML	10	10	SM9223B	05/09/2023 15:47	05/10/2023 16:27	BK/MM

Submission Number: 23050227 Sample Date: 05/09/2023
Sample Number: 003 Sample Time: 10:45
Sample Description: G2SD0032 Sample Method: Grab

Parameter	Result	Units	MDL	PQL	Procedure	Preparation Date/Time	Analysis Date/Time	Analyst
E- COLI BY IDEXX QUANTITRAY	135	#/100 ML	10	10	SM9223B	05/09/2023 15:47	05/10/2023 16:27	BK/MM

BENCHMARK EA SOUTH



NELAC Certification #E85086

Melinda Merchant

05/15/2023

Dale D. Dixon / Laboratory Director

Date

Melinda Merchant/ Laboratory Manager

DATA QUALIFIERS THAT MAY APPLY:

A = Value reported is an average of two or more determinations.
B = Results based upon colony counts outside the Ideal range.
H = Value based on field kit determination. Results may not be accurate.
I = Reported value is between the laboratory MDL and the PQL.
J1 = Estimated value. Surrogate recovery limits exceeded.
J2 = Estimated value. No quality control criteria exists for component.
J3 = Estimated value. Quality control criteria for precision or accuracy not met.
J4 = Estimated value. Sample matrix interference suspected.
J5 = Estimated value. Data questionable due to Improper lab or field protocols.
K = Off-scale low. Value is known to be < the value reported.
L = Off-scale high. Value is known to be > the value reported.
N = Presumptive evidence of presence of material.
O = Sampled, but analysis lost or not performed.

Q = Sample held beyond accepted hold time.
T = Value reported is < MDL. Reported for informational purposes only and shall not be used in statistical analysis.
U = Analyte analyzed but not detected at the value indicated.
V = Analyte detected in sample and method blank. Results for this analyte in associated samples may be biased high. Standard, Duplicate and Spike values are within control limits. Reported data are usable.
Y = Analysis performed on an improperly preserved sample. Data may be inaccurate.
Z = Too many colonies were present (TNTC). The numeric value represents the filtration volume.
! = Data deviate from historically established concentration ranges.
? = Data rejected and should not be used. Some or all of QC data were outside criteria, and the presence or absence of the analyte cannot be determined from the data.
* = Not reported due to interference.

NOTES:

J0 = All Quanti-tray wells were positive. Result may be higher than reported.
PQL = 4xMDL.
ND = Not detected at or above the adjusted reporting limit.
X = Value exceeds MCL.
1 dilution (10 mL) on each Enterococci and E.coli sample and 3 dilutions (1 mL, 10 mL, & 100 mL) on each Fecal Coliform sample.

COMMENTS:

For questions or comments regarding these results, please contact us at (941) 625-3137.

Results relate only to the samples.

Benchmark EnviroAnalytical, Inc.
 1001 Corporate Ave. Suite 102
 North Port, FL 34289
 (941) 625-3137 / (941) 240-3071 fax
 2023-05-01-46

BEAS Sample Receipt Temp. 3.1 °C
Checked at Benchmark EA South E85068
With Temperature Gun ID #7

Client Name: FDEP
Address: 2600 Blair Stone Rd.
 Tallahassee, FL, 32399
Phone: (239) 344-5719
Email: dep-overflowmicro@dep.state.fl.us

Bill to: Frances Moore
P.O. #: Lab B95341

Project Name: RQ-2024
Project Run/Location: SNAP / later gator

Laboratory Submission #:

523050227

Station ID
 Sample Type

Enterococci
 E.coli (SM9223B Quantitray)

Laboratory Sample #

E.Coli = 250 mL ; Entero = 120 mL ; Fecal Coliform = 120 mL

G3SD0100	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time: 5/9/23	900	1
G3SD0097	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time: 5/9/23	950	2
G2SD0032	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time: 5/9/23	1045	3
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		
	F / M / Grab	E. coli / Entero / Fecal coliform	Date & Time:		

Laboratory Sample Acceptability:
 pH < 2 : 0 BEAS Temperature: 3.1 °C

1. Sample Type is used to indicate whether the sample is being designated to drinking water (DW), groundwater (GW), surface water (SW), fresh surface water (FSW), saline surface water (SSW), soil sediment (SSA), or sludge (SLD).
2. Sample Matrix is used to indicate whether the sample is plastic (P) or glass (G).
3. Container Type is used to indicate whether the container is plastic (P) or glass (G).
4. Sample must be refrigerated or stored in wet ice after collection. The maximum temperature during storage should be 4°C (42°F). Under "Preservation," list any preservatives that were added to the sample container.
5. The client is responsible for documentation of the sampling event. Please note: special sampling events on the sample copy form.

1	Collector: <i>MONCK Jaeger</i>	Date: 5/9/23	Time: 1145	Received By: <i>Melinda McPherson</i>	Date: 5/9/23	Time: 1145
2	Relinquished by:	Date:	Time:	Received By:	Date:	Time:
3	Relinquished by:	Date:	Time:	Received By:	Date:	Time:
4	Relinquished by:	Date:	Time:	Received By:	Date:	Time:
5	Relinquished by:	Date:	Time:	Received By:	Date:	Time: