

Chemical Analysis Report

SO-DIST-2023-06-22-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **2023 SMP: Goin' Coastal**
Request ID: **RQ-2023-06-26-09**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 314
Fort Myers, FL 33901-3381
Attn: Monica Jaeger

For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Kerry Tate, Ph.D., Environmental Administrator

Date Certified: 11-JUL-2023 07:53

A handwritten signature in black ink that reads "Kerry Tate". The signature is written in a cursive, flowing style with a horizontal line extending from the end of the name.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Punta Rassa Cove

Collection Date/Time: 06/21/2023 10:10

Field ID: G2SD0047

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2419323	DEP SOP: NU-094-1	Color (true)	13		PCU	P431665	
	EPA 180.1 Rev. 2.0	Turbidity	3.3		NTU	P431664	
	EPA 300.0 Rev. 2.1	Chloride	1.6E+04		mg Cl/L	P431914	
		Sulfate	2.2E+03		mg SO4/L	P431833	
	SM 2320 B-2011	Total Alkalinity	120		mg CaCO3/L	P431797	
	SM 2540 C-2015	TDS	2.56E+04		mg/L	P432044	
	SM 2540 D-2015	TSS	7	IA	mg/L	P432048	
	SM 4500-F- C-2011	Fluoride	1.0		mg F/L	P431809	RPD
2419324	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P431801	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.52		mg N/L	P431750	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P431989	
	EPA 365.1 Rev. 2.0	Total-P	0.050		mg P/L	P431779	
	SM 5310 B-2011	Organic Carbon	7.4		mg C/L	P432072	

Ref. Method and Comment:

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P431665

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P431664

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431833

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431914

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P431801

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P431750

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P431989

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P431779

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: SM 2320 B-2011
Batch ID: P431797

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-2015
Batch ID: P432044

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P432048

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P431809

Component	Result	Code	Units
Fluoride	0.025	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P432072

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P431665

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.5		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P431664

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	99.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431833

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431914

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P431801

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	94.4		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P431750

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P431989

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P431779

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	99.4		P	90 - 110

Reference Method: SM 2320 B-2011
Batch ID: P431797

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	94.3		P	92 - 105

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 2540 C-2015
Batch ID: P432044

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	96.3		P	90 - 110

Reference Method: SM 2540 D-2015
Batch ID: P432048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	95.3		P	85 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P431809

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	100		P	94 - 106

Reference Method: SM 5310 B-2011
Batch ID: P432072

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.9		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431833

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2418864	Sulfate	101		P	90 - 110
2419270	Sulfate	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431914

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2418864	Chloride	104		P	90 - 110
2419270	Chloride	106		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P431801

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2418719	Ammonia-N	99.6	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P431750

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2418718	Kjeldahl Nitrogen	112	105	F/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P431989

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2419247	NO2NO3-N	106	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P431779

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2419115	Total-P	96.7	98.2	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Batch ID: P431809

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2415436	Fluoride	100	103	P/P	94 - 106

Reference Method: SM 5310 B-2011
Batch ID: P432072

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2419513	Organic Carbon	92.9	93.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P431665

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419221	Color (true)	1.07	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P431664

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419268	Turbidity	3.15	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431833

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2418864	Sulfate	1.22	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P431914

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2418864	Chloride	0.348	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P431801

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2418719	Ammonia-N	0.735	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P431750

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2418718	Kjeldahl Nitrogen	5.06	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P431989

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419247	NO2NO3-N	0.948	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P431779

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419115	Total-P	1.48	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
Batch ID: P431797

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419512	Total Alkalinity	1.59	Sample	P	0 - 3.9

Reference Method: SM 2540 C-2015
Batch ID: P432044

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419404	TDS	4.51	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011
Batch ID: P431809

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2415436	Fluoride	1.91	Spike	F	0 - 1.9

Reference Method: SM 5310 B-2011
Batch ID: P432072

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2419513	Organic Carbon	0.742	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A119978

Included Lab Sample IDs: 2419323

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	101	P/P	95 - 105

Reference Method: DEP SOP: NU-094-1

Run ID: A119979

Included Lab Sample IDs: 2419323

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	100	P/P	90 - 110

Reference Method: SM 2320 B-2011

Run ID: A120040

Included Lab Sample IDs: 2419323

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	95.2	92.4	P/P	90 - 110

Reference Method: SM 4500-F- C-2011

Run ID: A120040

Included Lab Sample IDs: 2419323

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A120041

Included Lab Sample IDs: 2419324

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.9	97.9	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A120051

Included Lab Sample IDs: 2419323

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	102	105	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A120053

Included Lab Sample IDs: 2419324

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.1	93.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A120078

Included Lab Sample IDs: 2419324

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	99.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A120124

Included Lab Sample IDs: 2419324

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A120177

Included Lab Sample IDs: 2419324

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	96.3	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
DEP SOP: NU-094-1	Color (true)	99.5			1.07	
EPA 180.1 Rev. 2.0	Turbidity	99.5			3.15	
EPA 300.0 Rev. 2.1	Chloride	99.5	104	106	0.348	
	Sulfate	103	101	105	1.22	
EPA 350.1 Rev. 2.0	Ammonia-N	94.4	99.6	101		0.735
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	112	105		5.06
EPA 353.2 Rev. 2.0	NO2NO3-N	105	106	105		0.948
EPA 365.1 Rev. 2.0	Total-P	99.4	96.7	98.2		1.48
SM 2320 B-2011	Total Alkalinity	94.3			1.59	
SM 2540 C-2015	TDS	96.3			4.51	
SM 2540 D-2015	TSS	95.3				
SM 4500-F- C-2011	Fluoride	100	100	103		1.91
SM 5310 B-2011	Organic Carbon	95.9	92.9	93.7		0.742

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2419323
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2419323
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2419323
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2419323
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2419324
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2419324
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2419324
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2419324
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2419323
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2419323
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2419323
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2419323
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2419324

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	06/22/2023			06/22/2023 15:09	Anna M. Plaviak	2419323
EPA 180.1 Rev. 2.0	06/22/2023			06/22/2023 13:33	Khloe J Berg	2419323
EPA 300.0 Rev. 2.1	06/22/2023			06/27/2023 06:01	James L Waggeberby	2419323
	06/22/2023			06/28/2023 14:06	James L Waggeberby	2419323
EPA 350.1 Rev. 2.0	06/22/2023			06/26/2023 11:49	Khloe J Berg	2419324
EPA 351.2 Rev. 2.0	06/22/2023	06/26/2023 14:39	Dana G. Hughes	06/27/2023 12:51	Ping Hua	2419324
EPA 353.2 Rev. 2.0	06/22/2023			06/30/2023 11:01	Beverly Y. Sanders	2419324
EPA 365.1 Rev. 2.0	06/22/2023	06/27/2023 13:15	Simonne.V. Calhoun	06/28/2023 15:12	Simonne.V. Calhoun	2419324
SM 2320 B-2011	06/22/2023			06/27/2023 02:28	Adam P Silver	2419323
SM 2540 C-2015	06/22/2023			06/23/2023 15:19	Yijie Li	2419323
SM 2540 D-2015	06/22/2023			06/23/2023 15:19	Yijie Li	2419323
SM 4500-F- C-2011	06/22/2023			06/27/2023 00:47	Adam P Silver	2419323
SM 5310 B-2011	06/22/2023			07/06/2023 02:27	Elise M. Gillis	2419324