

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: 155077

RQ: 2023-07-24

Project: SMP

~~Boldly "X" this box if there are quality data on this page.~~

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	B.Olin	7/24	0722	23.0	762.7	8.6	8.61	100.5		1.12	P/F	L/F
ICV	I		0731	23.0	762.7	8.6	8.62	100.7			P/F	L/F
CCV	M. Butler		1245	22.4	763.4	8.7	9.1	105.7			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	B.Olin	7/24	0733	221221B	01/24	1000	1000	P/F	L/F
ICV	I			221221A	01/24	100	99.6	P/F	L/F
CCV			1157	221221B	1/24	1000	996	P/F	L/F
CCV	B.Olin		1216	220425	11/23	50,000	50,000	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT.1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	B.Olin	07/24		1LD0599	12/23	7.	22.4	7.1	-35.7	P/F	L/F
Calibr.				1L30499	12/23	4.	22.1	4.0	-137.9	P/F	L/F
Calibr.				1LF0738	8/23	10.	22.1	9.9	-196.7	P/F	L/F
ICV				1L30999	12/23	10				P/F	L/F
CCV				1LF0738	8/23	4	22.3	4.0	-135.5	P/F	L/F
CCV				1LF0738	8/23	10.0	21.7	9.98	-196.6	P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	B.Olin	7/24	0720	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: