

Chemical Analysis Report

NW-DIST-2023-03-15-02

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Western Lake Run**
Request ID: **RQ-2023-03-13-26**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government Street
Suite 208
Pensacola, FL 32502-5794
Attn: Myranda K Butler

For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 30-MAR-2023 11:36

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Pond Creek West of Headwaters of Blackwater Bay **Collection Date/Time: 03/14/2023 12:20**

Field ID: G4NW0277

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2393968	EPA 8270E	Aldrin	0.72	U	ng/L	P427174	MS, RPD
		Alpha-BHC	0.11	U	ng/L	P427174	
		Alpha-Chlordane	0.22	U	ng/L	P427174	
		PCB-1016	2.2	U	ng/L	P427174	
		Beta-BHC	0.11	U	ng/L	P427174	
		PCB-1221	2.2	U	ng/L	P427174	
		Beta-Cyfluthrin	1.4	U	ng/L	P427174	
		PCB-1232	2.2	U	ng/L	P427174	
		Bifenthrin	0.56	U	ng/L	P427174	
		PCB-1242	2.2	U	ng/L	P427174	
		Carbophenothion	1.0	U	ng/L	P427174	
		PCB-1248	2.2	U	ng/L	P427174	
		Chlorothalonil	4.0	U	ng/L	P427174	
		PCB-1254	2.2	U	ng/L	P427174	
		Cis-Nonachlor	0.22	U	ng/L	P427174	
		PCB-1260	2.2	U	ng/L	P427174	
		Cypermethrin	1.7	U	ng/L	P427174	
		Dacthal	0.17	U	ng/L	P427174	
		DDD-p,p'	0.28	U	ng/L	P427174	
		DDE-p,p'	0.22	U	ng/L	P427174	
		DDT-p,p'	0.28	U	ng/L	P427174	
		Delta-BHC	0.45	U	ng/L	P427174	
		Deltamethrin	1.4	U	ng/L	P427174	
		Dichlobenil	0.28	U	ng/L	P427174	
		Dicofol	1.4	U	ng/L	P427174	
		Dieldrin	0.45	U	ng/L	P427174	
		Endosulfan I	0.45	U	ng/L	P427174	
		Endosulfan II	0.45	U	ng/L	P427174	
		Endosulfan Sulfate	0.33	U	ng/L	P427174	
		Endrin	0.45	U	ng/L	P427174	
		Endrin Aldehyde	0.84	U	ng/L	P427174	
		Endrin Ketone	0.45	U	ng/L	P427174	
		Esfenvalerate	0.84	U	ng/L	P427174	
		Ethalfuralin	0.17	U	ng/L	P427174	
		Fenpropathrin	0.51	I	ng/L	P427174	
		Gamma-BHC	0.14	U	ng/L	P427174	RPD
		Gamma-Chlordane	0.22	U	ng/L	P427174	
		Heptachlor	0.22	U	ng/L	P427174	
		Heptachlor Epoxide	0.28	U	ng/L	P427174	
		Hexachlorobenzene**	1.1	U	ng/L	P427174	
		Kepone	5.6	U	ng/L	P427174	
		Lambda-Cyhalothrin	0.84	U	ng/L	P427174	
		Methoprene	0.84	U	ng/L	P427174	

Field ID: G4NW0277

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2393968	EPA 8270E	Methoxychlor	0.33	U	ng/L	P427174	
		MGK-264	0.22	U	ng/L	P427174	
		Mirex	0.39	U	ng/L	P427174	
		Oxychlordane	0.33	U	ng/L	P427174	
		Permethrin	1.8	U	ng/L	P427174	
		Phenothrin	1.0	U	ng/L	P427174	
		Piperonyl Butoxide	0.17	U	ng/L	P427174	
		Prallethrin	2.2	U	ng/L	P427174	
		Tau-Fluvalinate	0.76	I	ng/L	P427174	
		Tetramethrin	0.84	U	ng/L	P427174	
		Trans-Nonachlor	0.22	U	ng/L	P427174	
		Triclosan Methyl	0.17	U	ng/L	P427174	
		Trifluralin	0.22	U	ng/L	P427174	
	EPA 8276	Chlordane	2.2	U	ng/L	P427174	
		Toxaphene	17	U	ng/L	P427174	

Sample Location: CRESCENT LAKE

Collection Date/Time: 03/14/2023 11:05

Field ID: G5NW0159

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2393964	DEP SOP: NU-094-1	Color (true)	25		PCU	P427216	
	EPA 180.1 Rev. 2.0	Turbidity	2.9		NTU	P427201	
	EPA 300.0 Rev. 2.1	Chloride	8.9		mg Cl/L	P427727	
		Sulfate	5.3		mg SO4/L	P427730	
	SM 2320 B-2011	Total Alkalinity	26		mg CaCO3/L	P427349	RPD
	SM 2540 C-2015	TDS	53		mg/L	P427644	
	SM 2540 D-2015	TSS	2	U	mg/L	P427537	
	SM 4500-F- C-2011	Fluoride	0.026	I	mg F/L	P427356	
2393965	EPA 350.1 Rev. 2.0	Ammonia-N	0.064		mg N/L	P427633	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.30		mg N/L	P427433	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.24		mg N/L	P427398	
	EPA 365.1 Rev. 2.0	Total-P	0.008		mg P/L	P427333	
	SM 5310 B-2011	Organic Carbon	3.1		mg C/L	P427625	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-2015: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500-F- C-2011: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P427216

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P427201

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427727

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427730

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P427433

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427398

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427333

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 8270E
Batch ID: P427174

Component	Result	Code	Units
Aldrin	0.65	U	ng/L
Aldrin	0.65	U	ng/L
Alpha-BHC	0.10	U	ng/L
Alpha-BHC	0.10	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P427174

Component	Result	Code	Units
Alpha-Chlordane	0.20	U	ng/L
Alpha-Chlordane	0.20	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-BHC	0.10	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Beta-Cyfluthrin	1.2	U	ng/L
Bifenthrin	0.50	U	ng/L
Bifenthrin	0.50	U	ng/L
Carbophenothion	0.90	U	ng/L
Carbophenothion	0.90	U	ng/L
Chlorothalonil	3.6	U	ng/L
Chlorothalonil	3.6	U	ng/L
Cis-Nonachlor	0.20	U	ng/L
Cis-Nonachlor	0.20	U	ng/L
Cypermethrin	1.5	U	ng/L
Cypermethrin	1.5	U	ng/L
Dacthal	0.15	U	ng/L
Dacthal	0.15	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDD-p,p'	0.25	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDE-p,p'	0.20	U	ng/L
DDT-p,p'	0.25	U	ng/L
DDT-p,p'	0.25	U	ng/L
Delta-BHC	0.40	U	ng/L
Delta-BHC	0.40	U	ng/L
Deltamethrin	1.2	U	ng/L
Deltamethrin	1.2	U	ng/L
Dichlobenil	0.25	U	ng/L
Dichlobenil	0.25	U	ng/L
Dicofol	1.2	U	ng/L
Dicofol	1.2	U	ng/L
Dieldrin	0.40	U	ng/L
Dieldrin	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan I	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan II	0.40	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endosulfan Sulfate	0.30	U	ng/L
Endrin	0.40	U	ng/L
Endrin	0.40	U	ng/L
Endrin Aldehyde	0.75	U	ng/L
Endrin Aldehyde	0.75	U	ng/L
Endrin Ketone	0.40	U	ng/L
Endrin Ketone	0.40	U	ng/L
Esfenvalerate	0.75	U	ng/L
Esfenvalerate	0.75	U	ng/L
Ethalfuralin	0.15	U	ng/L
Ethalfuralin	0.15	U	ng/L
Fenpropathrin	0.25	U	ng/L
Fenpropathrin	0.25	U	ng/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8270E

Batch ID: P427174

Component	Result	Code	Units
Gamma-BHC	0.12	U	ng/L
Gamma-BHC	0.12	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Gamma-Chlordane	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor	0.20	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Heptachlor Epoxide	0.25	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Hexachlorobenzene	1.0	U	ng/L
Kepone	5.0	U	ng/L
Kepone	5.0	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L
Lambda-Cyhalothrin	0.75	U	ng/L
Methoprene	0.75	U	ng/L
Methoprene	0.75	U	ng/L
Methoxychlor	0.30	U	ng/L
Methoxychlor	0.30	U	ng/L
MGK-264	0.20	U	ng/L
MGK-264	0.20	U	ng/L
Mirex	0.35	U	ng/L
Mirex	0.35	U	ng/L
Oxychlordane	0.30	U	ng/L
Oxychlordane	0.30	U	ng/L
PCB-1016	2.0	U	ng/L
PCB-1016	2.0	U	ng/L
PCB-1221	2.0	U	ng/L
PCB-1221	2.0	U	ng/L
PCB-1232	2.0	U	ng/L
PCB-1232	2.0	U	ng/L
PCB-1242	2.0	U	ng/L
PCB-1242	2.0	U	ng/L
PCB-1248	2.0	U	ng/L
PCB-1248	2.0	U	ng/L
PCB-1254	2.0	U	ng/L
PCB-1254	2.0	U	ng/L
PCB-1260	2.0	U	ng/L
PCB-1260	2.0	U	ng/L
Permethrin	1.6	U	ng/L
Permethrin	1.6	U	ng/L
Phenothrin	0.90	U	ng/L
Phenothrin	0.90	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Piperonyl Butoxide	0.15	U	ng/L
Prallethrin	2.0	U	ng/L
Prallethrin	2.0	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tau-Fluvalinate	0.25	U	ng/L
Tetramethrin	0.75	U	ng/L
Tetramethrin	0.75	U	ng/L
Trans-Nonachlor	0.20	U	ng/L
Trans-Nonachlor	0.20	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270E
Batch ID: P427174

Component	Result	Code	Units
Triclosan Methyl	0.15	U	ng/L
Triclosan Methyl	0.15	U	ng/L
Trifluralin	0.20	U	ng/L
Trifluralin	0.20	U	ng/L

Reference Method: EPA 8276
Batch ID: P427174

Component	Result	Code	Units
Chlordane	2.0	U	ng/L
Chlordane	2.0	U	ng/L
Toxaphene	15	U	ng/L
Toxaphene	15	U	ng/L

Reference Method: SM 2320 B-2011
Batch ID: P427349

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-2015
Batch ID: P427644

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2015
Batch ID: P427537

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500-F- C-2011
Batch ID: P427356

Component	Result	Code	Units
Fluoride	0.025	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P427625

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P427216

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P427201

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Turbidity	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427727

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427730

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	93.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P427433

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427398

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427333

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 8270E
Batch ID: P427174

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	35.9		P	10 - 92.0
Alpha-BHC	84.1		P	75 - 107
Alpha-Chlordane	53.9		P	50 - 108
Beta-BHC	72.5		P	36 - 138

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8270E
Batch ID: P427174

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Beta-Cyfluthrin	50.7		P	20 - 161
Bifenthrin	43.5		P	10 - 119
Carbophenothion	36.6		P	19 - 94.0
Chlorothalonil	41.9		P	26 - 186
Cis-Nonachlor	56.6		P	42 - 119
Cypermethrin	51.1		P	22 - 150
Dacthal	92.6		P	72 - 119
DDD-p,p'	58.5		P	45 - 107
DDE-p,p'	55.4		P	48 - 106
DDT-p,p'	66.3		P	48 - 110
Delta-BHC	75.7		P	54 - 140
Deltamethrin	54.3		P	22 - 189
Dichlobenil	56.7		P	36 - 117
Dicofol	51.8		P	46 - 155
Dieldrin	62.9		P	54 - 110
Endosulfan I	64.1		P	59 - 105
Endosulfan II	71.6		P	51 - 118
Endosulfan Sulfate	71.7		P	57 - 121
Endrin	62.2		P	10 - 120
Endrin Aldehyde	59.0		P	34 - 118
Endrin Ketone	87.5		P	69 - 177
Esfenvalerate	43.6		P	23 - 168
Ethalfuralin	54.2		P	49 - 99.0
Fenpropathrin	48.3		P	34 - 117
Gamma-BHC	88.5		P	72 - 117
Gamma-Chlordane	55.5		P	43 - 106
Heptachlor	50.8		P	41 - 98.0
Heptachlor Epoxide	79.1		P	57 - 106
Hexachlorobenzene	40.1		P	36 - 99.0
Kepone	125		P	16 - 215
Lambda-Cyhalothrin	36.0		P	21 - 177
Methoprene	35.2		P	10 - 119
Methoxychlor	61.0		P	60 - 112
MGK-264	60.5		P	26 - 122
Mirex	40.1		P	10 - 169
Oxychlordane	54.7		P	43 - 105
PCB-1016	76.9		P	48 - 112
PCB-1260	76.1		P	44 - 118
Permethrin	42.7		P	24 - 148
Phenothrin	31.6		P	10 - 106
Piperonyl Butoxide	75.0		P	10 - 139
Prallethrin	59.5		P	17 - 109
Tau-Fluvalinate	37.8		P	13 - 131
Tetramethrin	62.1		P	10 - 132
Trans-Nonachlor	55.4		P	44 - 106
Triclosan Methyl	64.9		P	59 - 109
Trifluralin	56.1		P	44 - 101

Reference Method: EPA 8276
Batch ID: P427174

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8276
Batch ID: P427174

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	80.5		P	61 - 116
Toxaphene	82.9		P	48 - 124

Reference Method: SM 2320 B-2011
Batch ID: P427349

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	95.8		P	94 - 106

Reference Method: SM 2540 C-2015
Batch ID: P427644

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TDS	105		P	90 - 110

Reference Method: SM 2540 D-2015
Batch ID: P427537

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
TSS	99.0		P	85 - 106

Reference Method: SM 4500-F- C-2011
Batch ID: P427356

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	103		P	96 - 104

Reference Method: SM 5310 B-2011
Batch ID: P427625

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427727

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2394077	Chloride	97.0		P	90 - 110
2394243	Chloride	97.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427730

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2394077	Sulfate	95.4		P	90 - 110
2394243	Sulfate	96.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395549	Ammonia-N	97.8	98.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P427433

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2393942	Kjeldahl Nitrogen	97.5	99.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427398

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2393849	NO2NO3-N	99.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427333

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2394042	Total-P	103	104	P/P	90 - 110

Reference Method: EPA 8270E
Batch ID: P427174

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2393976	Aldrin	38.4	39.6	P/P	20 - 82.0
2393976	Alpha-BHC	96.4	47.9	P/F	71 - 109
2393976	Alpha-Chlordane	59.5	59.1	P/P	42 - 106
2393976	Beta-BHC	80.8	74.4	P/P	69 - 180
2393976	Beta-Cyfluthrin	59.9	71.8	P/P	18 - 175
2393976	Bifenthrin	59.3	69.3	P/P	21 - 128
2393976	Carbophenothion	39.6	44.0	P/P	10 - 126
2393976	Chlorothalonil	40.1	28.6	P/P	10 - 300
2393976	Cis-Nonachlor	55.6	65.1	P/P	34 - 106
2393976	Cypermethrin	61.4	77.0	P/P	22 - 158
2393976	Dacthal	115	105	P/P	54 - 116
2393976	DDD-p,p'	63.5	62.5	P/P	34 - 107
2393976	DDE-p,p'	64.3	65.8	P/P	33 - 110
2393976	DDT-p,p'	72.3	76.7	P/P	50 - 122

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270E

Batch ID: P427174

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2393976	Delta-BHC	79.7	78.2	P/P	77 - 193
2393976	Deltamethrin	67.5	90.1	P/P	10 - 195
2393976	Dichlobenil	48.7	51.4	P/P	25 - 113
2393976	Dicofol	56.1	52.6	P/P	38 - 247
2393976	Dieldrin	66.9	71.1	P/P	45 - 118
2393976	Endosulfan I	66.1	69.8	P/P	51 - 106
2393976	Endosulfan II	70.4	73.1	P/P	37 - 122
2393976	Endosulfan Sulfate	72.6	67.3	P/P	43 - 132
2393976	Endrin	69.4	71.4	P/P	29 - 101
2393976	Endrin Aldehyde	62.3	51.6	P/P	19 - 110
2393976	Endrin Ketone	97.1	93.7	P/P	60 - 140
2393976	Esfenvalerate	53.8	74.2	P/P	26 - 199
2393976	Ethalfuralin	59.3	57.4	P/P	45 - 99.0
2393976	Fenpropathrin	50.1	57.2	P/P	35 - 120
2393976	Gamma-BHC	105	78.5	P/P	63 - 128
2393976	Gamma-Chlordane	61.3	65.3	P/P	40 - 104
2393976	Heptachlor	55.3	45.5	P/P	36 - 97.0
2393976	Heptachlor Epoxide	97.4	92.2	P/P	49 - 184
2393976	Hexachlorobenzene	40.3	44.0	P/P	39 - 96.0
2393976	Kepone	126	122	P/P	10 - 217
2393976	Lambda-Cyhalothrin	47.5	60.9	P/P	21 - 193
2393976	Methoprene	44.4	52.8	P/P	20 - 106
2393976	Methoxychlor	62.4	63.6	P/P	53 - 118
2393976	MGK-264	67.5	69.0	P/P	40 - 118
2393976	Mirex	45.6	55.0	P/P	26 - 92.0
2393976	Oxychlordane	60.5	64.6	P/P	44 - 99.0
2393976	Permethrin	52.1	66.0	P/P	33 - 182
2393976	Phenothrin	35.0	47.1	P/P	10 - 129
2393976	Piperonyl Butoxide	82.0	113	P/P	10 - 211
2393976	Prallethrin	61.7	66.0	P/P	24 - 113
2393976	Tau-Fluvalinate	50.6	63.7	P/P	14 - 149
2393976	Tetramethrin	59.8	65.7	P/P	17 - 151
2393976	Trans-Nonachlor	59.2	69.4	P/P	42 - 102
2393976	Triclosan Methyl	68.5	69.2	P/P	48 - 108
2393976	Trifluralin	59.5	61.5	P/P	47 - 96.0
2394026	PCB-1016	63.8	66.9	P/P	46 - 111
2394026	PCB-1260	64.1	66.9	P/P	45 - 114

Reference Method: EPA 8276

Batch ID: P427174

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2394026	Chlordane	86.2	84.0	P/P	53 - 126
2394026	Toxaphene	115	123	P/P	56 - 211

Reference Method: SM 5310 B-2011

Batch ID: P427625

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2393846	Organic Carbon	94.0	94.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P427216

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393839	Color (true)	0.00932	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P427201

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393841	Turbidity	15.8	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P427727

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2394077	Chloride	1.78	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395549	Ammonia-N	0.437	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P427433

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393942	Kjeldahl Nitrogen	1.11	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427398

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393849	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427333

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2394042	Total-P	1.36	Spike	P	0 - 20

Reference Method: EPA 8270E
Batch ID: P427174

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393976	Aldrin	3.08	Spike	P	0 - 41
2393976	Alpha-BHC	67.2	Spike	F	0 - 25

Quality Assurance Report Precision

Reference Method: EPA 8270E

Batch ID: P427174

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393976	Alpha-Chlordane	0.746	Spike	P	0 - 33
2393976	Beta-BHC	8.28	Spike	P	0 - 36
2393976	Beta-Cyfluthrin	18.0	Spike	P	0 - 42
2393976	Bifenthrin	15.6	Spike	P	0 - 53
2393976	Carbophenothion	10.4	Spike	P	0 - 49
2393976	Chlorothalonil	33.7	Spike	P	0 - 71
2393976	Cis-Nonachlor	15.8	Spike	P	0 - 29
2393976	Cypermethrin	22.5	Spike	P	0 - 40
2393976	Dacthal	9.25	Spike	P	0 - 26
2393976	DDD-p,p'	1.57	Spike	P	0 - 39
2393976	DDE-p,p'	2.22	Spike	P	0 - 33
2393976	DDT-p,p'	5.94	Spike	P	0 - 40
2393976	Delta-BHC	1.97	Spike	P	0 - 37
2393976	Deltamethrin	28.6	Spike	P	0 - 100
2393976	Dichlobenil	5.33	Spike	P	0 - 40
2393976	Dicofol	6.44	Spike	P	0 - 31
2393976	Dieldrin	6.02	Spike	P	0 - 27
2393976	Endosulfan I	5.48	Spike	P	0 - 23
2393976	Endosulfan II	3.70	Spike	P	0 - 28
2393976	Endosulfan Sulfate	7.67	Spike	P	0 - 38
2393976	Endrin	2.96	Spike	P	0 - 63
2393976	Endrin Aldehyde	18.7	Spike	P	0 - 60
2393976	Endrin Ketone	3.62	Spike	P	0 - 37
2393976	Esfenvalerate	31.9	Spike	P	0 - 52
2393976	Ethalfuralin	3.26	Spike	P	0 - 23
2393976	Fenpropathrin	13.1	Spike	P	0 - 32
2393976	Gamma-BHC	28.7	Spike	F	0 - 17
2393976	Gamma-Chlordane	6.39	Spike	P	0 - 40
2393976	Heptachlor	19.4	Spike	P	0 - 29
2393976	Heptachlor Epoxide	5.49	Spike	P	0 - 25
2393976	Hexachlorobenzene	8.75	Spike	P	0 - 36
2393976	Kepone	3.06	Spike	P	0 - 93
2393976	Lambda-Cyhalothrin	24.7	Spike	P	0 - 55
2393976	Methoprene	17.3	Spike	P	0 - 53
2393976	Methoxychlor	1.80	Spike	P	0 - 29
2393976	MGK-264	2.19	Spike	P	0 - 35
2393976	Mirex	18.7	Spike	P	0 - 46
2393976	Oxychlordane	6.60	Spike	P	0 - 29
2393976	Permethrin	23.5	Spike	P	0 - 44
2393976	Phenothrin	29.3	Spike	P	0 - 56
2393976	Piperonyl Butoxide	31.5	Spike	P	0 - 100
2393976	Prallethrin	6.78	Spike	P	0 - 42
2393976	Tau-Fluvalinate	22.9	Spike	P	0 - 54
2393976	Tetramethrin	9.35	Spike	P	0 - 51
2393976	Trans-Nonachlor	15.8	Spike	P	0 - 37
2393976	Triclosan Methyl	1.09	Spike	P	0 - 31
2393976	Trifluralin	3.30	Spike	P	0 - 30
2394026	PCB-1016	4.78	Spike	P	0 - 32
2394026	PCB-1260	4.29	Spike	P	0 - 31

Quality Assurance Report Precision

Reference Method: EPA 8276

Batch ID: P427174

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2394026	Chlordane	2.56	Spike	P	0 - 25
2394026	Toxaphene	7.08	Spike	P	0 - 26

Reference Method: SM 2320 B-2011

Batch ID: P427349

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2394010	Total Alkalinity	3.92	Sample	F	0 - 3.8

Reference Method: SM 2540 C-2015

Batch ID: P427644

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393812	TDS	5.30	Sample	P	0 - 10

Reference Method: SM 4500-F- C-2011

Batch ID: P427356

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2394010	Fluoride	0.989	Spike	P	0 - 1.6

Reference Method: SM 5310 B-2011

Batch ID: P427625

Replicated

Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2393846	Organic Carbon	0.266	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2393968
Field Sample ID: G4NW0277

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270E	Decachlorobiphenyl	99.1	P	29 - 130
EPA 8270E	Decachlorobiphenyl	99.1	P	30 - 118
EPA 8270E	Tetrachloro-m-xylene	73.3	P	29 - 104
EPA 8270E	Tetrachloro-m-xylene	78.5	P	30 - 97.0
EPA 8276	Decachlorobiphenyl	48.6	P	30 - 101
EPA 8276	PCNB	103	P	48 - 121

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A118016
Included Lab Sample IDs: 2393964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	100	P/P	95 - 105

Reference Method: DEP SOP: NU-094-1
Run ID: A118035
Included Lab Sample IDs: 2393964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	102	P/P	90 - 110

Reference Method: SM 2320 B-2011
Run ID: A118077
Included Lab Sample IDs: 2393964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	96.1	94.7	P/P	90 - 110

Reference Method: SM 4500-F- C-2011
Run ID: A118077
Included Lab Sample IDs: 2393964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A118103
Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	103	102	P/P	90 - 110

Reference Method: EPA 8276
Run ID: A118123
Included Lab Sample IDs: 2393968

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	96.1		P	80 - 120
Toxaphene	94.6		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A118124
Included Lab Sample IDs: 2393964

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	96.4	97.1	P/P	90 - 110
Sulfate	96.8	97.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A118134
Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A118134

Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.8	97.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A118158

Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.0	97.8	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A118179

Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.6	97.4	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A118183

Included Lab Sample IDs: 2393965

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.4	101	P/P	90 - 110

Reference Method: EPA 8270E

Run ID: A118203

Included Lab Sample IDs: 2393968

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
PCB-1016	104		P	80 - 120
PCB-1260	101		P	80 - 120

Reference Method: EPA 8270E

Run ID: A118251

Included Lab Sample IDs: 2393968

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	98.7		P	80 - 120
Alpha-BHC	103		P	80 - 120
Alpha-Chlordane	93.1		P	80 - 120
Beta-BHC	99.6		P	80 - 120
Beta-Cyfluthrin	91.2		P	80 - 120
Bifenthrin	101		P	80 - 120
Carbophenothion	96.1		P	80 - 120
Chlorothalonil	98.5		P	80 - 120
Cis-Nonachlor	103		P	80 - 120
Cypermethrin	95.8		P	80 - 120
Dacthal	103		P	80 - 120
DDD-p,p'	102		P	80 - 120
DDE-p,p'	105		P	80 - 120
DDT-p,p'	106		P	80 - 120
Delta-BHC	92.8		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270E
Run ID: A118251
Included Lab Sample IDs: 2393968

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Deltamethrin	89.3		P	80 - 120
Dichlobenil	107		P	80 - 120
Dicofol	97.2		P	80 - 120
Dieldrin	102		P	80 - 120
Endosulfan I	103		P	80 - 120
Endosulfan II	102		P	80 - 120
Endosulfan Sulfate	95.3		P	80 - 120
Endrin	105		P	80 - 120
Endrin Aldehyde	97.2		P	80 - 120
Endrin Ketone	114		P	80 - 120
Esfenvalerate	93.2		P	80 - 120
Ethalfuralin	97.0		P	80 - 120
Fenpropathrin	92.6		P	80 - 120
Gamma-BHC	102		P	80 - 120
Gamma-Chlordane	105		P	80 - 120
Heptachlor	105		P	80 - 120
Heptachlor Epoxide	103		P	80 - 120
Hexachlorobenzene	100		P	80 - 120
Kepone	80.4		P	80 - 120
Lambda-Cyhalothrin	95.4		P	80 - 120
Methoprene	104		P	80 - 120
Methoxychlor	98.2		P	80 - 120
MGK-264	92.7		P	80 - 120
Mirex	107		P	80 - 120
Oxychlordane	102		P	80 - 120
Permethrin	96.6		P	80 - 120
Phenothrin	95.4		P	80 - 120
Piperonyl Butoxide	100		P	80 - 120
Prallethrin	101		P	80 - 120
Tau-Fluvalinate	92.9		P	80 - 120
Tetramethrin	93.1		P	80 - 120
Trans-Nonachlor	104		P	80 - 120
Triclosan Methyl	101		P	80 - 120
Trifluralin	98.7		P	80 - 120

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
DEP SOP: NU-094-1	Color (true)	101				0.0093	
EPA 180.1 Rev. 2.0	Turbidity	100				15.8	
EPA 300.0 Rev. 2.1	Chloride	98.7	97.0	97.3		1.78	
	Sulfate	98.7	95.4	96.4			
EPA 350.1 Rev. 2.0	Ammonia-N	93.2	97.8	98.4			0.437
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100	97.5	99.0			1.11
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	99.0	99.0			0.0
EPA 365.1 Rev. 2.0	Total-P	101	103	104			1.36
EPA 8270E	Aldrin	35.9	38.4	39.6			3.08
	Alpha-BHC	84.1	96.4	47.9			67.2
	Alpha-Chlordane	53.9	59.5	59.1			0.746
	Beta-BHC	72.5	80.8	74.4			8.28
	Beta-Cyfluthrin	50.7	59.9	71.8			18.0
	Bifenthrin	43.5	59.3	69.3			15.6
	Carbophenothion	36.6	39.6	44.0			10.4
	Chlorothalonil	41.9	40.1	28.6			33.7
	Cis-Nonachlor	56.6	55.6	65.1			15.8
	Cypermethrin	51.1	61.4	77.0			22.5
	Dacthal	92.6	115	105			9.25
	DDD-p,p'	58.5	63.5	62.5			1.57
	DDE-p,p'	55.4	64.3	65.8			2.22
	DDT-p,p'	66.3	72.3	76.7			5.94
	Delta-BHC	75.7	79.7	78.2			1.97
	Deltamethrin	54.3	67.5	90.1			28.6
	Dichlobenil	56.7	48.7	51.4			5.33
	Dicofol	51.8	56.1	52.6			6.44
	Dieldrin	62.9	66.9	71.1			6.02
	Endosulfan I	64.1	66.1	69.8			5.48
	Endosulfan II	71.6	70.4	73.1			3.70
	Endosulfan Sulfate	71.7	72.6	67.3			7.67
	Endrin	62.2	69.4	71.4			2.96
	Endrin Aldehyde	59.0	62.3	51.6			18.7
	Endrin Ketone	87.5	97.1	93.7			3.62
	Esfenvalerate	43.6	53.8	74.2			31.9
	Ethalfuralin	54.2	59.3	57.4			3.26
	Fenpropathrin	48.3	50.1	57.2			13.1
	Gamma-BHC	88.5	105	78.5			28.7
	Gamma-Chlordane	55.5	61.3	65.3			6.39
	Heptachlor	50.8	55.3	45.5			19.4
	Heptachlor Epoxide	79.1	97.4	92.2			5.49
	Hexachlorobenzene	40.1	40.3	44.0			8.75
	Kepone	125	126	122			3.06
	Lambda-Cyhalothrin	36.0	47.5	60.9			24.7
	Methoprene	35.2	44.4	52.8			17.3
	Methoxychlor	61.0	62.4	63.6			1.80
	MGK-264	60.5	67.5	69.0			2.19
	Mirex	40.1	45.6	55.0			18.7
	Oxychlordane	54.7	60.5	64.6			6.60
	PCB-1016	76.9	63.8	66.9			4.78
	PCB-1260	76.1	64.1	66.9			4.29
	Permethrin	42.7	52.1	66.0			23.5
	Phenothrin	31.6	35.0	47.1			29.3
	Piperonyl Butoxide	75.0	82.0	113			31.5
	Prallethrin	59.5	61.7	66.0			6.78

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 8270E	Tau-Fluvalinate	37.8	50.6	63.7		22.9
	Tetramethrin	62.1	59.8	65.7		9.35
	Trans-Nonachlor	55.4	59.2	69.4		15.8
	Triclosan Methyl	64.9	68.5	69.2		1.09
	Trifluralin	56.1	59.5	61.5		3.30
EPA 8276	Chlordane	80.5	86.2	84.0		2.56
	Toxaphene	82.9	115	123		7.08
SM 2320 B-2011	Total Alkalinity	95.8			3.92	
SM 2540 C-2015	TDS	105			5.30	
SM 2540 D-2015	TSS	99.0				
SM 4500-F- C-2011	Fluoride	103				0.989
SM 5310 B-2011	Organic Carbon	94.8	94.0	94.3		0.266

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2393964
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2393964
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2393964
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2393964
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2393965
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2393965
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2393965
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2393965
EPA 8270E	Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS	2393968
EPA 8270E	PCBs in water matrices by GC/MS/MS	2393968
EPA 8276	Chlordane and Toxaphene in water matrices by GC/MS with Negative Chemical Ionization	2393968
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2393964
SM 2540 C-2015	Total Dissolved Solids in aqueous matrices	2393964
SM 2540 D-2015	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2393964
SM 4500-F- C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2393964
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2393965

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	03/15/2023			03/15/2023 17:37	Alexis Price	2393964
EPA 180.1 Rev. 2.0	03/15/2023			03/15/2023 16:00	Anna M. Plaviak	2393964
EPA 300.0 Rev. 2.1	03/15/2023			03/23/2023 13:59	Judith K. Clark	2393964
EPA 350.1 Rev. 2.0	03/15/2023			03/24/2023 10:50	Ping Hua	2393965
EPA 351.2 Rev. 2.0	03/15/2023	03/22/2023 10:42	Simonne.V. Calhoun	03/23/2023 11:29	Judith K. Clark	2393965
EPA 353.2 Rev. 2.0	03/15/2023			03/21/2023 10:09	Beverly Y. Sanders	2393965
EPA 365.1 Rev. 2.0	03/15/2023	03/20/2023 14:58	Adam P Silver	03/22/2023 11:10	James L Waggeberby	2393965
EPA 8270E	03/15/2023	03/16/2023 08:00	Fe-Val Siddell	03/23/2023 00:06	Julie Wi	2393968
	03/15/2023	03/16/2023 08:00	Fe-Val Siddell	03/23/2023 01:08	Julie Wi	2393968
EPA 8276	03/15/2023	03/16/2023 08:00	Fe-Val Siddell	03/18/2023 08:30	Lipi Saha	2393968
SM 2320 B-2011	03/15/2023			03/18/2023 11:52	Dale L. Simmons	2393964
SM 2540 C-2015	03/15/2023			03/17/2023 15:51	Yijie Li	2393964
SM 2540 D-2015	03/15/2023			03/17/2023 15:51	Yijie Li	2393964
SM 4500-F- C-2011	03/15/2023			03/18/2023 11:52	Dale L. Simmons	2393964
SM 5310 B-2011	03/15/2023			03/23/2023 16:51	Elise M. Gillis	2393965